MILITARY SPECIFICATION

CURETTES, PERIODONTAL (McCall)

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.2 Scope. This specification covers double-ended periodontal dental curettes of McCall design.

1.2 Classification. Curettes shall be of the following styles, as specified (see 6.1):

Style 1. - McCall, No. 13S and 14S.

Style 2. - McCall, No. 17S and 18S.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. The following specifications and standards form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents are those listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:43Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Federal

GG-I-526	- Instruments, Dental and Surgical; General Specifications.
PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-601	- Boxes, Wood, Cleated-plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-B-676	- Boxes, Setup.

Military

MIL-P-116	- Methods of Preservation.
MIL-L-10547	- Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Federal

FED-STD-66	- Steel, Chemical Composition and Hardenability.

Military

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.
MIL-STD-147	- Palletized Unit Loads.
MIL-STD-794	- Parts and Equipment, Procedures for Packaging of.

DRAWING

DEPENSE PERSONNEL SUPPORT CENTER (DPSC)

24034	- Curettes, Periodontal, McCall

(Copies of specifications, standards, and drawings required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity).

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:43Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise specified, the issue of the documents which are DOD adopted are those listed in the issue of the DODISS specified in the solicitation. Unless otherwise specified, the issues of documents not listed in the DODISS are the issues of the non-Government documents which are current on the date of the solicitation.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

B 135 - Specification for Seamless Brass Tube.

E 18 - Standard Test Methods for Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials.

E 92 - Standard Test Method for Vickers Hardness of Metallic Materials.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103).

(Nongovernment standards and other publications are normally available from the organizations which prepare or which distribute the documents. These documents also may be available in or through libraries or other informational services).

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification takes precedence. Nothing in this specification, however supersedes applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 General requirements. The curettes shall be in accordance with applicable requirements of GG-I-526 with the options and additional requirements specified herein.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:43Z
Check the source to verify that this is the currentversion before use.

3.2 Bid sample. Twelve bid samples which are representative of the curettes to be supplied shall be furnished complete with protocals and satisfactory report of tests performed in accordance with the specification to the Defense Personnel Support Center for evaluation by the Military Medical Services. When tested by periodontists under conditions of normal usage, the curettes shall be suitable for use in periodontal surgical procedures. The curettes shall be well balanced and the blade shall be finished and contoured to produce a sharp edge around the blade which shall effectively accomplish subgingival curettage, remove serumal calculus from root surfaces and thoroughly polish the cementum to assist in physiological tissue reattachment. The finish of the finger grip portion of the handle shall be professionally acceptable in providing proper control and comfort during operative use.

3.3 Material.

3.3.1 Operative ends. The operative ends shall be class 4, Type 440A, 440B, or 440C stainless steel.

3.3.2 Handles. Handle A shall be fabricated of class 4, Type 416 stainless steel as specified in GG-I-526 or ferritic stainless steel conforming to the 430F stainless steel in accordance with FED-STD-66 for composition only. Handle C shall be fabricated of class 3, austenitic stainless steel, Type 303. Handle B and D shall be chromium over nickel plated brass. Brass shall conform to copper alloy No. 260 or 330 as specified in ASTM B 135 for chemical composition only. The corrosion tests, as specified in 3.8 shall apply.

3.4 Hardness. After proper stress relieving, the Rockwell hardness of the operative ends shall be C52 to C56 for Type 440A, C54 to C58 for Type 440B, and C56 to C60 for Type 440C. Handles and shanks shall be properly tempered and shall have a hardness suitable for the purpose intended. Hardness shall be determined as specified in 4.4.1.

3.5 Style, design and dimensions. The curettes shall be double-ended McCall design curettes. The curettes shall be furnished in the style specified. The curettes shall be suitable for use in subgingival curettage during periodontal treatment. The style, design and dimensions of the curettes shall conform to Defense Personnel Support Center (DPSC) Drawing 24034 for the applicable style. Handles A, B, C, or D (with appropriate shank diameters), conforming to DPSC drawing 24034 shall be furnished.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:43Z
Check the source to verify that this is the current version before use.

3.5.1 Drawings. The drawings forming a part of this specification are specification control drawings. The supplier shall be responsible for preparing his own shop drawings. Where tolerances prescribed could cumulatively result in improper design, the supplier shall provide a product with dimensions within the tolerances prescribed on the drawing to insure proper design, function, fit, assembly and operation of the item. No deviation from the prescribed dimensions or tolerances is permissible without prior approval of the contracting officer.

3.6 Construction. The curettes shall be of double-ended construction consisting of a handle and an operative end at each end. The operative ends shall be permanently secured to the handle. The blades of the operative ends of each curette shall face in opposite directions. Handles A and C shall be of solid construction. Handles B and D shall be of hollow construction. Junction of the operative end to the handle shall be smoothly finished and with no crevices evident. There shall be no movement, loosening, or separation between the operative end and the handle when the operative end is subjected to an 85 pound (minimum) pull out force, when tested as specified in 4.4.2.

3.7 Finish. The operative end including shank shall have a smooth mirror or smooth bright finish, except for the surfaces that form the cutting edges which shall have a finely honed finish. The operative end (entire blade) shall have a sharp cutting edge and shall be free of defects such as rough grind marks, cracks, gouges, burrs, feathers, or other imperfections when viewed as specified in 4.3.1. Handles shall have a mirror or bright finish, except for the serrated or knurled portion. Finger grip or serrated portions of handles shall be free from barbs and shall provide adequate finger grip during use.

3.8 Corrosion resistance. The copper sulfate test shall apply to operative ends, ferritic and martensitic stainless steel handle. The boil test shall apply to the entire curette (operative ends and all handles).

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:43Z
Check the source to verify that this is the current version before use.

3.9 Identification marking. In addition to the requirements of GG-I-526, each curette shall be permanently and legibly marked with the identification "MC 13S-14S" or "MC 17S - 18S" as applicable for the style specified. All marking shall be on the smooth portion of the handle and shall be permanent and legible. Each curette shall be marked with the name or registered trademark of the manufacturer (and contractor, if other than the manufacturer). The word "stainless", when applicable, shall also be marked on instruments of foreign manufacture. As an alternate to the word "stainless", the use of a registered trade name identifying the steel as stainless shall be acceptable, provided prior approval is received from the cognizant contracting officer.

3.10 Workmanship. In addition to the requirements of GG-I-526, the operative end shall be free of pits, discoloration, and rounded edges.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Sampling for inspection shall be performed in accordance with MIL-STD-105, except where otherwise indicated hereinafter.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:43Z
Check the source to verify that this is the current version before use.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the raw material (stainless steel and brass), may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in table I, II and III (see 4.2.1, 4.2.2, and 4.3) and 4.3.2.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQLs) shall be as indicated in table I. The unit of product for examination purposes shall be one curette.

TABLE I. Examination criteria

Type Defect	Inspection level	AQL (Percent defective)
Major A	II	1.0
Major B	S-1	2.5
Minor	II	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II except that, unless otherwise specified, the acceptance number shall be zero (0) for all sample sizes. Where feasible, the same sample unit shall be used for the determination of two or more tests characteristics. The unit of product for test purposes shall be one curette.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:43Z
Check the source to verify that this is the current version before use.

TABLE II. Tests

Component	Characteristic	Rqmt.	Test Procedure	Inspection level
Entire curette	Chemical Composition	3.3.1, 3.3.2, & GG-I-526	GG-I-526	S-1 *
Operative ends	Hardness	3.4 and GG-I-526	4.4.1	S-2
Operative ends	Pull Test	3.6	4.4.2	S-2**
Operative ends and stainless steel handles (ferritic and martensitic)	Copper Sulfate	3.8	GG-I-526	S-2
Entire curette	Boil test	3.8	GG-I-526	S-2

*A sufficient quantity of each lot of metal used in the construction of the curettes may be selected for composition analysis in lieu of sample curettes, provided lot identity has been maintained and can be demonstrated to the Government.

**AQL shall be 1.0.

4.3 Examination. The curettes shall be examined to determine compliance with specification requirements. Examination shall be conducted in accordance with 4.2.1. The curettes shall be examined for defects, including but not limited to, those listed in table III.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:43Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Style and design not as specified for the applicable style and number curette.
102	Instrument not constructed as specified.
103	perative end finish not as specified.
104	Operative end not free from rough grind marks, cracks, burrs, gouges, feathers, or other imperfections that would render the curette unsuitable for use.
105	Cutting edges rounded or not sharp.
Major B	
151	Operative end dimensions not as specified.
Minor	
201	Handle finish not as specified.
202	Instrument not permanently and legibly marked with name or trademark of the contractor.
203	Instrument identification not present, properly located, or not complete.
204	Handle dimensions not as specified.
205	Curette not free from dirt, oil, compound, or foreign matter.
206	Handle knurling missing, not uniform or not complete.

4.3.1 Examination under magnification. The instruments shall be examined under four power (4X) magnification for the following number of defects: Categories 104 and 105. Examination under magnification shall include, but shall not be limited to, the defects listed herein.

4.4 Tests. Tests shall be conducted to determine compliance with the requirements of this specification (see 4.2.2).

4.4.1 Hardness. Hardness shall be determined in accordance with ASTM Designation E 18 (Rockwell) or E 92 (Vickers or Diamond Pyramid).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:43Z
Check the source to verify that this is the current version before use.

4.4.2 Pull test The handle and shank of the operative end shall be securely held in separate chucks or other holding devices. A slowly increasing pull out force, increasing from zero to 85 pounds (minimum), shall be applied to the shank of the operative end. The force shall be maintained at the high value for a minimum of one second. The instrument shall be examined for movement, looseness, or separation of the operative end with respect to the handle. The test shall be repeated for each operative end of the double ended instrument.

4.5 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C, as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Cleaning. Each curette shall be cleaned in accordance with MIL-P-116.

5.1.1.2 Drying. Each curette shall be dried in accordance with MIL-P-116.

5.1.1.3 Preservative applications. Preservatives shall not be used.

5.1.1.4 Unit package. Each curette shall be wrapped in neutral tissue or other suitable anticorrosive material. Curette shall be further packaged in accordance with MIL-P-116, method IC-2. As an alternate, the unit package shall be a rigid plastic tube having a metal screw cap or a plastic friction plug type closure. The curette shall be cushioned within the tube in such a manner to prevent damage to the operative ends.

5.1.1.5 Intermediate package. Twelve (12) curettes of one style only shall be packaged in a box conforming to PPP-B-566, PPP-B-676 or PPP-B-636, type CF, class domestic. Closure shall be as specified in the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:43Z
Check the source to verify that this is the current version before use.

5.1.2 Level C.

5.1.2.1 Unit package. Preservation for currettes shall conform to the requirements of MIL-STD-794 for level C.

5.1.2.2 Intermediate package. Twelve currettes of one style only shall be packaged in accordance with the requirements of MIL-STD-794 for level C.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.2.1 Level A. Twenty-four intermediate packages (288 units) containing curettes of one style only shall be packed in an exterior container designed for type 2 load and conforming to PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Twenty-four intermediate packages (288 units) containing curettes of one style only shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, type CF, class domestic. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. Twenty-four intermediate packages (288 units) containing curettes of one style only shall be packed in accordance with the requirements of MIL-STD-794 for level C.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:43Z
Check the source to verify that this is the currentversion before use.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination exceeds 250 lbs (excluding the pallet) or 20 cubic feet. Loads shall be unitized as specified in MIL-STD-147, on Type IV or Type V 4-way entry pallets. Pallet shall have a length of 40 inches and a width of 48 inches. Pallet loads, including the pallet, shall not exceed 54 inches in height, 40 inches in length and 48 inches in width. Quantity for shipment to one destination of less then 250 lbs or 20 cubic feet need not be palletized.

5.2.5.1 Levels A and B. Curettes packed as specified in 5.2.1 and 5.2.2, shall be unitized on pallets as specified in 5.2.5.

5.2.5.2 Level C. Curettes packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Levels A, B and C. Each unit pack, intermediate package, exterior container, and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2 and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be accordance with 4.5.

6. NOTES

6.1 Ordering data.

6.1.1 Acquisition requirements. Acquisition documents should specify the following:

(a) Title, number, and date of this specification.
(b) Style required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(d) Approval of tradename, if applicable (3.9).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:43Z
Check the source to verify that this is the current version before use.

6.2 National Stock Number (NSN) coverage. This specification covers the following items appearing in the Federal Supply Catalog:

NSN	Item Identification
6520-00-890-2050	CURETTE, PERIODONTAL, McCall, No. 13S and 14S
6520-00-890-2051	CURETTE, PERIODONTAL, McCall, No. 17S and 18S

6.3 Subject term (key word) listing

Curette
Periodontal
McCall

MILITARY INTERESTS

Custodians:

Army - MD
Navy - MS
Air Force - 03

Review Activities:

Army - MC
Navy - MS
Air Force - 03

Civil Agency Coordinating Activities:
VA-OSS
PHS

PREPARING ACTIVITY:

DOD-MB

Agent:

DLA-DM

Project No. 6520-1859

05454

Location: NCR - SYS/ATTH

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:43Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:43Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-C-36507C	Curettes, Peridontal, (McCall)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2020-10-21T08:43Z
Check the source to verify that this is the current version before use.