MILITARY SPECIFICATION

EVACUATION BOARD, CASUALTY

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers an evacuation board for use in rescuing injured personnel trapped in small enclosed spaces where manual removal is difficult.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

UU-P-271	Paper Wrapping, Waterproofed, Kraft.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-640	Box, Fiberboard, Corrugated, Triple-Wall.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWINGS

Defense Personnel Support Center Drawing (DPSC) No. 24075, Evacuation Board, Casualty.

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 First Article. Unless otherwise specified, prior to the start of production of the contract quantity, one sample of the evacuation board shall be assembled, completely finished, and tested by the contractor. The sample shall be submitted, together with the contractor's inspection and test reports, to the procuring contracting officer for examination and testing to determine compliance with the requirements of this specification. Written approval of the materials, construction, and performance of the sample shall be obtained from the procuring contracting officer, or his authorized representative, prior to the manufacture and assembly of the remaining quantity to be manufactured. Pending approval of the first article, acquisition of materials, components, or commencement of production is at the sole risk of the contractor. Accepted sample shall become the property of the procuring activity and shall be included in the quantity called for in the contract or order. In the event that the sample is not approved, the contractor shall submit a new sample for approval, or modify the existent sample so as to remove the deficiencies cited.

3.2 Material. Material shall be as listed on the drawing.

3.3 Design. Shall be a highly waxed board, fitted at one end with a winch and detachable sling, mounted between two vertical posts. Height of board mounting shall be adjustable in three inch steps. The board shall be free to pivot about its horizontal supporting member.

3.4 Construction. The evacuation board shall be constructed in accordance with Defense Personnel Support Center (DPSC) Drawing No. 24075.

3.5 Identification marking. Each evacuation board shall be fitted, in a suitable location, with a metal nameplate. The nameplate shall contain the following data:

Name or registered trademark of the manufacturer
Serial number
Federal Stock Number
Contract or purchase order number
The letters "U.S."
A nameplate secured by adhesives shall not be accepted.

3.6 Service data. The contractor shall enclose, with each unit, two copies each, of one or more service data booklets. In addition, the contractor shall mail 15 copies of the approved service data to the contracting officer for distribution. The service data shall contain complete instructions for installation, operation, maintenance, and a list of component parts. Each part of the unit other than common fasteners, shall be identified in the service data and parts list by the manufacturer's part number.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

3.6.1 Approval. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the contracting officer, via the cognizant Government inspector, two preliminary copies of his proposed service data, for review and comment by Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary service data shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of required revisions and corrections in the final service data without delaying delivery under the contract or purchase order.

3.6.2 Waiver of Review and Distribution Copies of Manual. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of contract, and proposes to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request.

3.7 Workmanship. The workmanship, finish, fit, safety, and details shall be first class in every respect. Evacuation boards shall be free from defects which detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier shall utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to insure that supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

4.2 Sampling. Sampling for examination shall be conducted in accordance with MIL-STD-105.

For visual examination	Inspection level	AQL (defects per hundred units)
Major defects	II	1.0
Total (major and minor combined)	II	4.0
Dimensions	S-3	2.5

4.2.1 For tests. Sampling for tests shall be conducted in accordance with Inspection level S-3 of MIL-STD-105. The acceptance level shall be zero.

4.3 Examination. The evacuation boards shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects.

Categories and defects*

Major

101 Any component missing, broken, or otherwise unuseable.
102 Unit not free of cracks, tears, ruptures, and deformed surfaces.
103 Unit not stable on a flat surface, (wobbles)
104 Exposed or accessible edges, corners, and surfaces, not free of burrs, sharpness, and sharp projections.
105 Legs or posts not smoothly shaped; not of equal length.
106 Wood part of board not free of warpage or open grains and upper side of board not smoothly polished or waxed.
107 Sling, and rope, not sufficiently flexible.

Minor

201 Identification marking not complete, permanent, or legible.
202 Rubber tips not snug fitting.
203 End of each leg not finished to preclude damage to rubber tips.

4.3.2 Dimensional examination. The evacuation boards shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:44Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. Each evacuation board, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Unit preparation. Each unit shall be disassembled as far as is practicable. Surfaces subject to corrosion shall be cleaned and treated with a suitable rust preventive. Al exterior surfaces shall be protected by wrapping with kraft paper and cushioned with flexible corrugated fiberboard or other suitable materials. Wrapping and cushioning material shall be secured by tying with cord or by sealing with kraft tape. Work shall be accomplished in such a manner as to insure against marring, scratching, or damage to the unit.

5.1.3 Level C. Units shall afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A. Each unit shall be overpacked in an exterior container designed for a type 3 load, and constructed in accordance with PPP-B-601, table II, using type III, class 1 plywood or PPP-B-640, class 2, style F. Each plywood box shall be lined with a case liner on the top, ends and sides conforming to UU-P-271, class L-2. Case liner shall not be required for fiberboard boxes. The unit shall be secured within the container by means of wood blocking, bracing, or other suitable devices to insure against shifting of or damage to the contents. Closure of wood boxes shall be as specified in the appendix of PPP-B-601. Closure and sealing of fiberboard boxes shall be as specified in the appendix of PPP-B-640.

5.2.2 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packing (see 5.1 and 5.2).
(c) When first article is not required (see 3.1).
(d) Where first article shall be sent (see 3.1).

6.2 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Category: N

*U.S. GOVERNMENT PRINTING OFFICE: 1987-251-518/5602

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION

MIL-E-36518(DSA-DM)

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 68 1426

REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:44Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
AUG 66

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:44Z
Check the source to verify that this is the current version before use.