MIL-H-3649B
22 July 1966

MILITARY SPECIFICATION

Cup And Holder, Dental Waste

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a metal receptacle and a disposable plastic insert (cup) suitable to accept waste material during dental operative procedures.

1.2 Classification. Dental waste cups and holders shall be of the following classes as specified (sec. 6.2).

1.2.1 Class 1. Cups (without holder)

1.2.2 Class 2. Holders (without cups)

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical, General Specifications.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Box, Setup.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:42Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWINGS

Defense Personnel Support Center

23032	Cup and Holder Assembly, Dental Waste

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Holder. The dental waste cup holder shall be fabricated of class 3, type 303 stainless steel conforming to GG-I-526.

3.1.2 Cup. The bottom section of the dental waste cup shall be fabricated of molded, semi-rigid, crystal clear polystyrene. The cap shall be fabricated of molded, flexible, white, opaque polyethylene. The plastic materials shall be suitable for the purpose intended.

3.2 Style, design, and dimensions. The waste cup holder and waste cup shall be, respectively, a metal receptacle and a disposable plastic insert suitable for use to accept waste material during dental operative procedures to eliminate the possibility of cross contamination. The waste cup holder shall weigh not less than 200 grams. The bottom shall be hexagonal in shape to provide stability, and the top shall be round. The corners at the bottom of the base shall be properly radiused to prevent scratching and marring. The waste cup shall be fabricated in two parts. The cap section shall have five slits radiating from the center. The slits shall have a radius on their upper aspect to facilitate entrance of waste from dental forceps, and shall be square on under surface to facilitate removal of waste from dental forceps. The top surface of cap shall have a peripheral bead to retain fluids and prevent skidding off cap. The bottom section of plastic insert shall have a gross liquid capacity of not less than 10 cubic centimeters. The design and dimensions shall comply with DPSC drawing No. 23032.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:42Z
Check the source to verify that this is the current version before use.

3.3 Finish. The waste cup holder shall have a mirror finish and waste cup shall have a smooth finish.

3.4 Resistance to boil test. The waste cup holder shall comply with the boil test requirements of GG-I-526.

3.5 Identification marking. The waste cup holder and waste cup shall be permanently and legibly marked with the name or registered trademark of the manufacturer. Marking shall appear on a surface where it will not adversely affect proper functioning.

3.6 Workmanship. Workmanship shall be first class throughout. The waste cup holder and waste cup shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table I. The unit of product shall be one dental waste cup or one dental waste cup holder.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:42Z
Check the source to verify that this is the current version before use.

Table I. Sampling criteria

Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Boil test	3.4	GG-I-526	S-2	*
Dimensions	3.2	S.M.I.	S-2	4.0

* Acceptance number shall be 0 for all sample sizes.

4.2.3 Raw materials. A sufficient quantity of each lot of metal used shall be selected for chemical analysis.

4.2.4 Where feasible, the same unit of product may be used for the determination of two or more test characteristics.

4.3 Examination. The waste cup holder and waste cup shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) holder, as specified, constitutes one unit of issue, one (1) each. Two hundred fifty (250) cups, as specified, constitute one unit of issue, one (1) package.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I. Packaging quantities

Item	Unit package	Intermediate package	Exterior container
Holder	1 unit	12 units	48 units
Cup	1 unit	Not Required	4 units

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:42Z
Check the source to verify that this is the current version before use.

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Unit package shall be a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors will be acceptable. Closure shall be adequate to prevent spilling of contents under normal handling.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors will be acceptable, or PPP-B-636. Closure shall be adequate to prevent spilling of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3,for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4, of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:42Z
Check the source to verify that this is the current version before use.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic and method I closure shall be utilized.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The metal receptacle and disposable plastic insert (cup) shall be suitable to accept waste material during dental operative procedures.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packing (see 5.1 and 5.2).
(c) Class required (see 1.2.1 and 1.2.2).

6.3 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Category: N

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:42Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity (as indicated on reverse hereof).

SPECIFICATION

MIL-H-36498

ORGANIZATION (Of submitter)	CITY AND STATE

CONTRACT NO.	QUANTITY OF ITEMS PROCURED	DOLLAR AMOUNT
		$

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES.

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES", IN WHAT WAY?

4. REMARKS (Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)

SUBMITTED BY (Printed or typed name and activity)	DATE

DD FORM 1426

SMD PPSO 33983

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:42Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

FOLD UNDER

DEFENSE SUPPLY AGENCY
DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20th ST
PHILADELPHIA, PA. 19101

POSTAGE FEES PAID
DEFENSE SUPPLY AGENCY

OFFICIAL BUSINESS

DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER

ATTN: DIRECTORATE MEDICAL MATERIEL

2800 SOUTH 20TH STREET

PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:42Z
Check the source to verify that this is the current version before use.