MIL-L-36531A
19 September 1977
SUPERSEDING
MIL-L-36531
6 September 1967

MILITARY SPECIFICATION

LIGHT, SURGICAL, FIELD

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one type of portable field operating light, suitable for use as a self-standing floor model.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

TT-E-527	- Enamel, Alkyd, Lusterless.
TT-E-529	- Enamel, Alkyd, Semi-Gloss.
TT-P-636	- Primer Coating, Synthetic, Wood and Ferrous Metal.
TT-W-572	- Wood-Perservative, Water-Repellant.
TT-P-1757	- Primer Coating, Zinc Chromate, Low-Moisture-Sensitivity.

MILITARY

MIL-M-7298	- Manual, Technical, Commercial Equipment.
MIL-A-8625	- Anodic Coatings, for Aluminum and Aluminum Alloys.
MIL-P-15328	- Primer, Pretreatment (Formula No. 117 for Metals).

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:45Z
Check the source to verify that this is the current version before use.

STANDARDS

FEDERAL

FED-STD-595 - Colors.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWINGS

Defense Personnel Support Center (DPSC)

25010 - Light, Surgical, Operating, Field.

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

American National Standards Institute (ANSI)

C73.11 - Plugs and Receptacles: General Purpose, 125 volts, 15 amperes, 2 pole, 3 wire, Grounding Type.

(Applications for copies may be addressed to the American National Standards Institute, 1430 Broadway, New York, NY 10016.)

American Welding Society

Welders Handbook

(Application for copies may be addressed to the American Welding Society, 345 Eash 47th Street, New York, NY 10017.)

3. REQUIREMENTS

3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.1.2.1 and 6.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:45Z
Check the source to verify that this is the current version before use.

3.2 Material. Materials used in the construction of the field operating light shall be as specified herein and as listed in the Bill of Materials to DPSC Drawing No. 25010. Commercial tolerances shall apply to all thicknesses of materials.

3.3 Design and construction. The field operating light shall consist of the following:

a. Standard No. 3, Medical Field Chest, modified.
b. Vertical and horizontal support members.
c. Reflector assembly.
d. Power pack.
e. Spare parts.

Total weight of the light, including chest, shall not exceed 68 pounds. Design and construction shall be in accordance with the cognizant drawing and the Bill of Materials thereto.

3.3.1 Standard No. 3, medical field chest, modified. When specified (see 6.2), the basic chest shall be furnished as Government Furnished Material and the modifications shall be accomplished by the contractor to comply with the cognizant drawing. Chest shall comply with requirements for Chest, Medical Instrument and Supply Set (see 6.4), modified in accordance with the cognizant drawing.

3.3.2 Vertical and horizontal supports. Vertical and horizontal support members shall be in accordance with the cognizant drawing. The support members shall consist of a base bracket, a three-sectioned vertical support, a two-sectioned horizontal support, four vertical stiffening cables and one horizontal stiffening cable.

3.3.3 Reflector assembly. Reflector assembly shall consist of a heat cylinder refelctor, reflector arm, spring housing bracket and power cable. Power cable shall be complete with an "off-on" through switch. The cord shall be furnished with a plug for use with receptacles specified in ANSI C73.11. Construction shall be in accordance with the cognizant drawing. The heat cylinder shall be made of blue-green Aklo glass with performance characteristics as specified in 3.4.3. The reflector shall be aluminum sheet as specified in the Bill of Materials, having a specular reflecting surface. The reflecting surface shall have a reflectivity of 83 percent minimum when tested as specified in 4.4.1.3. The reflective surface shall have a specular finish protected by a film of clear anodize of sufficient thickness to prevent clouding, staining, oxidizing or peeling during the life of the unit, and shall be tested as specified in 4.4.1.8.

3.3.4 Power pack. Power pack shall be in accordance with DPSC Drawing No. 25010 and Bill of Materials thereto. Two power cables shall be furnished as indicated in the Bill of Materials. One cable shall terminate in a 3-prong plug with ground, for connection to a 115 volt, alternating current source. The other shall terminate with insulated battery clips for connection to a 24 volt battery. Switchover to battery operation in the event of faulure of the 115 volt power supply shall be automatic.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:45Z
Check the source to verify that this is the current version before use.

3.3.5 Spare parts. Each unit shall be equipped with the following spare parts:

a. Fuse. In addition to the 10 ampere fuse mounted in the fuse holder, five 10 ampere fuses shall be supplied in the spare fuse holder.

b. Lamp. In addition to the lamp mounted in the reflector assembly, one lamp shall be inserted in the heat cylinder canister and another in the lamp bulb extractor canister.

c. Heat cylinder. In addition to the heat cylinder mounted in the reflector assembly, a heat cylinder shall be included in the heat cylinder canister.

3.4 Performance.

3.4.1 Light beam characteristics. The light shall provide the following beam characteristics:

a. Color temperature shall be 4000°K. $\pm$ 500°K. with trichromatic coefficients of .30 to .40 for "X" and .30 to .41 for "Y" when tested as specified in 4.4.1.2.

b. A minimum light output of 1800 foot candles measured 38 inches from the reflector rim.

c. Depth of field from 24 to 42 inches (see 4.4.1.4).

d. Lamphead shall not demonstrate an oscillatory motion which results in the focal spot in the light field exceeding a total lateral distance of one inch in a direction which is normal to the light beam. Further, counterbalancing assemblies shall provide for immediate "dampening" (2 seconds or less) for the lamphead assembly to return to the designated position without any indication of "drifting" (a position above or below the designated position) (see 4.4.1.5 and 4.4.2.4).

3.4.2 Lamp life. The lamp shall provide at least 200 hours of useful life at a cycle rate of 4 hours "ON" and 1/2 hour "OFF" when tested as specified in 4.4.2.5.

3.4.3 Heat cylinder. The Aklo glass heat cylinder shall absorb more than 90% of the infrared rays so that a temperature rise, as measured by a thermocouple on the blackened disk, shall be less than 4° C. per 1000 foot candles above ambient with a total temperature rise not exceeding 10° C. (see 4.4.1.6).

3.4.4 Dielectric withstand. The unit shall be capable of withstanding without breakdown, for one minute, a potential of 1240 volts, 60 Hz applied between live and dead metal parts (see 4.4.1.7 and 4.4.2.3).

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:45Z
Check the source to verify that this is the currentversion before use.

3.5 Finish.

3.5.1 Metal parts. All exposed metal parts, plus interior of power unit box and cover, shall be treated by prepriming with a wash-primer coat in accordance with MIL-P-15328, and then treated with a primer spray coat in accordance with TT-P-1757. Finish coat shall consist of two spray coats of baking type synthetic, semi-gloss enamel in accordance with TT-E-529. Color shall be 26081 conforming to FED-STD-595. The following parts shall not be painted:

a. Stiffening cables.

b. Reflecting surface.

c. Electrical components.

d. Vertical and Horizontal support members.

e. Support member locking pin assemblies.

f. Sterilizable handle.

3.5.2 Wood parts. Wood parts shall be dressed to the correct dimensions and then immersed in pentachlorophenol solution, in accordance with TT-W-572. Immersion shall be for a period of not less than four hours. Treated wood shall be thoroughly air-dried or kiln-dried prior to painting. Wood parts shall then be given a prime coat in accordance with TT-P-636, followed by two top coats of semi-gloss enamel in accordance with TT-E-529. Color shall be 26081 of FED-STD-595.

3.5.3 Medical supply chest. Any abrasions or scratches resulting from the modification of the chest to accommodate the light, shall be touched up with synthetic lusterless olive drab enamel in accordance with TT-E-527. Color shall be No. 24087 in accordance with FED-STD-595. Chest shall be rainproof after modification (see 4.4.1.9).

3.5.4 Vertical and horizontal support members. The vertical and horizontal support members shall be finished in a Type III hard anodic coating not less than 1-1/2 thousandths thick, in accordance with MIL-A-8625. Color shall be 26081 of FED-STD-595.

3.6 Identification marking. Each light and carrying case shall be permanently and legibly marked with the following:

Manufacturer's name or registered trademark.
Contract or purchase order number.
National stock number.
Electrical characteristics (volts, amperes, watts, AC-DC).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:45Z
Check the source to verify that this is the current version before use.

3.7 Service data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance and lists of component parts. The manual shall comply with the requirements of MIL-M-7298. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer two preliminary copies of his proposed technical manual for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract, or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of the contract and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the Contracting Officer. Waiver of submission of drafts and submission of 15 additional copies shall be at the option of the Government.

3.8 Workmanship. Workmanship shall be first class throughout. Item shall be free from defects which detract from its appearance or may impair its serviceability. All burrs and sharp edges shall be removed from the item.

3.8.1 Welding. Welding shall be durable, sound in every detail and in accordance with the latest recommended by the American Welding Society. All welds shall be produced by qualified welding operators with satisfactory welding equipment using satisfactory filler metal or electrode where applicable to provide the best possible weld. Surfaces to be welded shall be free from grease, oil, oxides or any other foreign matter. Any cleaning method which will remove these substances and leave the surface clean is satisfactory. All weld joints shall blend into the adjacent metal in gradual, smooth curves. Beads shall be smooth and free of slag, excessive undercut, or excessive splatter. Sufficient metal shall be added to provide a suitable fillet or reinforcement unless otherwise specified. Welds shall not be oxidized. All welds shall be continuous with no visible voids, pin holes, or cracks which affect the quality, and have more strength than required for the purpose intended. For the purpose of requirements in this specification, welded fully, fully welded and continuously welded and continuous weld shall all have the same meaning such that all the mating surfaces between the parts being welded are closed by welds without any breaks or voids. Grinding wheels used to remove excess weld metal and to prepare weld surfaces for subsequent polishing operations shall be of a suitable grit, free of metal filings or other iron materials such as iron oxides which would tend to cut into the parent metal surfaces and allow for rusting at a later date. Spot welds shall be free from surface burning, cracks, and porosity, and shall be of sufficient strength to pull the spot out of the parent metal when pulled apart.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:45Z
Check the source to verify that this is the currentversion before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2.1 Classification of inspection. The inspection requirements specified herein are classified as follows:

a. First article inspection.
b. Quality conformance inspection.

4.1.2.1.1 First article inspection. First article inspection shall be performed on the light when a first article sample is required (see 3.1). This inspection shall include examination of 4.3 and tests of 4.4.1.

4.1.2.1.2 Quality conformance inspection. Quality conformance inspection shall include the examination of 4.3 and tests of 4.4.2.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:45Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I. The unit of product for sampling purposes shall be the light.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and Minor combined)	II	4.0
For dimensional examination	S-3	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II. The unit of product for sampling purposes shall be as indicated in table II. Acceptance number is zero. Any test failure constitutes a rejection during these tests.

TABLE II. Sampling for tests.

Component (Unit of product)	Characteristic	Requirement	Test procedure	Inspection level
Light	System test		4.4.1.1 & 4.4.2.1	II
	Color temperature	3.4.1	4.4.1.2	1/
	Reflectivity	3.3.3	4.4.1.3	1/
	Depth of Field	3.4.1	4.4.1.4	1/
	Vibration	3.4.1	4.4.1.5 & 4.4.2.4	S-3
	Heat cylinder	3.4.3	4.4.1.6	1/
	Dielectric withstand	3.4.4	4.4.1.7 & 4.4.2.3	100%
Lamp	Life test	3.4.2	4.4.2.5	S-1
Reflector	Sealing test	3.3.3	4.4.1.8 & 4.4.2.2	2/
Modified chest Light including chest	Rainproof test	3.5.3	4.4.1.9	1/
	Weight test	3.3	4.4.1.10 & 4.4.2.6	S-1

1/ First article only.
2/ One per lot.

4.2.3 Raw materials. A sufficient quantity of each lot of each metal used in the manufacture of the light shall be selected for chemical analysis.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:45Z
Check the source to verify that this is the current version before use.

4.3 Examination. The light shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table III. Examination shall not be restricted to the classified possible defects listed below.

TABLE III. Classification of defects.

Categories	Defects
Major	
101	Any component part missing, broken, or other wise unusable.
102	Exposed edges not free of burrs or sharpness.
103	Components not properly aligned or fitted.
104	Any weld cracked, broken, or not free of pin holes.
105	Vertical upright not easily adjusted throughout entire range.
106	Power cord without suitable strain relief.
107	CRES cable not properly swaged with shank ball.
108	CRES cable frayed.
Minor	
201	Identification marking not complete, permanent or legible.
202	Identification marking not securely attached to unit.
203	Finish not free of chipped areas, cracks, scratches, or roughness.
204	Service data not as specified.

4.3.2 Dimensional examination. The light shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Group I Tests.

4.4.1.1 System test. A system operation simulation test shall be performed to insure proper functioning of all applicable components, including the 24 volt lead-acid vehicle battery.

4.4.1.2 Color temperature. Color temperature as specified herein establishes the isothermal relationship between the Planckian Locus and a point, which is a plot of the X, Y, coordinates of color on the chromaticity diagram. Measurement shall be performed at room temperature by means of a suitable spectrophotometer of standard make and model. The chromaticity diagram and certification by the manufacturer of the heat cylinder and certification by the lamp manufacturer shall be acceptable as evidence that the heat cylinder and lamp meet the color temperature requirements as specified herein.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:45Z
Check the source to verify that this is the current version before use.

4.4.1.3 Reflectivity measurements. Measurements shall be performed at room temperature by means of a suitable reflectometer of standard make and model. Certification by the company applying the finished to the reflector shall be acceptable as evidence that the finished reflector meets the reflectivity requirements specified herein.

4.4.1.4 Depth of field. Using a suitable scale having minimum graduations of 1/16 iches, measurements shall be taken from the rim of the reflector for measuring the depth of the field. All measurements shall be recorded. Foot candles shall also be noted during these measurements.

4.4.1.5 Lamphead vibration test. The light shall be set up with the lamphead assembly located approximately 78 inches from the ground level of the room. The lamphead shall be placed face down and positioned in a place which is parallel to the room floor. Using photographic equipment, the focal point shall be determined against an appropriate background. Location of focal point coordinates shall be recorded from the top and left hand edges of a sheet of paper in a fixed position. The lamphead handle shall be grasped and the lamphead raised at least 4 inches in vertical height and returned to its original position and released. Using a stopwatch, the time shall be recorded in seconds from the release of the lamphead assembly after its return to the original position to the final oscillation in this position. During the dampening motions of the lamphead, the lateral displacement of the focal point shall be recorded using time-lapse photographic equipment and an appropriate scale. The same procedure in its entirety shall be followed when the lamphead is lowered at least four inches in vertical height from the original position.

4.4.1.6 Heat cylinder. A black disk shall be placed under the lamp at a spot in the light beam where the radiant heat energy is at a maximum. A thermocouple shall be attached to the center of the disk and with the light off, the room ambient temperature shall be noted. The light shall then be turned on for a period of three continuous hours with the located spot concentrated on the center of the black disk. Temperature readings shall be taken during this period denoting temperature rise above ambient.

4.4.1.7 Dielectric withstand. A slowly rising potential, form 0 to 1240 volts, 60 Hz, shall be applied between live and dead metal parts. This potential shall be maintained at maximum for one minute. There shall be no evidence of breakdown.

4.4.1.8 Sealing test. Dissolve 1g of aluminum blue 2LW dye in 50 ml of distilled or deionized water. The pH of the dye test solution shall be adjusted to 5.0 + 0.5 and maintained at this value with the addition of acetic acid or sodium hydroxide as required. A drop of the water solution shall be applied to the specular surface of the reflector at room temperature. Allow it to stand for 5 minutes then rinse with running water. Saturate a soft cloth with a solution of mild soap and water and abrade the area tested. There shall be no evidence of stain or discoloration.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:45Z
Check the source to verify that this is the current version before use.

4.4.1.9 Rainproof test. The chest shall be exposed to water spray from a 1/2 - 3/4 inch diameter hose. The spray shall simulate falling rain stabilized at a constant rate of 2.0 ± 0.5 inches per hour. Each side and the top of the chest shall be exposed to the spray for a period of not less than 3 minutes. Pressure from the public water supply shall be between 40 - 60 pounds per square inch. After the spray test, the chest shall be examined for water leakage.

4.4.1.10 Weight test. The light including the chest shall be placed upon a suitable weighing scale and the weight recorded in pounds.

4.4.2 Group II Tests.

4.4.2.1 System test. Shall be as specified in 4.4.1.1.

4.4.2.2 Sealing test. Shall be as specified in 4.4.1.8.

4.4.2.3 Dielectric withstand. Shall be as specified in 4.4.1.7.

4.4.2.4 Lamphead vibration test. Shall be as specified in 4.4.1.5.

4.4.2.5 Lamp life test. Using a 24 volt input at a cycle rate of 4 hours on and 1/2 hour off, the lamp shall be tested ar room temperature ambient conditions.

4.4.2.6 Weight test. Shall be as specified in 4.4.1.10.

4.5 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of section 5. Defects shall be scored as specified in table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality shall be 4.0 defects per hundred units.

TABLE IV. Examination for preparation for delivery.

Examination	Defect
Containers	Not as specified.
Workmanship	Inadequate application of components, such as incomplete closure of container, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:45Z
Check the source to verify that this is the current version before use.

5. PACKAGING

5.1 Packaging. Packaging shall be level A.

5.1.2 Level A.

5.1.2.1 Unit package. Each light shall be disassembled and the disassembled components of the vertical and horizontal support members, reflector assembly, power pack, and spare parts shall be positioned within a government furnished container as indicated in DPSC drawing No. 25010. The government funrished container shall conform to 3.3.1.

5.2 Packing. No further packing is required.

5.3 Marking. Each unit package shall be marked as required for the exterior container in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The light shall be suitable for use in field, surgical operating rooms as a self-standing floor model.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents.

a. Title, number and date of this specification.
b. When a first article is required (see 3.1, 4.1.2.1 and 6.3).
c. Whether Government Furnished Material is to be supplied (see 3.3.1).

6.3 First article. When a first article is required, it shall be tested and approved under the appropriate provisions provisions of 7-104.55 of Armed Services Procurement Regulation. The contract or purchase order specifies the quantity, procedure for delivery and other specific instructions for inspection and approval of the first article.

6.4 The modified medical field chest referenced in 3.3.1 shall comply with requirements for NSN 6545-00-914-3480.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:45Z
Check the source to verify that this is the current version before use.

6.5 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6530-00-937-2204	LIGHT, SURGICAL, FIELD, 115 Volt, AC or 24 Volt, DC

6.6 Changes from previous issue. Asterisks are not used in this revision to identify changes with respect to the previous issue, due to the extensiveness of the changes.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DoD - MB

Preparing activity:
DLA-DM

Project No. 6530-1401

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:45Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:45Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS: This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.

NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER *(Number)* AND TITLE

MIL-L-36531A LIGHT, SURGICAL, FIELD

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1. ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 OCT 76 1426 EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:45Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

POSTAGE AND FEES PAID



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:45Z
Check the source to verify that this is the current version before use.