MIL-M-36510
12 December 1966

MILITARY SPECIFICATION

MANIKIN, RESUSCITATION TRAINING

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers one type of adult male, full size, life like manikin used for the training of personnel in closed chest cardiac compression, mouth-to-mouth- and mouth-to-nose resuscitation.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-636 Box, Fiberboard.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129 Marking for Shipment and Storage.

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

6910

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:43Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 First Article. Unless otherwise specified, prior to the start of production of the contract quantity, sample(s) of the manikin shall be assembled, completely finished, and tested by the contractor. The sample(s) shall be submitted, together with the contractor's inspection and test reports, to the procuring contracting officer for examination and testing to determine compliance with the requirements of this specification. Written approval of the materials, construction, and performance of the sample shall be obtained from the procuring contracting officer, or his authorized representative, prior to the manufacture and assembly of the remaining quantity to be manufactured. Pending approval of the first article(s), acquisition of materials, conponents, or commencement of production is at the sole risk of the contractor. Accepted sample shall become the property of the procuring activity and shall be included in the quantity called for in the contract or order. In the event that the sample is not approved, the contractor shall submit a new sample for approval, or modify the existent sample so as to remove the deficiencies cited.

3.2 Material. All material used in the fabrication of the resuscitation training manikin shall be of the best commercial quality, non-toxic and suitable in every respect for the purpose intended.

3.2.1 Plastic. The plastic used in the fabrication of the hands, feet, body, and skull shall be polyvinyl chloride. The skin shall be fabricated of polyvinyl chloride and plasticizers.

3.3 Construction. The training manikin shall be ruggedly constructed to withstand a reasonable amount of rough handling by the resuscitation trainee. The manikin shall consist of an inflatable body and removable shoulder, lung, and head assemblies, with each assembly being easily removable for cleaning. The manikin shall not be stained or discolored when tested as specified in 4.4.2. The head and shoulders shall be approximately the same weight as that of a human being.

3.3.1 Body. The body of the training manikin shall be fabricated of 0.5 millimeters (mm) thick plastic and shall be provided with an air intake and plug to enable inflation by a pump furnished as an accessory. The body shall have molded plastic hands, shoes and stockings each being firmly secured to the body material. The shoes and stockings shall be painted black. All seams shall be air tight and the body, when tested as specified in 4.4.1 shall show no signs of air leakage. The top of the body shall be so constructed to secure the shoulder assembly.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:43Z
Check the source to verify that this is the current version before use.

3.3.2 Shoulder assembly. The shoulder shall be molded plastic containing a mounting for the head assembly and an air passage to the lungs. The shoulder section shall be attached to prevent its separating from the inflatable part of the manikin. The air passage shall begin at the center of the head assembly mount and extend through a windpipe which connects to the lungs. Venting holes and a non-return air valve are provided to prevent the air in the lungs from venting back towards the trainee. The non-return valve shall be 2-1/2 inches in overall length and of suitable diameter to fit securely into the 27/32 inch diameter valve receptacle located in the shoulder assembly. The venting holes shall vent the air from the lungs to the back of the shoulder assembly. The windpipe shall be 1-3/4 inches long with a 5/16 inch inner diameter and a 19/32 inch outer diameter. A hole shall be provided in the side of the windpipe for the insertion of the lung resuscitation training gage or the locking tab mounting of the lungs.

3.3.3 Head assembly. The head assembly shall consist of the neck, skull, and face mask. The head assembly shall be fabricated of plastic (see 3.2.1) and so constructed to enable complete rotation when mounted on the shoulder assembly. The skull shall be moveable to enable tilting back on the neck and shall have a moveable jaw. The bony structures of the head and jaw shall be realistic and conform closely to the actual anatomy of the human head. The mask shall cover the entire head assembly and shall be painted to simulate the hair of a male adult. The facial expression of the mask shall be that of an unconscious male adult. The air passage from the mouth to the mounting device shall be an integral part of the mask and shall be blocked to prevent air passage when the skull is not tilted back, simulating the conditions which exist in an unconscious person. The inside of the mouth shall be realistic having simulated tongue and teeth. The texture of the oral cavity and associated structures shall be soft and natural to the touch.

3.3.4 Lung assembly. The lung shall be fabricated of plastic with the front molded to simulate the rib construction of an adult male. The lungs shall have the capacity and provide the resistance of air intake as would be found in an unconscious person. A plastic belt, of such length to encircle the inflated body, shall be attached to the back of the lungs. A hole shall be provided at the top of the lungs to receive the windpipe. A locking tab shall be provided on the molded front of the lungs to prevent the windpipe from becoming disconnected from the lungs when the training gage is not used.

3.3.5 External cardiac resuscitaion assembly. Each resuscitation training manikin shall be provided with an external cardiac resuscitation assembly complete with heart sponge pad, neckband and tubes. The heart sponge pad covering the neckband shall be fabricated of plastic. The heart pad shall be provided with a tabe on each side containing a snap connector to mate with snap connector provided on the plastic coated steel chest band. The flexible tubes shall be transparent plastic tubing having a 1/8 inch inner diameter and a 3/16 inch outer diameter. The tubes connecting the heart pad and the neckband shall be 7 inches in length and

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:43Z
Check the source to verify that this is the current version before use.

e extra tube, for connection to the visual training gage, shall be not less an 24 inches in length. All tubes, connected to the assembly, shall be serted into the molded receiving holes. Each tube shall be inserted not less an 1-1/2 inches to insure a good connection. The neckband shall contain an rtight pouch at the end of each tube for the purpose of transmitting a pulse en the heart pad is compressed. The neckband shall encircle the neck, being cured by snap connectors located at the back. The tube connections shall be cated on each side of the neck simulating the location for testing the cartoid se in a human being. The extra tube shall be provided with a removable plug enable use of the assembly without attachment to the visual training gage.

3.3.6 Chest band. The chest band shall be fabricated of 3/64 inch thick eel and coated evenly with plastic. The chestband shall be 1-3/4 inches wide, rming a rigid loop with an overall height of 8-1/4 inches and length of 10-1/4 ches. The top of the chestband shall be constructed of two steel strips run- ng parallel, between which shall fit the heart sponge pad (see 3.3.5). The wer steel strip shall contain snap connectors secured to the plastic coating r securing the heart sponge pad. The upper steel strip shall contain two short astic belts with snap connectors, for securing the plastic lung belt (see 3.3.4).

3.3.7 Sweat shirt and pants. The resuscitation training manikin shall be pplied with a sweat shirt and pants fabricated of cotton jersey material. The nts shall have an elastic draw-type belt and metal slide fasteners at the leg enings to enable removal of the pants over the shoes. The sweat shirt shall blue with a white shoulder area. The waist and arm openings shall be elastic nd drawn for a snug fit. The neck opening shall have a blue elastic stretch ck band. The cotton jersey clothes shall show no noticeable change in color size when tested as specified in 4.4.2. Two metal slide fasteners shall be ovided to enable removal of the shoulders, lung and heart attachment without moval of the shirt. Each slide fastener shall be not less than 5 inches in ngth and shall be jam free.

3.3.8 Carrying case. The training manikin and accessories shall be stored in rugged, shock-resisting carrying case. The case shall have a hinged cover, a rrying handle, and two locking catches. The case shall be fabricated of pressed ard with a washable coating on both sides. The case shall have a wooden frame which the pressed board is riveted. Two supporting straps shall be riveted ross both the top and the bottom surfaces. All edges and corners of the case all be reinforced smooth and rounded. The carrying case shall be furnished with protection box to protect the head assembly from damage during transportation and orage. A compartment, complete with lid, shall be permanently affixed internally one end of the case. This compartment shall secure the loose accessories during orage. Two cross belts shall be secured to the case walls and shall serve as straint belts for the deflated manikin during storage in the case. A third belt all be secured to the wood frame at one end of the case to serve as a restraint belt r the head protection box.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:43Z
Check the source to verify that this is the current version before use.

3.4 Design. The resuscitation training manikin shall be designed for training personnel in closed chest cardiac compression and mouth-to-mouth and mouth-to-nose resuscitation. The manikin shall resemble a human body, being life like and approximately 61 inches in height. The facial features and hair shall resemble that of a male adult. The head and jaw shall be moveable so that the practice of mouth-to-mouth resusciation is realistic. The air blown into the mouth by the trainee shall be vented away from the trainee. Special plastic film face pieces shall be provided for maintenance of clean and sanitary conditions. All parts of the manikin shall be easily cleaned by ordinary methods. The training manikin shall provide the same breast bone reaction as an unconscious human being and shall provide the means for simulating cartiod pulse pressure. The manikin shall be constructed to enable connection to visual training gages which indicate when heart-lung resuscitation is being correctly performed. A separate head section shall be furnished for use in demonstrating air passage construction and the dependency of natural and artificial respiration on the movement of free air.

3.5 Accessories. The resuscitation training manikin shall be furnished with the following accessories:

a. Visual training gage set
b. Eye overlay
c. 300 plastic sheets
d. Forming apparatus for disposable sanitary face masks.
e. Air pump
f. Replacement non-return air valve.
g. Return air valve.
h. Repair kit for inflatable body.
i. Separate head section.
j. Funnel

3.5.1 Visual training gage set. The visual training gage set shall consist of a plastic case, with cover, two gages, and a plastic tube. The gages shall be insulated to prevent damage in the event the case is dropped. Two external case-mounted, slip-on, tube connectors shall be provided for connecting the gages to the tubing for use with the resuscitation training manikin. Each gage meter face shall be divided to show meter readings of "too little", "correct", and "too much". The lung resuscitation gage shall be identified by the word "VENTILATION" and the heart resuscitation gage identified by the word "CHEST DEPRESSION". The plastic tube shall be approximately two feet in length and of such inside diameter to connect securely with the slide-on connectors mounted on the plastic case. The opposite end of the tubing shall have a metal connector for insertion into the hole provided in the lung and windpipe (see 3.3.4). The cover of the plastic case shall be of the same material as the bottom of the case. The cover shall be completely removable and shall contain directions for assembly of the training gages for use with the resuscitation training manikin.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:43Z
Check the source to verify that this is the currentversion before use.

3.5.2 Eye overlay. An eye overlay shall be furnished with each unit to acquaint the trainee with the appearance of a dilated eye, as would be found in a patient within one minute after cessation of circulation. One eye shall be dilated and the other normal. The overlay shall be fabricated of the same plastic and plasticizers used in the skin of the face mask and shall fit smoothly over the eye region of the face mask. The eyes shall be covered with eyelids which shall have eye lashes. Instructions shall be included with the eye overlay concerning the correct method of opening a closed eye lid without damage to the lid or eye. The eye overlay shall be contained in a suitable paper board box.

3.5.3 Plastic sheets. The resuscitation training manikin unit shall be furnished with 300 transparent plastic sheets, each being 5 inches in diameter and (0.025 millimeters) thick. The plastic sheets shall be suitable for forming by use of the forming apparatus (see 3.5.4). The formed sheets shall fit smoothly around the nose and over the lips of the face mask to prevent the trainee from coming in contact with the mouth and nose area of the face mask during resuscitation training, thus eliminating the need for cleaning the face mask after each trainee.

3.5.4 Forming apparatus. A forming apparatus, suitable for the forming of disposable sanitary face pieces, shall be furnished with each unit. The apparatus shall be of rugged molded plastic construction consisting of a base and hand operated pressure cover, cutting guide and piercing probe. The base and pressure cover shall form the plastic sheet (see 3.5.3) such to enable the sheets to fit smoothly about the mouth and nose area of the face mask. The cutting guide shall have cutouts for the nose and mouth of the face mask and shall be ridged or textured on the back surface. The cutting guide shall be suitable for securing the formed plastic sheet on the face mask to enable use of the piercing probe. The piercing probe shall be pointed at one end for piercing the formed plastic sheet to provide nose openings for mouth-to-nose resuscitation. The opposite end of the probe shall have a tapered cutting edge for use in cutting a mouth opening for mouth-to-mouth resuscitation.

3.5.5 Air pump. An air pump shall be furnished with each unit for use in inflating the body. The pump shall be of rubber and plastic construction with a molded plastic tapered connector which shall fit the pump receptacle on the side of the manikin body. The pump shall quickly return to initial height after a depression force has been applied and then released.

3.5.6 Replacement non-return valve. The resuscitation training unit shall be furnished with a replacement non-return valve. The valve shall be the same as the non-return valve specified in 3.3.2. The replacement valve shall be contained in a molded plastic container having a 27/32 inch inner diameter with an inside depth of not less than 1-3/4 inches and an overall length of approximately 2-3/8 inches.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:43Z
Check the source to verify that this is the current version before use.

3.5.7 Return air valve. A return air valve shall be furnished with the resuscitation training unit to be inserted in the shoulder assembly when the venting of the lung air out of the mouth is desired. The return valve shall be molded of plastic and shall be 2-3/4 inches long with a 3/4 inch inner diameter. The outer diameter shall be such that the return valve shall fit tightly into the valve receptacle of the shoulder assembly.

3.5.8 Repair kit. A repair kit shall be furnished for repairing the inflatable body material. The kit shall consist of a tube of adhesive, two strips of patch material, instructions for patching, and a paper board container. The patch material shall be the same as that used in the fabrication of the body. Each strip shall be approximately 4 inches by 7 inches. The paper board container shall contain external marking identifying it as a repair kit.

3.5.9 Separate head section. Each training unit shall be furnished with a separate head section for use in demonstrating how air passages can be obstructed or opened. The head section shall be fabricated of plastic laminations and shall be 13 inches in height and 9 inches wide. The head section shall have a movable chin and skull and shall be equipped with an interlocking cross piece for use as a stand.

3.5.10 Funnel. A funnel shall be furnished, suitable for use in the cleaning of the manikin. The funnel shall be fabricated of translucent plastic and shall be capable of being depressed from any side (at room temperature) without showing signs of cracking. The funnel shall be approximately 4-1/2 inches in height, 4 inches in diameter at the large end and 1/2 inch in diameter at the small end. The spout shall be not less than 2 inches long and shall contain longitudinal ribs for added strength. A tab shall be provided at the top edge of the funnel. The tab shall be an integral part of the funnel and shall have a 1/8 inch hole at its center.

3.6 Dimensions. All dimensions specified herein shall have the following tolerances: up to 1 inch ± 1/32 inch, up to 6 inches ± 1/8 inch, over 6 inches ± 1/2 inch. Compliance with these tolerances shall not release the contractor from providing air tight fits where required.

3.7 Weight. The weight of the resuscitation training manikin, accessories and carrying case shall not excee 40 pounds.

3.8 Identification marking. The carrying case shall be permanently and legibly marked with the following information:

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:43Z
Check the source to verify that this is the current version before use.

Manufactuerer's name or registered trademark
Federal Stock number
Contract or purchase order number
The letters U.S.

Markings secured by adhesives shall not be used.

3.9 Service data. The contractor shall furnish, with each unit, two copies of service data containing complete instructions for assembly, operation, maintenance, and a list of component parts. The service data shall include such detailed information which shall permit easy assembly and maintenance of the item. All information contained in the manual shall be suitably indexed to permit ready reference to the required information. Drawings and descriptive material shall include adequate information so that the replacement parts can be ordered by the manufacturer's part number, and without confusion as to the item identification. Identifcal or interchangeable replacement parts shall be identified by a single part number only. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the contracting officer, via the cognizant Government inspector, two preliminary copies of his proposed service data, for review and comment (see 6.1). The preliminary service data shall be submitted sufficiently early to permit adequate review (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of required revisions and corrections in the final service data without delaying delivery under the contract or purchase order. One set of the service data, as indicated above, shall be placed in a sealed, waterproof envelope, and affixed securely to the outside of the exterior container. The envelope shall be secured to the exterior container by use of metal ties, nails, or similar means, and shall be located at the end of the exterior container, preferably, when space permits, at the end bearing the markings required by MIL-STD-129. In cases where complete service data is too bulky to be conveniently attached to the outside of the exterior container, a note shall be made on the exterior envelope which shall indicate where the service data shall be located. In addition, the contractor shall mail 15 copies of the approved service data to the contracting officer for distribution. If the supplier has furnished acceptable annuals for the identical item within two years preceding the date of contract, and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request.

3.10 Workmanship. Workmanship shall be first class throughout. The training manikin, and accessories, shall be free of defects which detract from their appearance of may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:43Z
Check the source to verify that this is the currentversion before use.

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, or auditory, investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

	Inspection Level	AQL (defects per hundred units)
Major defects	II	1.0
Total (major and minor defects)	II	4.0

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105.

4.3 Examination. The resuscitation training manikin and accessories shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects:

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:43Z
Check the source to verify that this is the current version before use.

Categories and defects*

Major

101. Any component part missing, damaged, or otherwise unusable.
102. Any moveable part does not function smoothly without binding.
103. Mating parts of any connection do not fit together easily for a tight fit.
104. Service data missing or not complete.
105. The carrying case handle, locking catches or hinges ot securely attached to the case.
106. Painted areas not free of blisterd, scratched, or peeling areas.

Minor

201. Identification marking not legible, permanent or complete.
202. Surface of the carrying case not free of roughness, mars, scratches or cuts.
203. Any component not free of stain marks, grease, or otherwise unclean.

*Examination shall not be restricted to the classified possible defects listed above.

4.4 Tests

4.4.1 Leakage test. The inflatable body shall be tested to determine compliance with the requirements specified in 3.3.1. After removal of attachments the inflated body shall be completely submerged in water and examined for emission of air bubbles.

4.4.2 Cleaning test. The resuscitation training manikin shall be tested to determine the capability of the components to withstand cleaning by ordinary methods. The entire manikin shall be washed with a mild soap and water solution. Special emphasis shall be given to cleaning the mouth and nose region of the face mask. The cotton jersey clothes shall be washed and dryed in the same manner as gym clothes.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One manikin, as specified, constitutes one unit of issue.

5.1.2 Level A.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:43Z
Check the source to verify that this is the currentversion before use.

5.1.2.1 Unit package. Each unit shall be packaged in a carrying case as specified in 3.3.8.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Each unit, in carrying case as specified in 5.1.2.1, shall be packed in an exterior container constructed in accordance with PPP-B-636, compliance symbol V3C. Interior fittings shall be provided to prevent movement of carrying case. Closure shall be in accordance with appendix of specification, and waterproofing shall conform with 30.4, of PPP-B-636. Case liner shall not be required.

5.2.1.2 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Each unit, in carrying case as specified in 5.1.2.1 shall be packed in an exterior container constructed in accordance with PPP-B-636, type CF, class domestic, fabricated from not less than 275 pound test fiberboard. Interior fittings shall be provided to prevent movement of carrying case. Method I closure shall be utilized.

5.3 Marking.

5.3.1 Unit package. Each carrying case shall be marked as specified in 3.8.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Where service data is to be sent.
(c) Selection of applicable levels of packing (see 5.1 and 5.2).
(d) When First Article is required (see 3.1)
(e) Where First Article shall be sent (see 3.1)

6.2 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:43Z
Check the source to verify that this is the current version before use.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Category N

Preparing activity:
Defense Supply Agency - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

*U.S. GOVERNMENT PRINTING OFFICE: 1987-251-515/3607

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:43Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements.

SPECIFICATION
MIL-M-36510

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?
   ☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:43Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
AUG 66

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:43Z
Check the source to verify that this is the current version before use.