MIL-M-36552A
12 April 1976
SUPERSEDING
MIL-M-36552
15 January 1968

MILITARY SPECIFICATION

MIRROR, MOUTH EXAMINING, PLASTIC, PLANE GLASS, DISPOSABLE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a mouth examining mirror suitable for patient use in preventive dentistry programs.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Boxes, Shipping, Fiberboard.
PPP-B-676	Boxes, Setup.

Military

MIL-L-10547	Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.
MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)
23170 Dental Mirror, Disposable

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:47Z
Check the source to verify that this is the current version before use.

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

D256	Impact Resistance of Plastics and Electrical Insulating Materials
D638	Tensile Properties of Plastics

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

3. REQUIREMENTS

3.1 Material. Material shall be as hereinafter specified, first class throughout, and suitable in all respects for the intended purpose.

3.1.1 Glass. Glass for the mirror shall be of good quality polished glass. Composition of glass shall be of such quality as to withstand prolonged use and repeated soap and water washings without dulling of surface, when tested as specified in 4.4.2. The back of the glass shall receive a coating of chemically deposited silver, protected by a film of electrolytically deposited copper. The copper surface shall be provided with a heavy coating of water-resistant enamel. The glass, silver, copper and enamel interfaces shall be firmly adherent. The front and back surfaces of the glass shall be flat and parallel. The thickness of the coated glass shall be 0.045 to 0.075 inch.

3.1.2 Plastics. Plastics used for the mirror frame and handle shall be cellulose acetate, polystyrene, or acrylonitrile-butadiene-styrene (ABS) with a minimum average Izod impact strength of 1.5 foot-pounds per inch of notch, and a minimum average tensile strength of 4000 pounds per square inch, when tested as specified in 4.4.3.

3.2 Design and construction. The mirror shall consist of a reflective glass disc securely encased in a one-piece molded plastic frame and handle. The reflective disc shall be plane and shall provide a clear, undistorted image when tested as specified in 4.4.1, in order for a patient to interpret results of flossing and brushing stained tooth areas under normal, unintensified illumination. The reflective disc shall be firmly secured within

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.

the mirror frame. The mirror shall show no signs of deterioration or impairment of function when tested as specified in 4.4.2. The mirror shall be in accordance with Defense Personnel Support Center (DPSC) Drawing 23170.

3.3 Finish. The mirror face shall have a smooth reflective surface free from pits, scratches, cracks, ripples or other surface imperfections affecting reflectivity. The mirror frame and handle shall have a smooth finish free from sharp edges or projections.

3.4 Identification marking. The name or registered trademark of the manufacturer or contractor shall be legibly and permanently marked on the handle or back of the mirror frame. In addition, mirrors of foreign (other than U.S.) manufacture shall be marked with the name of the country where manufactured.

3.5 Workmanship. Workmanship shall be first class throughout. The mirror shall be free from defects which may affect proper functioning, appearance or serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of compliance. When available, certificates of compliance supplied by the manufacturer of the component or raw material may

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.

be furnished in lieu of actual performance of such testing by the contractor. When certificates of compliance are submitted, the Government reserves the right to check test such items to determine the validity of the certification.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The unit of product for examination shall be the individual mirror. Inspection and acceptable quality levels (AQL's) shall be as indicated in table I.

Table I. Inspection and Acceptable Quality Levels for Examinations

Type defect	Inspection level	AQL (Percent defective)
Major A	II	1.0
Major B	S-2	1.5
Minor A	II	2.5
Minor B	S-2	4.0

4.2.2 For end item testing. Sampling for end item testing shall be conducted in accordance with MIL-STD-105 and table II. The unit of product for testing shall be the individual mirror.

Table II. Inspection and Acceptable Quality Levels for Tests

Characteristic	Requirement	Test Procedure	Inspection level	AQL (percent defective)
Reflection	3.2	4.4.1	S-2	1.0
Washing	3.2	4.4.2	S-2	2.5

4.2.3 For raw materials testing. In the absence of certification (see 4.1.3), a sufficient quantity of raw material shall be sampled for testing from each lot of raw material used.

4.3 Examination. Mirrors shall be examined for defects, including, but not limited to, those listed in table III.

Table III. Classification of Defects

Categories	Defects
Major A:	
101	Mirror not of design and construction specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.

Table III. Classification of Defects (con't)

102	Mirror glass not free from workmanship defects such as scratches, cracks, chips, seeds, bubbles, or breaks in reflective coating.
103	Mirror glass not firmly secured in the plastic head.
104	Coatings (see 3.1.1) not present on back of mirror glass. 1/
105	Glass for mirrors not plane.
106	Plastic frame shows workmanship defects such as cracks, chips, sharp edges, or flash.
Major B:	
151	Thickness of mirror glass not as specified 1/
152	Dimensions of mirror frame not as specified on drawing.
153	Mirror frame not angled as shown on drawing.
Minor A:	
201	Plastic does not have a smooth finish.
202	Handle shows workmanship defects such as cracks, chips, sharp edges, or flash.
203	Identification marking not as specified.
Minor B:	
251	Dimensions of handle not as specified on drawing.

1/ In process examination may be used for this defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.

4.3.1 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Table IV. Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging & Packing Materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

4.4 Tests.

4.4.1 Reflection test. This test consists of viewing the image formed by the straight edge of a steel ruler as the image scans the mirror face. The following procedure shall be followed: Place the ruler on a flat surface with the straight edge away from the viewer. Locate the mirror between 2 and 2 1/4 inches from the straight edge, and support it in a vertical position resting the mirror edge on the flat surface. Slowly pivot the mirror about its edge forward and backward to cause the reflected image to vertically scan the mirror face. The image shall be examined to determine compliance with 3.2.

4.4.2 Soap and water washings. Mirrors shall be thoroughly cleansed with soap and water, and then rinsed thoroughly with distilled water. The procedure shall be repeated five times. Mirrors shall then be examined for compliance with requirements specified in 3.1.1 and 3.2.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.

4.4.3 Testing of plastics. Strength tests performed on the raw plastic material shall be conducted in accordance with ASTM D256 Method A for impact strength and ASTM D638 for tensile strength at break.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.1).

5.1.1 Unit of issue. One box containing seventy-two (72) mirrors, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Unit package. Seventy two mirrors shall be packaged in a snug fitting box constructed in accordance with PPP-B-566 or PPP-B-676. Closure of the box shall be adequate to prevent accidental opening.

5.1.2.2 Intermediate package. Ten unit packages shall be packaged in a snug fitting box constructed in accordance with PPP-B-636, class domestic. Closure shall be adequate to prevent accidental opening.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.2.1 Level A. Four intermediate packages (40 units) shall be packed in a box designed for a type 2 load, and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2 or PPP-B-636, type CF, class weather-resistant. Closure and strapping shall be in accordance with the appendix of the applicable box specification.

5.2.1.1 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard boxes shall be waterproofed in accordance with PPP-B-636.

5.2.2 Level B. Four intermediate packages (40 units) shall be packed in a box designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 Specification changes. Because of the extensiveness of changes to this document, the margins have not been marked with an asterisk to indicate where changes from the previous issue have been made.

6.3 National Stock Number (NSN) Coverage. This specification covers the following item appearing in the Federal Supply Catalog:

NSN	Item Identification
6520-00-935-4135	MIRROR, MOUTH EXAMINING, Plastic, Plane Glass, Disposable, 72s

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project Number 6520-1399

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS. The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers, vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-M-36552A MIRROR, MOUTH EXAMINING, PLASTIC, PLANE GLASS, DISPOSABLE

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

   A. GIVE PARAGRAPH NUMBER AND WORDING

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.