MIL-S-36524A
27 September 1978
SUPERSEDING
MIL-S-36524
31 January 1967

MILITARY SPECIFICATION

SPLINT, WRAP AROUND

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a combination leg, arm, back and neck splint.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

V-T-276	- Thread, Cotton
AA-C-295	- Chair, Rotary, Steel Tubular, Special Purpose.
JJ-W-155	- Webbing, Textile, Cotton Elastic.
QQ-S-763	- Steel Bars, Shapes, and Forgings, Corrosion Resisting.
CCC-C-419	- Cloth, Duck, Cotton, Unbleached, Plied Yarns, Army and Numbered.
CCC-D-950	- Dyeing and Aftertreating Processes for Cotton Cloths.
PPP-B-585	- Boxes, Wood, Wirebound.
PPP-B-601	- Boxes, Wood, Cleated-Plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	- Boxes, Fiberboard.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-ATT, 2800 So. 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-F-495	- Finish, Chemical, Black, for Copper Alloys.
MIL-W-530	- Webbing and Tape, Textile, Cotton, General Purpose, Natural or in Colors.
MIL-B-543	- Buckles, Tongueless and Web Strap.
MIL-T-3530	- Thread and Twine, Mildew Resistant or Water Repellent Treated.
MIL-R-5001	- Rubber, Latex Foam Sponge.
MIL-W-5664	- Webbing, Textile, Elastic, Cotton.
MIL-L-10547	- Liners, Case, and Sheet, Overwrap, Water-Vaporproof or Waterproof, Flexible.
MIL-G-16491	- Grommet, Metallic.

STANDARDS

FEDERAL

FED-STD-151	- Metal, Test Methods.
FED-STD-191	- Textile Test Methods.
FED-STD-751	- Stitches, Seams and Stitchings.

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.
MS 27980	- Fasteners, Snap, Style 2.
MS 27983	- Fasteners, Snap, Style 4.
MS 35425	- Nut, Plain, Wing.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC).

20219	- Splint, Wrap Around, Leg, Arm, Back and Neck, Canvas.

(Copies of specifications, standards, drawing and publications required by contractors in connection with specific procurement functions should be obtained from from the procuring activity or as directed by the contracting officer).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

Color Association of the United States, Incorporated.

United States Army Standard Color Card for the Official Standardized Shades of Sewing Threads.

(Application for copies should be addressed to the Color Association of the United States, Incorporated, 200 Madison Avenue, New York, NY 10016).

3. REQUIREMENTS

3.1 Component parts. The wrap around splint consists of a padded wrap around leg splint, a removable, adjustable traction bar, a padded extension section, and a padded cuff.

* 3.2 Materials. Except as specified herein, materials shall conform to the applicable portions of the specifications, standards and drawing as specified herein. When materials to be used are not specifically designated, those used shall be of the highest commercial quality, and entirely suitable for the purpose intended.

3.2.1 Cloth, cotton duck. The cloth, before finishing, shall be cotton duck conforming to type III, of CCC-C-419, 12.29 ounces per square yard minimum.

3.2.1.1 Color and treatment. The cotton duck shall be vat dyed Olive Drab No. 7, water-repellent and mildew-resistant, conforming to type I, class B, of CCC-D-950, except that the cloth shall be made mildew-resistant by treatment limited to 2, 2' methylene-bis (4-chlorophenol).

* 3.2.2 Webbing, cotton. Webbing shall be Olive Drab No. 7, class 4 or 7, MIL-W-530, except when class 4 webbing is used, maximum total copper shall be 0.23 percent. The webbing attached to the proximal end of the splint shall be type IIb, one inch wide, class 4 or 7 of MIL-W-530 and the webbing for the cuff shall be type IIb, 1/2 inch wide, class 4 or class 7. Raw ends of webbing shall be dipped in a vinyl resin to prevent raveling and shall be tested as specified (see 4.4.1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

3.2.3 Thread. Except as specified herein, the thread shall conform to V-T-276, and shall be type IB3, machine thread, 4-ply, ticket no. 12, table II for needle thread and type IA3, machine thread, 4-ply, ticket no. 12, table I for the bobbin thread. The thread shall be vat dyed olive drab shade S-1, conforming to U.S. Army Standard Color Card. Thread shall be treated for mildew resistance in accordance with MIL-T-3530, using type I, class 1 mildew inhibiting agent. The finished thread shall show fair colorfastness to laundering, chlorine bleaching, and weather.

3.2.4 Webbing, elastic. Elastic webbing for the securing straps shall conform to class 1 of MIL-W-5664. The webbing shall be one inch wide. The color shall be dyed olive drab no. 7 to match the color of the splint proper. The webbing shall be water repellent and mildew resistant in accordance with JJ-W-155.

3.2.5 Lumber. Lumber for the stays shall be long leaf yellow pine, white oak, beech, birch, maple, or ash. Lumber shall be clear, straight grain, and shall be edge grain cut. Lumber shall be seasoned to a moisture content at the time of fabrication; not to exceed 12 percent. It shall be free from knots, splints, decay, birdpicks, checks, and any other defects affecting its appearance or which may affect serviceability. Moisture content shall be tested as specified (see 4.4.2).

3.2.6 Fasteners, snap.

* 3.2.6.1 Fasteners for elastic webbing straps. Snap fasteners for the elastic webbing straps shall consist of -1B button, -6B socket, -7B stud, and -8B eyelet of MS 27980. Material shall be brass with a black chemical finish conforming to MIL-F-495. The male elements shall be for attachment to the straps; the female elements shall be for attachment to the splint as shown in Drawing 20219.

* 3.2.6.2 Fasteners for the padded extension. Snap fasteners for the padded extension shall consist of -6 cap, -7 socket, -8 stud, and -9 post, or -1 button, -2 socket, -3 stud of MS 27983 and -8B eyelet of MS 27980. The male elements shall be for attachment to the splint. The female elements shall be for attachment to the padded extension as shown in the drawing.

* 3.2.7 Buckles, web strap. Buckles for the cotton webbing strap shall conform to type III, style 3, construction A, class 1 or class 3, one inch size of MIL-B-543 and shall accept one inch wide webbing (see 3.2.2). The buckles shall be given a black finish as specified in MIL-B-543.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:44Z
Check the source to verify that this is the currentversion before use.

3.2.8 Grommets. Grommets shall conform to type I, class 3, size no. 4 of MIL-G-16491.

3.2.9 Traction bar clamp assembly.

3.2.9.1 Bolts, steel. Bolts for the traction bar clamps shall be No. 10 bolt, 24 threads per inch, National Coarse Series, 1-1/16 ± 1/32 inch long shank. They shall be carriage bolts, square neck, with unslotted roundhead. The thread length shall be sufficient to enable the wingnut (see 3.2.9.2) to be tightened securely to hold the traction bar in place when used as intended. Head diameter shall be 7/16 ± 1/64 inch. The bolts shall be made of carbon steel, commercial grade machine screw stock and shall be given a black chemical finish.

3.2.9.2 Wingnuts. Wingnuts for the traction bar clamp bolts shall be steel and shall conform to MS 35425-37, undrilled. They shall be given a black chemical finish.

3.2.9.3 Clamps. Clamps shall be fabricated of phosphor bronze, 4X hard, 0.0329 inch (No. 21, United States Standard Gage) thick of a temper suitable for the purpose intended. Commercial tolerances shall apply. The temper shall be such as to permit the clamp to securely hold the traction bar (see 3.2.10) in place when the wingnut (see 3.2.9.2) is tightened and to allow release of the traction bar when the wingnut is loosened without a permanent setting of the clamp.

3.2.10 Traction bar. Steel for the traction bar shall be cold-finished corrosion-resisting steel, class 302, 303, or 304, conforming to the chemical requirements of QQ-S-763. Steel shall be a 1/4 inch diameter round bar. The dimensional tolerances shall be as specified in QQ-S-763. The traction bar shall be given a lustreless black chemical finish.

3.2.11 Wire mesh. Wire mesh for the padded extension section and the padded cuff shall be made from soft annealed low carbon steel wire, 0.041 inch diameter before galvanizing, 5 inches ± 1/4 inch wide. Commercial tolerances shall apply. The wire mesh shall be woven with a uniform two by two square mesh, two wires to the inch. The edges of the long borders shall be made of a single or double wire. The wire mesh shall be free of protruding wires which may puncture the padding and duck. The wire mesh shall be heavily galvanized with molten zinc by the hot-dipped process so that all joints are securely locked.

3.2.12. Padding. Padding shall be either bound hair, foam rubber or polyurethane foam.

3.2.12.1 Hair, bound. Bound hair shall be 15 percent cattle-tail hair and 85 percent hog hair, with a binder of polychloroprene latex, natural rubber latex, butadiene-styrene latex, or a mixture of natural rubber latex and butadiene-styrene latex. Density shall be 0.0020 to 0.0025 pounds per cubic inch.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

3.2.12.2 Foam rubber. The foam rubber padding shall be type II, soft class, grade B or C of MIL-R-5001.

3.2.12.3 Polyurethane foam. Polyurethane foam shall conform to AA-C-295.

3.2.13 Adhesive. Adhesive for cementing the slabs of foam polyurethane or rubber shall be capable of cementing the layers of foam material together as required for the padded extension section and the cuff so that the joint shall be stronger than the foam material itself. On separating the layers, the foam material shall fail before the joint becomes spearated.

3.3 Construction. Construction shall meet the requirements as specified herein, and as shown in Drawing 20219.

3.3.1 Stitching. All stitching shall conform to type 301, of FED-STD-751, eight to ten stitches per inch. Thread breaks shall be overstitched not less than one inch beyond ends of each break. Stitching shall be backstitched not less than one inch at ends, except where ends are turned under in a hem or held down by other stitching. Thread tension shall be maintained so that there will be no loose stitching and that the lock will be embedded in the stitched materials. All thread ends shall be trimmed.

3.3.2 Attachment of grommets. Grommets shall be tightly clinched without cutting the duck. The hole punched in the duck and padding shall be smaller than the outside diameter of the grommet barrel so that the barrel must be forced through the hole.

3.3.3 Attachment of elastic webbing. In the construction of the splint, a double thickness of elastic webbing shall be used as shown in the drawing. The fastener elements shall include both thicknesses of the elastic webbing where attached.

3.3.4 Attachment of fastener elements. Holes shall be punched in the fabric and padding before insertion of the tube of the fastener elements. Holes shall be small enough to require tubes to be forced through the material. Fastener elements shall be tightly clinched, without cutting the duck, and shall not become detached from the fabric in normal use.

3.3.5 Attachment of clamp assembly. The bolts (see 3.2.9.1) shall be located as shown in the drawing and shall be centered between the sides of the wood stays. Holes shall be punched in the duck and padding to permit the bolts to be inserted without tearing the duck. Holes shall be 1/32 larger than the diameter of the bolts. The end of the bolts shall be peened over sufficiently to prevent loss of the wingnuts.

3.3.6 Traction bar. The traction bar (see 3.2.10) shall be fabricated as shown in the drawing. The bottom of the "U" of the bar shall have a bend to prevent lateral slipping of a cravat bandage or other securing device.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

* 3.3.7 Foam padding and wire mesh of the extension section and the cuff. The wire mesh (see 3.2.11) shall be sandwiched between the two layers of foam material when they are cemented together as shown in the drawing. The foam material shall extend beyond the sides and ends of the wire mesh. Thickness of the foam padding for the padding extension and padded cuff shall be 1/4 inch.

* 3.3.8 Padding for splint proper. Padding for the splint proper (see 3.2.12) shall be a single piece and shall be secured in the stitching of the splint and in the attachment of the hardware to the splint proper, as shown in the drawing. Bound hair padding thickness under no load shall not be less than one inch. Foam padding shall be not less than 7/16 inch thick.

3.4 Marking for identification. The three parts of each wrap around splint shall be legibly and permanently marked in black as shown in the drawing.

3.5 Workmanship. Workmanship shall be first class throughout. The finished wrap around splint shall be clean, well made, and free from defects which detracts from its appearance or may impair its serviceabilitiy.

3.6 Operating instructions manual. The contractor shall supply with each complete splint an operating instruction manual. The manual shall consist of four pages and shall be printed on good quality paper. Page size shall be 8 by 10 inches. Manual shall be placed in a waterproof envelope of approrpiate size. Negatives of the instruction manual will be furnished by the Government to the successful bidder, to be returned to the Government upon completion of contract. The envelope shall be prominently marked: "Operating Instruction Manual".

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:44Z
Check the source to verify that this is the currentversion before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the splint materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one wrap around splint.

TABLE I. Sampling for examination.

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	2.5
Minor defects	II	4.0
For dimensional examination	S-1	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II. The inspection level shall be S-2 and the acceptance number shall be zero. Unit of product shall be item shown in table II.

TABLE II. Sampling for tests.

Component	Characteristic	Requirement	Test procedure
Splint	Webbing end finish test	3.2.2	4.4.1
Stays	Moisture content of wood	3.2.5	4.4.2

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

4.3 Examination. The wrap around splint shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table III. Examination shall not be restricted to the classified possible defects listed below.

TABLE III. Classification of defects.

Categories	Defects
Major	
101	Any hole, cut, tear, weak place, open place, broken or missing yarn in cloth or webbing.
102	Webbing frayed, scalloped edges or not tightly woven.
103	Webbing ends not dipped to prevent fraying.
104	Hardware broken, malformed, failing to serve intended purpose, burs, sharp edges or corroded edges.
105	Hardware finish partly or totally omitted, incomplete, misplaced or insecurely attached.
106	Lumber contains knots, splinters, decay or other defects.
107	One or more broken stitches joining seam, two or more skipped stitches or runoffs.
108	Wrong seam type, loose seam, or seam too tight.
109	Stitching ends not backstitched or secured as specified.
Minor	
201	Raw edge of cloth not secured as shown in drawing.
202	One or more stitches less than specified or one or more stitches greater than specified.
203	Thread breaks overstitched less than one inch beyond each end of break.
204	Sling not free of dirt, grease, stains or foreign material.
205	Marking omitted, misplaced, illegible, incorrect or size or characters not as specified.

4.3.2 Packaging inspection. The inspection of the unit packaging, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and marking requirements as specified in MIL-STD-129.

4.3.3 Dimensional examination. The sling shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Webbing end finish test. The vinyl resin on the raw ends of web straps shall not become tacky after being conditioned for four hours at 180° Fahrenheit (F.), shall not crack or peel off when bent over a four inch mandrel at minus 60°F. (after webbing has been at minus 60°F for one hour), and shall not swell up or become gelatinous after being conditioned for four hours in an atmosphere of 100 percent relative humidity at 100°F. In addition, after being subjected to the accelerated aging in accordance with method 5852 of FED-STD-191, the vinyl-resin-dipped end of the strap shall be capable of passing the tests described above.

4.4.2 Moisture content of wood. Wood specimen shall be weighed initially. Specimen shall be placed in a drying oven and dried at a temperature of 100 to 105°C. to a constant weight. Oven air circulation about piece should not be restricted. Test specimen should be weighed immediately after removal from oven and moisture percentage calculated in accordance with the formula shown below.

$$\% \text{ moisture content} = \frac{(\text{Original weight} - \text{oven dried weight}) \times 100}{\text{Oven dried weight}}$$

As an alternate, moisture content may be determined by a moisture meter calibrated to Bureau of Standards requirement.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C, as specified (see 6.2).

5.1.1 Unit of issue. Each splint constitutes a unit of issue.

5.1.2 Level A.

5.1.2.1 Unit packaging. Each splint shall be clean, neatly folded and wrapped in kraft paper or other suitable material. The wrapped splint shall be further packaged in a box conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.1.3 Level C. Each splint shall be clean and packaged in a manner that will afford adequate protection against deterioration and physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.3).

5.2.1 Level A. Four unit packages shall be packaged in a shipping container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:44Z
Check the source to verify that this is the currentversion before use.

5.2.2 Level B. Four unit packages shall be packed in a shipping container designed for a type 2 load and conforming to PPP- B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The packaged splints shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the carrier's rules and regulations.

5.3 Marking. Unit packages and shipping containers shall be marked as specified in MIL-STD-129.

6. NOTES

6.1 Intended use. The splints are intended for use in emergency and first aid treatment to immobilize the leg, arm, back or neck.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) National Stock Number (NSN).
(c) Selection of applicable level of packaging and packing required. (see 5.1 and 5.2)

6.3 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6515-00-680-0887	SPLINT, WRAP AROUND, Leg, Arm, Back and Neck, Canvas.

6.4 This specification does not include all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover one type which is normally procured to meet military requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:44Z
Check the source to verify that this is the currenversion before use.

* 6.5 Changes from previous issue. The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Logistics Agency - DM

Project Number 6515-2172

Review activity:
DoD - MB

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS: This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.

NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER *(Number)* AND TITLE

SPLINT, WRAP AROUND MIL-S-36524A

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1. ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY *(Printed or typed name and address – Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 OCT 76 1426 EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:44Z
Check the source to verify that this is the current version before use.

FOLD

Defense Personnel Support Center
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL
2800 SOUTH 20TH STREET
PHILADELPHIA, PENNSYLVANIA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:44Z
Check the source to verify that this is the current version before use.