MIL-T-0036527 (DSA-DM)
14 February 1967

MILITARY SPECIFICATION

THERMOMETER, SELF INDICATING, LIQUID IN GLASS,
Film Processing

1. SCOPE.

1.1 Scope. This specification covers one type of liquid in glass, floating, tank-type thermometer, suitable for use in photographic chemical solutions.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up, Paperboard.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

FSC 6685

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:45Z
Check the source to verify that this is the current version before use.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Materials.

3.1.1 Glass. The registering column shall be fabricated from lens tubing, and the thermometer shall be resistant to acid and alkali solutions as used in photo-processing.

3.1.2 Thermometric liquid. The thermometric liquid shall be blue or black colored alcohol.

3.1.3 Ballast. The ballast shall be lead shot.

3.1.4 Cement. The cement used to secure the ballast and the scale shall show no evidence of softening when tested as specified in 4.4.1.

3.2 Design. The thermometer shall be cylindrical in shape. A printed paperboard scale shall be securely affixed to the registering column, and the assembly shall be securely mounted inside the body tube. The thermometer shall be weighted with sufficient ballast so that it shall float in a vertical position with the top portion 1-1/2 ± 1/2 inches (in.) above the surface of the solution (specific gravity 1.1). The ballast shall be sealed dry and immobile by means of a fused partition or a suitable cement. (see 3.1.4)

3.3 Dimensions. The thermometer shall have an overall length of 8 ± 1/4 inch, and a diameter of 9/16 ± 1/16 inch.

3.4 Scale. The thermometer shall have a minimum temperature range of 30° to 100° F. The graduations shall be in increments of 1° F. throughout the entire scale. Each 10 degree mark shall be longer than the intervening marks and shall be clearly and appropriately numbered.

3.5 Special marking. Each thermometer shall have a special marking in red on the scale to indicate distinctly the scale position of 68° F.

Source: https://assist.dla.mil -- Downloaded: 20261021T08:45Z
Check the source to verify that this is the current version before use.

3.6 Accuracy. The thermometer shall be accurate to within plus or minus one degree Fahrenheit at all temperatures within the range of the instrument. The thermometer shall be tested as specified in 4.4.2.

3.7 Identification marking. Each thermometer shall be legibly and permanently marked with the thermometer range and the name or registered trademark of the manufacturer.

3.8 Workmanship. Workmanship shall be first class throughout. Thermometers shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for Inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to insure that supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:45Z
Check the source to verify that this is the currentversion before use.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

	Inspection level	AQL (defects per 100 units)
For visual examination		
Major defects	II	1.0
Total (major and minor combined) defects	II	4.0
For dimensional examination	S-2	0.65

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105.

Table I.

Characteristic	Requirement	Test Procedure	Inspection level*
Cement	3.1.3	4.4.1	S-1
Accuracy	3.6	4.4.2	S-3

* Acceptance number shall be zero.

4.3 Examination. The thermometers shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects.

Categories and defects*

Major
101 Thermometer glass not free of bubbles, stones and cords.
102 Thermometer glass not free of cracks and checks.
103 Thermometer glass not clear.
104 Separation of thermometer liquid.
105 Major graduations not longer than minor graduations.
106 Ballast improperly sealed.
107 Scale card not free of wrinkles, warping and delamination.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:45Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements.

SPECIFICATION: MIL-T-0036527 (DSA-DM)

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426

REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:45Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
AUG 66

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:45Z
Check the source to verify that this is the current version before use.