MIL-T-36522B
26 August 1976
SUPERSEDING
MIL-T-36522A
29 May 1974

MILITARY SPECIFICATION

TABLE, ANESTHETIST'S

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a mobile, four-drawer table for use as an anesthetist's stand.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-601 - Boxes, Wood, Cleated Plywood.
PPP-B-640 - Boxes, Fiberboard, Corrugated, Triple-wall.

MILITARY

MIL-P-116 - Preservation-packaging methods Of.
MIL-A-8625 - Anodic Coatings for Aluminum and Aluminum Alloys.
MIL-L-10547 - Liners, Case, and Sheet, Overwrap, Water-Vaporproof or Waterproof, Flexible.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, Code ATT, 2800 South 20th Street, Philadelphia, Pa. 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

STANDARDS

FEDERAL

FED-STD-151 - Metals Test Methods.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

25002 - Anesthetist's Table

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply

AMERICAN IRON AND STEEL INSTITUTE (AISI)

AISI Steel Products Manual - "Stainless and Heat Resistant Steels."

(Application for copies should be addressed to the American Iron and Steel Institute, 1000 16th St., N.W. Washington, DC 20036.)

NATIONAL FIRE PROTECTION ASSOCIATION (NFPA)

No. 56A Standard for the Use of Inhalation Anesthetics

(Application for copies should be addressed to the National Fire Protection Association, 470 Atlantic Ave., Boston, MA 02210.)

3. REQUIREMENTS

3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.1.2.1 and 6.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

3.2 Materials. Unless otherwise specified herein, material shall conform to DPSC Drawing No. 25002.

3.2 1 Corrosion-resisting steel. Corrosion-resisting steel shall comply with AISI type 302 or 304 as indicated in the AISI Steel Products Manual when tested as specified in 4 4.1.

3.3 Design and construction The table shall be in accordance with DPSC Drawing No. 25002 and as specified herein. As an alternate, the guardrail may be attached with screws and lockwashers. The table shall be equipped with four flush-closing drawers mounted on roller-type slides. Drawers shall move freely without binding and means shall be provided to restrain the drawers from opening when the table is moved. Drawers shall be readily removable for cleaning. Spacing around all closed drawer fronts shall be uniform, not exceeding 1/8 inch. The uniform difference between surfaces of drawer fronts and the surface of the table front shall not exceed 1/32 inch. The table shall be mounted on four conductive casters. When placed on a flat level surface, the table shall rest evenly on the casters, without rocking, and shall move freely when tested as specified in 4.4.3

3.3 1 Drawers. Drawers shall be designed to prevent inadvertent removal and shall be free of sharp corners and edges. Alternate construction to prevent drawers from being removed inadvertently may be employed, provided they function as effectively as that shown in DPSC Drawing No. 25002. The double-walled drawer fronts shall be insulated by filling with uniformly distributed glass wool or by a special mastic (see 6.5). Drawer pulls shall be corrosion-resisting steel, satin-chrome plated brass, or anodized aluminum alloy (see 3.6), with lockwashers to secure the mounting screws.

3 4 Electrical conductivity The completely assembled table shall meet the requirements of NFPA 56A.

3 5 Identification marking Each table shall be marked, permanently and legibly, with the contract or purchase order number or the name or registered trademark of the manufacturer.

3 6 Finish All exposed corrosion-resisting steel surfaces shall have a No. 4 commercial finish. Exposed carbon steel surfaces and hardware shall be cadmium or zinc plated. Chrome plating, where used, shall have a nickel undercoat. Anodizing shall be colorless and in conformance with type II of MIL-A-8625.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

3.7 Workmanship. Workmanship shall be first class throughout. The table shall be free from defects which detract from its appearance or may impair its serviceability. All burrs and sharp edges shall be removed from the table.

3.7.1 Welding. Welding shall be durable, sound in every detail and in accordance with the latest practices recommended by the American Welding Society. All welds shall be produced by qualified welding operators with satisfactory welding equipment using satisfactory filler metal or electrode where applicable to provide the best possible weld. Surfaces to be welded shall be free from grease, oil, oxides or any other foreign matter. Any cleaning method which will remove these substances and leave the surface clean is satisfactory. All weld joints shall blend into the adjacent metal in gradual, smooth curves Beads shall be smooth and free of slag, excessive undercut, or excessive splatter. Sufficient metal shall be added to provide a suitable fillet or reinforcement unless otherwise specified. Welds shall not be oxidized. All welds shall be continuous with no visible voids, pin holes, or cracks.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government. upon the Government's request at any time, or from time to time. during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

4.1.2.1 Classification of inspection. The inspection requirements specified herein are classified as follows:

a. First article inspection.
b. Quality conformance inspection.

4.1.2.1.1 First article inspection. First article inspection shall be performed on tables when a first article sample is required (see 3.1). This inspection shall include examination of 4.3. and tests of 4.4.

4.1.2.1.2 Quality conformance inspection. Quality conformance inspection shall include the examination of 4.3 and tests of 4.4.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contractor number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I. The unit of product for sampling purposes shall be one table.

TABLE I. Sampling for examination

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Minor defects	II	4.0
For dimensional examination	S-1	2.5

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II, for an acceptance level of zero.

TABLE II. Sampling for tests

Component (Unit of product)	Characteristic	Requirement	Test procedure	Inspection level
Metal	Chemical composition	3.2.1	4.4.1	1/
Table	Electrical conductivity	3.4	4 4 2	100%
	Stability and motion	3.3	4.4 3	I

1/ See 4.2.3

4.2 3 Raw materials. A quantity of each lot of each metal used in the manufacture of the tables shall be selected for chemical analysis as specified in 4.4.1.

4.3 Examination. The tables shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects Examination shall be conducted in accordance with the classification of defects listed in table III. Examination shall not be restricted to the classified possible defects listed below

TABLE III. Classification of defects

Categories	Defects
Major	
101	Any component missing
102	Components not properly aligned nor fitted
103	Metal exhibits cracks, tears, ruptures, or deformed surfaces.
104	Any weld not sound, not free of cracks.
105	Caster cracked, broken, not securely attached.
106	Front of drawer not flush with front surface of table, spacing not uniform.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects (cont'd).

Categories	Defects
Major	
107	Drawers do not slide freely.
108	Drawers not provided with stops which prevent inadvertent removal.
109	Drawers not provided with means which prevent opening when table is moved.
110	Plated finish missing, not proper type.
111	Any exposed or accessible edge, corner, or surface has burrs, sharpness, or sharp projections.
Minor	
202	Finish not uniform, not free of coarse grind marks. scratched areas
203	Any weld not continuous, not cleaned, shows pitting, splatter, or discoloration.
204	Any visibly exposed weld not ground smooth to blend with adjacent surfaces.
205	Unit not clean has dirt, oil. grease, or metal chips
206	Identification marking not present, complete, and legible.
207	Any rubber drawer-bumper missing
208	Corrosion-resisting steel finish not No. 4.

4.3.2 Packaging inspection The inspection of the packaging shall be in accordance with the group A quality conformance inspection requirements of MIL-P-116. The inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4 3.3 Dimensional examination The tables shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4 4 Tests Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

4.4.1 Chemical composition. Four ounces from each lot of corrosion-resisting steel shall be tested in accordance with Fed. Test Method Std. 151.

4.4.2 Electrical conductivity. The completely assembled unit shall be tested using the procedure and equipment described in NFPA No. 56A. The test shall be conducted with the metal plate alternately under each caster

4.4.3 Stability and motion The table shall be placed on a flat level surface and moved in a figure-8 pattern Any tendency of the table to rock, or of the casters to balk, shall be observed.

5. PACKAGING

5.1 Packaging Packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A

5.1.1.1 Unit package. Casters shall be removed, packaged in a bag of suitable size and design, and secured to the unit Trays shall be wrapped in kraft paper and the package secured in such a manner as to prevent shifting or damage. Each unit shall be further protected by wrapping with kraft paper or other suitable material. Cushioning shall be interposed between points of contact of the container and wrapped unit. The required wrapping and cushioning material shall be secured in position. Packaging shall conform to MIL-P-116, method III.

5.1.2 Level C Tables shall be individually packaged in a manner that will afford adequate protection against corrosion, deterioration and physical damage during shipment from supply source to the first receiving activity.

5.2 Packing Packing shall be level A, B or C as specified (see 6.2)

5.2.1 Level A. Each table shall be packed in an exterior container designed for a type 3 load and constructed in accordance with PPP-B-601, overseas type. or PPP-B-640, class 2, style E. Container design shall include suitable interior fittings to prevent shifting or damage to the unit. Closure of box shall be as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

5.2.1.1 Case liner. Each level A wood box shall be lined with waterproof case liner conforming to MIL-L-10547. Closure and ·aling shall conform to applicable paragraphs of appendix thereto. se liner shall not be required for fiberboard boxes. Fiberboard ıx shall be waterproof and sealed as specified in the appendix PPP-B-640.

5.2.2 Level B. Each table shall be packed as specified in 2.1, except that boxes shall be nonweather resistant and domes- c type. Case liner for wood box and waterproofing and sealing the fiberboard box is not required.

5.2.3 Level C The packaged table shall be packed in shipping ıntaıners in a manner that will afford adequate protection against ımage during shipment from the supply source to the first receiv- ıg activity These packs shall conform to the applicable carrier ıles and regualtions.

5 3 Marking Each exterior container shall be marked as spec- ied in MIL-STD-129.

6. NOTES

6 1 Intended use The table shall be used as an anesthetist's and in the operating room, or for general utility.

6 2 Ordering data Purchasers should select the preferred ıtıons permitted herein and include the following information ı procurement documents.

a. Title, number and date of this specification.
b Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
c. When a first article is required (see 3.1, 4.1.2 1 and 6.3).

6.3 First article When a first article is required, it shall . tested under the appropriate provisions of 7-104 55 of Armed ervices Procurement Regulation. The contract or purchase order ecifies the quantity procedure for delivery and other specific ıstructions for inspection and approval of the first article.

6 4 This specification covers the following item appearing in ıe Federal Supply Catalog

National Stock Number	Item Identification
6530-00-926-9014	TABLE, ANESTHETIST'S

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

6.5 An acceptable mastic for that specified in 3.2.1 is Nataz no. 75 Airless Coating, manufactured by Pure Asphalt Company, Chicago, IL.

6.6 Changes from previous issue. Asterisks are not used in this revision to identify changes with respect to the previous issue, due to the extensiveness of the changes.

Custodians
Army - MD
Navy - MS
Air Force - 03

Preparing activity
DSA-DM

Project No. 6530-1394

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No 22 R255

INSTRUCTIONS The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government Fold on lines on reverse side, staple in corner, and send to preparing activity Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements Attach any pertinent data which may be of use in improving this document If there are additional papers, attach to form and place both in an envelope addressed to preparing activity

DOCUMENT IDENTIFIER AND TITLE
TABLE, ANESTHETIST'S MIL-T-36522B

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1 HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2 COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3 IS THE DOCUMENT RESTRICTIVE

☐ YES ☐ NO *(If "Yes", in what way?)*

4 REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

GPO 1974— 603 680/ 6126 2 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:44Z
Check the source to verify that this is the current version before use.