MIL-T-36550A
19 October 1976
SUPERSEDING
MIL-T-36550
13 December 1967

MILITARY SPECIFICATION

TEMPLE, SPECTACLE, GOLD-PLATED, CABLE TYPE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one type and class of all metal gold-plated spectacle temple of the comfort cable type.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

FEDERAL

QQ-B-750	- Bronze, Phosphor Bar, Plate, Rod, Sheet, Strip, Flat Wire and Structural and Special Shaped Sections.
PPP-B-566	- Box, Folding, Paperboard.
PPP-B-636	- Box, Shipping, Fiberboard.
PPP-B-676	- Box, Setup.

MILITARY

MIL-G-45204	- Gold Plating, Electrodeposited.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6540

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.

STANDARDS

FEDERAL

Fed. Test Method Std. No. 151 - Metals; Test Methods.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

23081 TEMPLE, SPECTACLE, GOLD-PLATED, COMFORT CABLE TYPE, 6-1/2 INCH.
24033 FRAME, SPECTACLE, GOLD-PLATED, BAYONET TEMPLES.

(Copies of specifications, standards and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Materials. Materials used for the temples shall conform to the requirements specified herein. Materials not specifically designated shall be of the highest commercial quality and suitable for the intended purpose.

3.1.1 Shank, hinge joint and cable winding. The base metal for these components shall be a 15 percent nickel silver alloy commercially designated as Copper Alloy No. 754. The alloy shall conform to the chemical requirements shown in table I.

3.1.2 Cable core. The cable core used for support of the windings shall be copper alloy No. 754 or phosphor bronze conforming to QQ-B-750, Composition A.

3.1.3 Electroplating. The shank and the outer windings of the cable portion shall be electroplated with a minimum of 100 millionths of an inch of bright copper and overplated with a minimum of 60 millionths of an inch of 23 carat bright gold. The complete temple shall be finished with a flash plating of 18 carat bright gold.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.

TABLE I. Copper alloy no. 754

Element	Minimum	Maximum
Copper	63.5	66.5
Nickel	14.0	16.0
Lead	--	0.10
Iron	--	0.25
Manganese	--	0.50
Zinc	Remainder	
Total other elements	--	0.50

3.2 Design and construction. The temple shall consist of a solid shank flat core connected to a flexible earpiece section. The earpiece section shall consist of nickel silver cable windings on a nickel silver or phosphor bronze core. The temples shall be gold plated as specified in 3.1.3. The hinge joint shall fit the spectacle front component of Frame, Spectacle, Gold Plated, Bayonet Temples as shown on Defense Personnel Support Center Drawing (DPSC) No. 24033. The temple shall conform to Defense Personnel Support Center Drawing 23081. The temple unit shall consist of one pair, consisting of one left and one right temple.

3.3 Finish. The finish exhibited by the temple shall be the plating described in 3.1.3. The plating shall be smooth, unstained, free of discoloration and as specified herein.

3.4 Identification marking. The temple shall be distinctly and inconspicously marked by metal stamp with the manufacturer's name or registered trademark, and the overall temple length on the inside of the temple butt.

3.5 Workmanship. Workmanship shall be first class throughout. Temples shall be free from all defects which may affect their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specifications where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:47Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the metal, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as shown in table II. The unit of product for examination shall be one pair.

TABLE II. Sampling for examination.

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	4.0
Dimensional	S-3	2.5

4.2.2 For tests. Sampling for tests shall be in accordance with MIL-STD-105 and as shown in table III. The sample unit shall be one pair.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.

TABLE III. Sampling for tests.

Characteristic	Sample unit	Inspection level	AQL (percent defective)
Base metal	4 ounces from each lot of each metal used	--	1/
Plating thickness	One pair	S-3	1/

1/ Acceptance number shall be zero.

4.3 Examination.

4.3.1 Dimensions. Any dimension not within the specified tolerance shall be classed as a defect and be cause for rejection of the temple.

4.3.2 Defects. The temples shall be examined for defects including but not limited to those listed in table IV. Additional defects shall be classified as described in MIL-STD-105.

TABLE IV. Classification of defects.

Categories	Defects
Major	
101	Temple not gold plated.
102	Plating stained or discolored which seriously affects appearance.
103	Plating not continuous or cut through.
104	Plating not smooth, not uniform, or non adherent.
105	Temple bent out of shape, twisted, malformed or otherwise damaged which seriously affects serviceability or appearance.
106	Sharp edges and burrs present.
107	Incomplete solder joint which seriously affects serviceability or appearance.
108	Solder spatter or excess solder which seriously affects appearance or serviceability.
109	Metal damaged or deformed in soldering, which seriously affects serviceability or appearance.
110	Improper and loose cable winding.
Minor	
201	Front and temple not properly matched, not marked, or marking not legible.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.

TABLE IV. Classification of defects.-continued

Categories	Defects
Minor	
202	Evidence of plating porosity.
203	Foreign matter embedded in temple.
204	Stained or discolored plating which does not seriously affect appearance.
205	Excess solder or solder spatter which does not seriously affect appearance or serviceability.
206	Metal slightly damaged or deformed in soldering which does not seriously affect serviceability or appearance.

4.4 Packaging inspection. The inspection of the unit, intermediate package, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.5 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.5.1 Chemical analysis of the metals. The samples selected in accordance with 4.2.2 shall be analyzed in accordance with methods 111.2 of 112.2 of Fed. Test Method Std. No. 151 for compliance with requirements of Section 3 and subsidiary specifications. In case of dispute, method 111.2 shall be used for referee analysis.

4.5.2 Plating thickness. The thickness of the gold plating and the copper underplating shall be determined by microscopic, beta-radiation back scatter or x-ray fluorescence as described in applicable portions of MIL-G-45204. Determinations shall include the flat portion of the shank and the top winding of the cable portion.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. One pair of temples, as specified, constitutes one unit of issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.

5.1.2 Level A.

5.1.2.1 Unit package. One pair of temples shall be packaged in an envelope of suitable size and design. The material from which the envelope is fabricated shall be such as to prevent marring or scratching of the temples. The flap (closure) of the envelope shall not be glued or sealed.

5.1.2.2 Intermediate package. Twelve unit packages shall be packed in box conforming to PPP-B-566 or PPP-B-676. Closure shall be as specified in the appendix of the applicable box specification.

5.1.3 Level C. One pair of temples shall be packaged in a manner that will afford adequate protection against corrosion, deterioration and physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Forty intermediate pacakges (480 units) shall be packaged in an exterior container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

5.2.2 Level B. Forty intermediate packages (480 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The packaged temples shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carrier rules and regulations.

5.3 Marking.

5.3.1 Interior packages. Each unit and intermediate package shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior containers shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use. The spectacle temples covered by this specification are intended as an alternate to the bayonet temples used with Frame, Spectacle, Bayonet Temples, Gold-Plated, Double Bar, NSN 6540-00-782-6502.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(c) National stock number.

6.3 This specification is intended to cover the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6540-00-926-9013	TEMPLE, SPECTACLE, GOLD-PLATED, Comfort Cable Type, 6-1/2 inch.

6.4 Changes. Asterisks to indicate changes have been omitted due to the extensiveness of the changes.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Logistic Agency - DM

Project No. 6540-1044

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-T-36550A - TEMPLE, SPECTACLE, GOLD-PLATED, CABLE TYPE

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20th STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:47Z
Check the source to verify that this is the current version before use.