MILITARY SPECIFICATION

TABLE, EXAMINING AND TREATMENT

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers a cabinet-type examining and treatment table having adjustable head and leg sections.

2. APPLICABLE DOCUMENTS

2.1 The following documents of issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

QQ-S-698	Steel Sheet and Strip, Low Carbon.
TT-C-490	Cleaning Methods and Pretreatment of Ferrous Surfaces for Organic Coatings.
TT-E-491	Enamel Gloss, Synthetic (for Metal and Wood Hospital Furniture).
TT-P-659	Primer-Surfacer, Synthetic, Tints and White, (for Metal and Wood Surfaces).
UU-P-271	Paper, Wrapping, Waterproofed Kraft.
UU-P-510	Paper Sheeting, Bacteriostatic.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.

STANDARDS

Federal

Fed. Std. No. 66 Steel, Chemical Composition and Hardenability.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

U.S. DEPARTMENT OF COMMERCE

Commercial Standard CS45 Douglas Fir Plywood.

(Copies of Commercial Standard CS45 may be obtained upon application, accompanied by money order, coupon, or cash, to the Superintendant of Documents, Government Printing Office, Washington 25, D.C. Price may be obtained from the Superintendant of Documents.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Steel.

3.1.1.1 Carbon steel. The sheets and strips used in the table shall conform to QQ-S-698 for chemical composition, tolerances on thickness, and bending properties. The exposed surfaces shall be either cold-rolled, condition 4 or 5, regular bright finish, stretcher leveled; or hot-rolled, annealed condition, pickled and oiled, stretcher leveled. The unexposed surfaces shall be of the same steel as the exposed surfaces, except that stretcher leveling may be omitted. Sheet steel shall be of uniform gage, commercial tolerances being permitted for all gages specified. Unless otherwise specified, gage of meterials shall be United States Standard (U.S.S.) gage.

3.1.1.2 Corrosion-resisting steel. Corrosion-resisting steel (C.R.S.) shall conform to AISI 302 or 304 of Fed. Std. No. 66.

3.1.2 Plywood. Plywood shall be Douglas fir plywood conforming to grade A veneer, interior type grade A-A of Commercial Standard CS45.

3.1.3 Upholstery.

3.1.3.1 Vinyl. Vinyl for upholstery shall be a heavy duty reinforced 32 ounce vinyl upholstery material (see 6.3).

3.1.3.2 Filler. Filler shall be firm premium grade polyurethane.

3.1.3.3 Upholster's nails. Upholster's nails shall be over-headed needle-point upholster's nails, 5/8 inch long with a 7/16 inch head diameter, nominal dimensions.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

3.2 Design. The table shall be a steel cabinet type table with three section upholstered top and manually adjustable back and leg sections. The table top shall be 53 ± 1/2 inches long, 24 ± 1/2 inches wide and 33 ± 1/2 inches high at the foot end. With leg rest extended, the table shall have an approximate length of 74 inches.

3.3 Construction.

3.3.1 Cabinet. The cabinet section shall be formed of not less than No. 20 U.S.S. gage, 0.0375 inch, carbon steel; except that the table top shall be No. 18 U.S.S. gage, 0.050 inch, carbon steel with a 1 inch continuous flange. The cabinet body shall be 22 ± 1/2 inches wide, 42 ± 1/2 inches long and 28 ± 1/2 inches high. The cabinet construction shall be compartmentized so that the table shall contain the storage compartments listed in Table I.

Table I.

Table section	Number of compartments	Width * (inches)	Height * (inches)	Depth * (inches)
Foot end	1-upper	22	10 3/4	13 1/2
	1-lower	22	6	13 1/2
Head end	1-lower	22	10 3/4	13 1/2
	1-upper	22	10 1/4	13 1/2
Center section (accesible from either side)	1-lower	14 1/2	11 3/4	22
	1-upper	14 1/2	9 1/4	22

* The Dimensional tolerance shall be ± 1/4 inch.

The center section shelves shall be covered in C.R.S. There shall also be a pass through drawer with C.R.S. fronts located in the upper center section. The table shall stand on square tubular legs equipped with plastic guides.

3.3.2 Upholstered top. The upholstered top shall consist of a stationary seat section 17 by 24 inches, an adjustable head piece 36 by 24 inches tapered to eliminate the need for a pillow, and a swing up leg rest 15 inches wide by 22 inches long. Dimensional tolerance for each section shall be ± 1/4 inch. The top shall be made of minimum 3/4 inch veneer plywood. The center and head pieces shall be padded with a minimum of 1 1/2 inch thick polyurethane foam, and the leg rest with a minimum of 1 inch thick foam. The edges shall be reinforced with 1 inch thick vertical maple side cleats for maximum rigidity. The center and head pieces shall be joined by means of a continuous steel hinge. The head piece shall be capable of being raised to positions of 22 1/2, 30, 45, and 60 degrees by a rack type locking device which shall be controlled by handles located on either side. After testing as specified in 4.4.1, the head section shall show not sign of distortion. The leg rest shall swing up from the vertical position and be capable of being securely locked in the horizontal position. After testing as specified in 4.4.2, the leg rest shall show no sign of distortion.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

3.3.3 Footstep. The table shall have a strong and sturdy pull-out footstep, with a nonskid rubber top. The footstep shall rest evenly on a flat surface and retract under the cabinet when not in use. After testing as specified in 4.4.3, the footstep shall show no sign of distortion.

3.3.4 Stirrups. The foot end shall be equipped with two adjustable stirrups. The stirrups shall extend horizontally and vertically for GU and GYN examinations and shall lock automatically in any position desired without tee handles or hand wheels. When not in use, they shall retract into the table. The stirrups shall be a satin chrome plated steel.

3.3.5 Paper roll attachment. The head end of the table shall be equipped with a paper roll attachment for holding rolls of paper 18 ± 1/16 inches wide. A steel spring or other suitable device shall be provided to hold the paper tight on the roll. The other end of the table shall have suitable means for retaining the stretched paper in place; and a paper cutter. Paper rolls shall be in accordance with UU-P-510. Twelve rolls of paper shall be supplied with each table.

3.4 Finish.

3.4.1 Cleaning and surface preparation. Cleaning and surface preparation shall be accomplished as specified for type I of TT-C-490. All welds shall be thoroughly and completely cleaned of flux, and scales, prior to application of primer.

3.4.2 Primer. After cleaning, all exposed metal surfaces shall be given one coat of a synthetic primer conforming to TT-P-659. The primer shall be baked on for not less than 45 minutes, at a temperature of 250° Fahrenheit (F.), or for an equivalent baking period at a higher or lower temperature as recommended by the paint manufacturer and approved by the contracting officer. The primer shall be applied so as to level out and produce a uniform surface free from pebbles or wrinkles.

3.4.3 Enamel. After the primer has been allowed to cool, all surfaces shall then be covered with two coats of synthetic gloss enamel conforming to class B of TT-E-491. Each coat shall then be baked on separately for not less than 45 minutes, at a temperature of 250° F. or for an equivalent baking period at a higher or lower temperature as recommended by the paint manufacturer and approved by the contracting officer. The enamel coats shall be applied so as to level out and produce a final, smooth, uniform surface without runs, wrinkles, orangepeel, or grit, and without areas of thin or no film. The total dry film thickness shall be not less than 0.002 inch at any point.

3.4.4 Color. The color of the cabinets shall be coral, green, blue, or white. Cabinet and top upholstery combinations are: coral with charcoal; green with olive; blue with turquois; or white with buff. Color shall be as specified by the Procuring Activity (see 6.2).

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:48Z
Check the source to verify that this is the current version before use.

3.5 Identification marking. A metal nameplate shall be firmly and securely attached to each table. The marking shall be distinct and legible. The nameplate shall bear the following information:

Name or registered trademark of the manufacturer.
Stock number.
Contract or purchase order number.
U.S.

Nameplates attached by adhesives shall not be used.

3.6 Service data. The contractor shall enclose, with each unit, two copies each, of one or more service data booklets. In addition, the contractor shall mail 15 copies of the approved service data to the contracting officer for distribution. The service data shall contain complete instructions for installation, operation, maintenance, and a list of component parts. Each part of the unit other than common fasteners, shall be identified in the service data and parts list by the manufacturer's part number. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the contracting officer, via the cognizant Government inspector, two preliminary copies of his proposed service data, for review and comment by Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary service data shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of required revisions and corrections in the final service data without delaying delivery under the contract or purchase order. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of contract, and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request.

3.7 Workmanship. Workmanship shall be first class throughout. Examining tables shall be free from defects which detract from their appearance or may impair their serviceability. The table shall be free from sharp edges.

3.7.1 Welding. Welding shall be durable, well fused, sound in every detail and free from blisters or blow-holes. Welding shall be accomplished without any resultant distortion and shall be free from twists or strain. Spot welds shall be accomplished so as to insure sound welds of sufficient strength to tear the spot out of the parent metal when pulled apart. All exposed welds shall be ground and polished smooth to eliminate traces of weld marks, pits, discontinuities, and laps. Spot welds shall be smoothly dressed to blend with adjacent surfaces.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total (major and minor combined)	II	4.0
For dimensional examination	S-3	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and Table II. The acceptance level shall be zero.

Table II.

Component	Characteristic	Requirement	Test Method	Inspection level
Head section	Load strength	3.3.2	4.4.1	S-4
Foot section	Load strength	3.3.3	4.4.2	S-4
Mounting step	Load strength	3.3.2	4.4.3	S-4
Finish	Paint	3.4	*	*

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

*See 3.4 for applicable paint specification.

4.3 Examination. The table shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects:

Categories and defects**

Major:

101	Component missing.
102	Upholstery fabric not properly tailored, form-fitting or properly secured to its base.
103	Fabric color not as specified; not free from blemishes, and cracks.
104	Head and leg sections not freely moving; not secure in required positions.
105	Drawer not furnished with stainless steel front.
106	Drawer not freely moving.
107	Center section of top not securely attached to top of cabinet.
108	Foot step not as specified.

Minor:

201	Table does not rest evenly on level floor.
202	Identification marking missing, not complete, correct or legible.
203	Table not free from sharp or uneven edges.
204	Cabinet color not as specified.
205	Covering on pull-out footstep not as specified; not smoothly attached to base.
206	Table not free from grease, dirt, extraneous matter, dents and scratches.

**Examination shall not be restricted to the defects listed above.

4.3.2 Dimensional examination. The table shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests. Static load tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

4.4.1 Head Section. A 50 pound weight shall be placed on the head-end section and secured in position. The loaded head-end section shall then be raised to the full vertical operating position and secured in position. The section shall then be lowered to each of its operating positions allowing the support bracket to rest in its respective place. This operation shall be repeated 15 times. At the end of this procedure, the weight shall be increased to 100 pounds and the operation described above repeated. The section, table, and fittings shall be examined for distortion.

4.4.2 Foot End Section. The foot-end section of the table shall be raised to the horizontal position, and a 250 pound load applied evenly on the fully extended section. The table shall then be examined for distortion.

4.4.3 Pull-out Footstep. A 250 pound weight shall be placed on the footstep and shaken by hand 10 times. The footstep shall then be examined for distortion.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C (see 6.1).

5.1.1 Unit of issue. One table, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Unit packaging. Each unit shall be disassembled as far as is practicable. Disassembled parts shall be properly protected and packed within the unit in such a manner as to prevent movement. Doors and drawers shall be closed and secured in position. All exterior surfaces shall be protected by wrapping with kraft paper and cushioning with flexible corrugated fiberboard. Wrapping and cushioning material shall be secured by tying with cotton cord or tape, or by sealing with kraft gummed paper tape. Work shall be accomplished in such a manner, as to insure against marring, scratching, or damage to the unit.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A of C (see 6.1)

5.2.1 Level A. Each unit shall be overpacked in an exterior container, designed for type 3 load, and constructed in accordance with PPP-B-621, table IV, or PPP-B-601, table II, type I, class 1 minimum 5/16 inch thick plywood. The unit shall be secured to the base of the exterior container in such a manner that the weight of the unit shall not be carried directly by its legs. In addition, the unit shall be further secured within the exterior container by means of bolts or wood braces to insure against

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

shifting, breakage, or damage. Each unit shall be protected by a shroud or lined with waterproof barrier material conforming to UU-P-271, class L-2(a) or L-2 (b). The exterior container design shall include three skids of not less than 1 1/2 inches in thickness applied crosswise of base, as to facilitate handling with a forklift truck and to maintain unit in an upright position. Closure shall comply with appemdix to applicable specification. Strapping will not be required for shipments forwarded to a recieving activity within the continental limits of the United States for storage.

5.2.2 Level C. Each unit shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each container shall be marked to identify contents.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES.

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packing (see 5.1 and 5.2).

6.2 Availability of color chips. Color chips of the colors referenced in 3.4.4 are available from Naval Supply Depot, Philadelphia.

6.3 Vinyl covering is available from E.I DuPont DeNemours Co. under the trade name "Cordero".

6.4 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

Custodians:

Army - MD
Navy - MS
Air Force - 03

Review activity:

Army - MD
Navy - MS
Air Force - 03

Preparing activity:

Defense Supply Agency (DM)

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION

MIL-T-36555

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO (If "yes", in what way?)

4. REMARKS (Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)

SUBMITTED BY (Printed or typed name and activity - Optional)

DATE

DD FORM 1426 REPLACES EDITION OF 1 OCT 54, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL CODE ATS
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
AUG 66

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.