MIL-V-36535A
28 July 1975
Superseding
MIL-V-36535
10 April 1967

MILITARY SPECIFICATION

VALVE, SURGICAL DRAIN, CHEST, HEIMLICH, DISPOSABLE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a sterile, disposable, surgical drain valve of the Heimlich design.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-S-36586	Sterilization Strip Test, Bacterial Spore.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:46Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Personnel Support Center (DPSC)

No. 24090 VALVE, SURGICAL DRAIN, CHEST, HEIMLICH, DISPOSABLE

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions shall be obtained from the procuring activity or as directed by the contracting officer).

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

THE UNITED STATES PHARMACOPEIAL CONVENTION, INC.

The Pharmacopeia of the United States of America (USP)

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

U.S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE

Bureau of Biologics. Food and Drug Administration Code of Federal Regulation Part 273. Minimum requirements (as applicable).

(Application for copies should be addressed to the Food and Drug Administration, 5600 Fishers Lane, Rockville, MD 20852.)

3. REQUIREMENTS

3.1 Plastic and rubber components. All plastic and rubber components shall be of the best quality and suitable in all respects for the intended purpose. All plastic and rubber components shall be free from pyrogens, when tested as specified in 4.4.2. All plastic components shall meet the requirements of the USP for Class II plastics, when tested as specified in 4.4.3. The inlet port and rubber flutter vent tubing components only, shall pass the safety (toxicity) test specified in 4.4.4.

3.1.1.1 Rubber flutter vent tubing. The rubber flutter vent tubing shall be a single piece of formed pure gum rubber, amber colored, lightly powdered with talc. The vent tubing shall be open at the proximal end, flat in section, with internal surfaces in contact for a sufficient length

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:46Z
Check the source to verify that this is the current version before use.

from the distal end to prevent reverse flow, even during passage of fluid, air or clots. The open end shall be secured to the distal end of the plastic inlet port in a leakproof manner.

3.1.1.2 Flutter vent tubing sleeve. Flutter vent tubing sleeve shall be fabricated from amber colored, latex rubber.

3.1.1.3 Plastic inlet port. Inlet port shall be fabricated from rigid plastic and shall be colored for identification purposes. The proximal end shall have an external taper to accommodate catheters or plastic tubing as appropriate. The distal end shall contain internal grooves which shall accommodate the plastic body casing.

3.1.1.4 Plastic body casing. Body casing shall be fabricated from rigid transparent plastic. The distal end of the body casing shall have an external taper (outlet port) to accommodate catheters or plastic tubing as appropriate. The proximal end of the body casing shall be open and have formed protrusions which shall snap into the grooves of the plastic inlet port to form a leakproof assembly, after the rubber flutter vent tubing and the flutter vent tubing sleeve have been assembled to the plastic inlet port.

3.2 Style, design, and dimensions. Shall be a sterile, surgical drain valve, of the Heimlich design suitable for providing chest drainage. Valve shall be designed to provide a safe, one-way drainage device. The valve assembly shall consist of the components specified above. The inlet and outlet ports shall be designed to accommodate catheters or tubing in a secure manner without the need of clamping devices. The valve assembly shall be in accordance with DPSC Drawing No. 24090.

3.3 Finish. The valve shall be smoothly finished. Surfaces shall be free from cracks, pits, excessive flash, holes, and foreign matter.

3.4 Security of assembly.

3.4.1 Rubber flutter vent tubing to plastic inlet port. The rubber flutter vent tubing shall not separate or loosen from the plastic inlet port when tested as specified in 4.4.5.1.

3.4.2 Plastic inlet port to plastic body casing. The plastic inlet port and plastic body casing junction shall not separate or loosen when tested as specified in 4.4.5.2.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:46Z
Check the source to verify that this is the current version before use.

3.5 Leakage.

3.5.1 Plastic inlet port to body casing. The plastic inlet port and body casing junction shall not exhibit leakage when tested as specified in 4.4.6.1.

3.5.2 Drain valve assembly. The complete drain valve assembly shall not lose total vacuum when tested as specified in 4.4.6.2. In addition, the assembly shall permit free air flow from the inlet port to formed rubber flutter vent tubing.

3.6 Sterility. All components of the drain valve shall be sterile when tested with a biological indicator as specified in 4.4.1.

3.6.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item identification.
b. National stock number.
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.)
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or a designated qualified alternate.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, PA 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:46Z
Check the source to verify that this is the current version before use.

3.6.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and simultaneously sterilized in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.1. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

3.7 Certificates of freedom from pyrogens and biological tests. A certificate listing all lot or control numbers shall state that representative samples from each lot or control have been tested in accordance with methods described in the USP and found to be free from pyrogens, and have passed the biological tests. These certificates shall include the specific test performed and the test results. These certificates shall be submitted by the contractor in the manner specified in 3.6.1.

3.8 Certificate of safety. A certificate of safety, for the inlet port and rubber flutter vent tubing components only, listing all lot or control numbers, shall state that representative samples from each lot or control have been tested for safety in accordance with methods described in the USP and found to be safe (nontoxic). This certificate shall include the specific tests performed and the test results. This certificate shall be submitted by the contractor in the manner specified in 3.6.1.

3.9 Manufacturer's record. The manufacturer shall place a code or lot number on the paper label, or on the printed carton, which can be verified against the manufacturer's records of all tests. These records shall be made available for inspection by any duly qualified official of the Federal Food and Drug Administration. No shipment shall be made until all the tests have been completed, verified, and recorded in the permanent records of the manufacturer.

3.10 Identification markings. The contractor's name or registered trademark shall appear on the body casing. In addition, the direction of drainage flow shall be indicated on the casing.

3.11 Workmanship. Workmanship shall be first class throughout. The surgical drain valves shall be free from defects which detract from their appearance or impair their serviceability. The flutter vent tubing shall be completely flat along its full length and free from pits or holes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:46Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the latex rubber may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number (NSN), the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in Table I. The unit of product for examination purposes shall be one complete drain valve.

Table I. Sampling Criteria for Examination

Type Defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:46Z
Check the source to verify that this is the current version before use.

4.2.2 For tests.

4.2.2.1 For sterility test. Sampling for sterility test shall be in accordance with the USP for parenteral devices and as specified in 4.4.1.

4.2.2.2 For pyrogen and safety tests. Sampling for pyrogen test and safety test shall be as indicated in the USP for transfusion and infusion assemblies. Samples shall be selected which are to be representative of each lot of each material used in the fabrication of the drain valves. As an alternate, sampling shall be on a sterilizer lot basis when sterilizer lots are smaller than, or equal to, production lots. Only the inlet port and the rubber flutter vent tubing shall be subjected to the safety tests.

4.2.2.3 For other tests. Sampling for other tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. The unit of product for testing purposes shall be one complete drain valve. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

Table II. Sampling Criteria for Tests

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL
Valve	Dimensions	3.2	--	S-1	2.5
	Security of assembly-vent tubing to inlet port	3.4.1	4.4.5.1	S-2	1.5
	Security of assembly-inlet port to body casing	3.4.2	4.4.5.2	S-2	1.5
	Leakage - inlet port to body casing	3.5.1	4.4.6.1	S-2	1.5
	Leakage - complete assembly	3.5.2	4.4.6.2	S-2	1.5

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:46Z
Check the source to verify that this is the current version before use.

4.2.3 Sampling for biological tests. Sufficient samples to perform the biological tests shall be selected from each lot and shall be representative of each batch of plastic used in the manufacture of the drain valves.

4.3 Examination. The drain valves shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Tables I and III. Examination, however, shall not be restricted to defects listed in Table III. Any valves in the sample containing one or more defects shall be rejected, and if the number of defective valves in any sample exceeds the acceptance number for that sample, the lot represented by that sample shall be rejected.

Table III. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Valve not of the design specified.
102	Vent tubing restricted or occluded at the distal end of the inlet port.
103	Inlet port and body casing not fabricated from rigid plastic.
104	Inlet port not secured to body casing and vent tubing.
105	Valve not free from cracks, pits, holes, excessive flash or foreign matter.
106	Vent tubing not flat along its full length.
Minor	
201	Valve does not possess markings specified in 3.10.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Sterility test. Drain valves shall be tested for sterility in accordance with the method specified in the USP for parenteral devices using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:46Z
Check the source to verify that this is the current version before use.

distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in paragraph 4.4.1.1. The minimum number of test units, regardless of lot size, shall be 10, except for innoculated carrier which shall be 20 for Type I Sterilization and 30 for Type II Sterilization including inoculated carrier and product tests units.

4.4.1.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65°C
Dry Heat	B. stearothermophilus, or	55° - 65°C
	B. subtilis var. niger (or globigii)	30° - 35°C
Gas	B. subtilis var. niger (or globigii)	30° - 35°C
Radiation	B. pumilus	30° - 35°C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

4.4.1.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.1.1 Sterility test and 4.4.1.1 Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.4.1.1. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

4.4.1.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method.

4.4.1.4 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP.

4.4.2 Freedom from pyrogens. Samples of the drain valves shall be tested for freedom from pyrogens in accordance with the USP for transfusion and infusion assemblies.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:46Z
Check the source to verify that this is the current version before use.

4.4.3 Biological tests for Class II plastics. All plastic components of the drain valve shall be tested in accordance with the requirements of the USP for Class II plastics.

4.4.4 Safety test. The plastic inlet port and rubber flutter vent tubing components of the drain valve shall be tested for safety in accordance with the requirements of the USP for transfusion and infusion assemblies.

4.4.5 Security of assembly. (See DPSC Drawing No. 24090).

4.4.5.1 Rubber flutter vent tubing to plastic inlet port. The formed rubber flutter vent tubing shall be subjected to an axial force of 2 pounds for 10 seconds with the vent tubing assembled to the vent tubing sleeve and the inlet port appropriately secured. This test can be performed during production of the valve or can be performed on the end item (destructive testing).

4.4.5.2 Plastic inlet port to plastic body casing. The plastic inlet port to plastic body casing junction shall be subjected to a 10 pound axial force, with the body casing appropriately secured, for at least 10 seconds.

4.4.6 Leakage

4.4.6.1 Plastic inlet port to body casing. The inlet port shall be plugged in a leakproof manner. With the drain valve placed in the upright position, a suitable liquid shall be admitted into the outlet port until the drain valve is completely filled. The junction of the plastic inlet port to body casing shall be observed for leakage for at least 3 minutes.

4.4.6.2 Drain valve assembly. The complete drain valve assembly shall be subjected to a vacuum of 0.5 inch of mercury applied to the inlet port with the outlet port plugged in a leakproof manner. The vacuum shall be applied for at least 10 seconds.

4.5 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Source: https://assist.dla.mil -- Downloaded: 20261O-21T08:46Z
Check the source to verify that this is the current version before use.

Table IV. Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and packing materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.1).

5.1.1 Unit of issue. One package containing ten (10) valves, as specified, constitutes one unit of issue.

5.1.1.1 Level A.

5.1.1.1.1 Immediate container. Each valve shall be packaged in a peel type sealed container capable of maintaining the sterility of the contents until manually opened.

5.1.1.1.2 Unit package. Each unit of issue (10 valves) shall be packaged in a box of appropriate size and constructed in accordance with PPP-B-566, PPP-B-676, or PPP-B-636, type CF class domestic. Closure shall be adequate to prevent accidental opening. Each box shall be further protected by being overwrapped or bagged in plastic film and sealed.

5.1.4 Level C. Each unit of issue shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:46Z
Check the source to verify that this is the current version before use.

5.2 Packing. Packing shall be level A, B or C as specified. (see 6.1).

5.2.1 Level A. Twelve (12) units of issue shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type, PPP-B-621, class 2; or PPP-B-636, class weather resistant, grade V3C. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof as specified in the appendix of PPP-B-636.

5.2.2 Level B. Twelve (12) units of issue shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, overseas type; PPP-B-621, class 1; or PPP-B-636, type CF, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. Units of issue shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Immediate container. Each immediate container shall be marked in accordance with the requirements of the Federal Food, Drug and Cosmetic Act. Marking shall include item name, manufacturer's name and address, sterility lot control number, the legends "STERILE" and "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED" or similar statement.

5.3.2 Unit package and exterior container. Each unit package and exterior container shall be marked as specified in MIL-STD-129. Marking shall include the sterility lot control number. Each unit package and exterior container shall be marked with the method of sterilization (G-GAS, R-RADIATION, C-COLD, S-STEAM, D-DRY HEAT STERILIZATION). Marking shall follow the level of packaging and date of packaging. Example: "A-6/75-G). Final sterilization shall be on or after indicated packaging date. In addition, each unit package shall be marked with the legend: "STERILE - CONTENTS OF INNER WRAP ARE STERILE UNLESS IT HAS BEEN OPENED OR DAMAGED" or similar statement.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:46Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. Purchasers should exercise any desired options offered herein, and the procurement document should specify the following:

(a) Title, number, and date of this specification.
(b) National Stock Number.
(c) Applicable levels of packing and packaging (see 5.1 and 5.2).

6.2 This specification covers National Stock Number 6515-00-926-9150 VALVE, SURGICAL DRAIN, Chest, Heimlich, Disposable, 10s, appearing in the Federal Supply Catalog.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1963
is

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:46Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:46Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-V-36535A VALVE, SURGICAL DRAIN, CHEST, HEIMLICH, DISPOSABLE

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 20261O-21T08:46Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:46Z
Check the source to verify that this is the current version before use.