

FED. SUP CLASS
6640





PART NUMBER	DIMENSIONS (cm) L	W	H	OPER TEM °C	CHAMBER (cm) DIA.	LENGTH	POWER REQUIREMENTS VOLTS AC*	Hz	W
MS36158-1	44	19	30	982	3.5	30	115	60	550
MS36158-2	30.5	30.6	37.1	2000	5.0	30.5	120	50/60	1300
MS36158-3	33	27	42	1093	5.2	30.5	20 AC/DC	-----	1000
MS36158-4	34	25	35	165 to 1000	2.5	30.5	115	60	800
MS36158-5	84	40	38	3800	8	61	240	50/60	3800
MS36158-6	135	58	56	11000	15	91	240	50/60	11000
MS36158-7	23	50	29	960	2.5	30	115/230	50/60	960
MS36158-8	50	23	37	1400	5	30	115/230	50/60	1400
MS36158-9	80	30	37	3400	6.4	61	115/230	50/60	3400

*Unless otherwise specified.

MATERIAL: Furnaces shall be made of steel, ceramic insulation and chromel A or nickel-chromium resistance wire.

DESIGN: MS36158-1 and -2 furnaces shall be cylindrical shape mounted on an iron or steel frame. The chromel A heating wire shall be wound on a grooved refractory through which the combustion tube, or chamber, is inserted and held in position by grooves of the refractory. The heat insulating chamber is enclosed in a metal jacket. To regulate temperature below maximum voltage, an autotransformer or another type of control device must be used.

MS36158-3 furnace shall be the same construction as -1 and -2, except it shall be powered by a 115 volt, 60 Hz transformer with a 20 volt output.

MS36158-4 shall be a hinged tube furnace with a built-in input adjuster to provide manual control of temperatures from 165° to 1000°C. Heating elements shall be nickel-chromium wire, helically wound and mounted on high temperature ceramic supports. The case shall be finished with heat resistant enamel and equipped with a heavy top supporting hinge, handle and snap catch.

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 82

APPROVED 22 JUN 1983 REVISED

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		FURNACE COMBUSTION TUBE, LABORATORY	MS36158
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 2

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1184
STANDARD DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:35Z
Check the source to verify that this is the current version before use.

MS36158-5 and -6 furnaces shall have the heating chamber divided into three zones for providing temperature control by use of three thermocouples. Each furnace shall have a controller unit, which will serve as the mounting base.

MS36158-7, -8, and -9 furnaces shall be single zone, but with a split cylindrical chamber that is heated by two semi-cylindrical heating units, providing radial heat uniformity. The split-hinge design permits convenient placement of the process tube and easy operation. The heating elements are helically-coiled high temperature resistance wire. Each furnace shall have a controller unit, which will serve as the mounting base.

All electrical components shall conform to the standards of Underwriters Laboratories, Incorporated.

NOTE: Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

USER ACTIVITIES:

REVIEWER ACTIVITIES:

APPROVED 22 JUN 1983 REVISED

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A.: Other Gust	INTERNATIONAL INTEREST	TITLE: FURNACE COMBUSTION TUBE, LABORATORY	MILITARY STANDARD MS36158
Procurement Specifications	SUPERSEDES:		SHEET 2 OF 2

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1184
STANDARD DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:35Z
Check the source to verify that this is the current version before use.