MILITARY SPECIFICATION

NEEDLES, HYPODERMIC, CARTRIDGE TYPE, DISPOSABLE

This Amendment forms a part of Military Specification MIL-N-36319A, dated 9 April 1970, and is mandatory for use by all Departments and Agencies of the Department of Defense.

Page 4:

3.2 Style, design, and construction. Delete period at end of paragraph and add the following:

", except for the below - listed changes in the drawing:

ASSEMBLY --- Lancet, delete '50% - 60%' and substitute '30% to 50%'.

CANNULA DIMENSIONS --- After B ℄, delete '.005' and substitute '10%'.

Under B delete '.080' wherever it appears and substitute '.094'.

Under B delete '.068' wherever it appears and substitute '.078'."

Page 6:

Delete 4.1 thru 4.1.2 and insert the following:

"4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

"4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 20261O-21T08:24Z
Check the source to verify that this is the current version before use.

"4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

"4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the stainless steel and plastic, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements."

Page 12:

5.1.1 Unit of issue. Line 1, delete "250" and substitute "100".

5.1.2 Packaging quantities. In Table V, in column headed "Exterior container", delete "12 units" and substitute "30 units".

5.1.3.2 Unit package. Delete and substitute:

"5.1.3.2 Unit package. One hundred individual packages shall be packaged in a box of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676 except that commercial colors are acceptable. Closure shall be adequate to prevent spilling of contents under normal handling. Each box shall be overwrapped in clear plastic and the wrap heat sealed."

Page 13:

5.3.2 Unit package. Line 2, delete the words "and each bag".

Page 14:

5.3.3 Exterior container. In subparagraph headed "Note:", line 4, delete "250s" and substitute: "100s".

6.2 Delete FSN's as listed and substitute FSN's in same sequence as listed below.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:24Z
Check the source to verify that this is the current version before use.

Stock No.	Item Identification
6515-181-7403	NEEDLE, HYPODERMIC, Cartridge Type, Disposable, 25 Gage, 13/16 inch, 100s
6515-181-7404	NEEDLE, HYPODERMIC, Cartridge Type, Disposable, 25 Gage, 1-3/8 inch, 100s
6515-181-7412	NEEDLE, HYPODERMIC, Cartridge Type, Disposable, 27 Gage, 1-3/8 inch, 100s
6515-181-7425	NEEDLE, HYPODERMIC, Cartridge Type, Disposable, 27 Gage, 13/16 inch, 100s

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1591

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:24Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:24Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION
MIL-N-36319A (Admendment-1)

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:24Z
Check the source to verify that this is the currenversion before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE [illegible]
2600 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:24Z
Check the source to verify that this is the current version before use.