MILITARY SPECIFICATION

OBJECT DISKS, MICROTOME

1. SCOPE

1.1 Scope. This specification covers object disks, suitable for placement use with an automatic rotary microtome (see 6.1).

1.2 Classification. Object disks shall be of the following sizes, specified (see 6.2):

Size 1. - 7/8-inch diameter.
Size 2. - 1-1/8-inch diameter.
Size 3. - 1-1/2-inch diameter.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of vitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up, Paperboard.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:23Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Military Medical Supply Agency

2066	Object Disk, Microtome.

(Copies of specifications, standards, and drawings required by supplier connection with specific procurement functions should be obtained from the p curing activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. Material shall be brass.

3.2 Style and design. Style and design shall be in strict acco with Drawing 2066. Dimensions shall be as shown on the drawing.

3.3 Finish. Finish shall be chromium over nickel plate.

3.4 Workmanship. Workmanship shall be first class throughout. disks shall be free from defects which detract from their appearance impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified contract or purchase order, the supplier is responsible for the perfo of all inspection requirements as specified herein. Except as otherw specified, the supplier may utilize his own facilities or any commerc laboratory acceptable to the Government. The Government reserves the to perform any of the inspections set forth in the specification wher inspections are deemed necessary to assure supplies and services conf prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, i defined as both examination (such as visual, auditory, etc., investig without the use of special laboratory appliances or procedures) and t (determination by technical means of physical and chemical properties item.

4.1.2 Certificates of quality. When available, certificates of supplied by the manufacturer of the component or material, listing th fied test method and test results obtained, may be furnished in lieu performance of such testing by the contractor.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:23Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted accordance with MIL-STD-105.

	Inspection level	AQL(percent defective)
For visual examination		
Major defects	II	2.5
Minor defects	II	4.0
For dimensional examination	S-1	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance :h MIL-STD-105.

4.3 Examination. The disks shall be examined to determine compliance :h all the requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in :ordance with the following classification of defects:

Categories and defects*

Major	
101	Unit not free of sharp edges.
102	Rod not secured to disk.
103	Soldering not sound; not free of porosity and cracks.
104	Plating does not adhere.
Minor	
201	Soldered areas not smooth.
202	Surface not smooth; not free of pits and burs.

amination shall not be restricted to the classified possible defects ted above.

4.3.2 Dimensional examination. The disks shall be examined for ects in dimensions. Any dimension not within the tolerance specified ein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with :ification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:23Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) object disk, of the size specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	288 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676. Closure shall be adequate to prevent spillage of contents under normal handling.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676. Closure shall be adequate to prevent spillage of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose within the continental limits of the United States of America.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:23Z
Check the source to verify that this is the current version before use.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for
e 2 load, and constructed in accordance with PPP-B-585, style 3, for
ss 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621,
le III; or PPP-B-636, type CF, class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a water-
of case liner conforming to MIL-L-10547. Closure and sealing shall conform
applicable paragraphs of appendix thereto. Case liner shall not be required
fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with
endix of applicable box specification. Closure of each fiberboard box
ll be as specified in the appendix, and waterproofing shall be in ac-
dance with 30.4 of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance
a appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a
eiving activity within the continental limits of the United States for
rage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to
-B-601, table I, using type III, class 1 plywood; or shall be designed
type 2 load, and constructed in accordance with PPP-B-585, style 3,
class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic,
method I closure shall be utilized.

5.2.3 Level C. The commodity shall be packed in substantial
nercial containers of the type, size, and kind commonly used for the
pose, so constructed as to insure acceptance and safe delivery by
ion or other carriers, at the lowest rate, to point of delivery called
in the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:23Z
Check the source to verify that this is the currentversion before use.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The object disks are intended for use with Federal stock number 6640-433-6000, MICROTOME, Rotary, Automatic.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Size required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is curre
as of the date of this docu
For future coordination of
changes to this document, d
circulation should be based
the information in the curr
DODISS.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:23Z
Check the source to verify that this is the currentversion before use.

SPECIFICATION ANALYSIS SHEET

FORM APPROVED BUDGET BUREAU NO. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use of the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity.

CATION MILITARY SPECIFICATION MIL-O-36312(DM), dated 30 October 1964
OBJECT DISKS, MICROTOME

ATION | CITY AND STATE

T NO. | QUANTITY OF ITEMS PROCURED | DOLLAR AMOUNT $

. PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN JREMENT USE?
. GIVE PARAGRAPH NUMBER AND WORDING

. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

NTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

SPECIFICATION RESTRICTIVE?
☐ YES ☐ NO IF "YES" IN WHAT WAY?

S (ATTACH ANY PERTINENT DATA WHICH MAY BE OF USE IN IMPROVING THIS SPECIFICATION. RE ARE ADDITIONAL PAPERS, ATTACH TO FORM AND PLACE BOTH IN AN ENVELOPE ADDRESSED PARING ACTIVITY)

BY (PRINTED OR TYPED NAME AND ACTIVITY) | DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:23Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCURE

ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGE:

DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFF

NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR A

old Under

DEFENSE SUPPLY AGENCY
DEFENSE MEDICAL SUPPLY CENTER
3RD AVENUE AND 29TH STREET
BROOKLYN, N.Y., 11232

OFFICIAL BUSINESS

POSTA
DEFENSE

Defense Supply Agency
Defense Medical Supply Cente
ATTN: Directorate of Technical Operations (Code T-M5)
3rd Avenue and 29th Street
Brooklyn, N.Y., 11232

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:23Z
Check the source to verify that this is the current version before use.