MIL-R-36360C
25 April 1972
SUPERSEDING
MIL-R-36360B
1 September 1970

MILITARY SPECIFICATION

RESUSCITATOR AND ASPIRATOR, INTERMITTENT POSITIVE PRESSURE, MANUAL CYCLING

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers one type of portable resuscitator and aspirator unit complete in a carrying case, suitable for use by medical and paramedical personnel in emergency resuscitation of humans.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-585	Box, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-L-10547	Liners, Case, and Sheet Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129	Marking for Shipment and Storage.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:28Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)

ANSI Z79.2 American Standard Specification for Anesthetic Equipment: Endotracheal Tube Connectors and Adapters.

(Application for copies should be addressed to the American National Standards Institute, 1420 Broadway, New York, N.Y. 10018).

DRAWING

Defense Personnel Support Center (DPSC)

20703 Resuscitator and Aspirator, Intermittent Positive Pressure, Manual Cycling

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions, should be obtained as directed by the contracting officer.)

3. REQUIREMENTS

* 3.1 Material. The material shall be as indicated herein and of highest quality suitable in all respects for the purpose intended. Springs for check valves and the aspirator bellows shall be made of austenitic stainless steel. Metal components other than aluminum in contact with aluminum shall be austenitic stainless steel.

* 3.1.1 Breathing bag. The breathing bag shall be made of suitable elastomeric material as specified for FSN 6515-018-9494, RESUSCITATOR, HAND OPERATED, INTERMITTENT POSITIVE PRESSURE.

* 3.1.2 Non-rebreathing valve housing. The valve housing shall be fabricated from a crystal clear plastic. The patient end of the non-rebreathing valve shall be in accordance with Fig. 5 of ANSI Standard Z79.2 and as specified for 6515-018-9494, RESUSCITATOR, HAND OPERATED, INTERMITTENT POSITIVE PRESSURE.

* 3.1.3 Face mask. Face mask shall be fabricated from natural rubber dipped in conductive neoprene or conductive natural rubber as specified for 6515-018-9494, RESUSCITATOR, HAND OPERATED, INTERMITTENT POSITIVE PRESSURE.

* 3.1.4 Pump base and lock screw. Pump base shall be fabricated from plastic coated steel, austenitic stainless steel, or anodized aluminum with a suitable organic finish.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

3.1.5 Bellows. The bellows shall be fabricated of a suitable neoprene rubber.

3.1.6 Aspirator jar. Aspirator jar shall be fabricated of a suitable plastic; shall not crack or become distorted when tested as specified in 4.4.1.

* 3.1.7 Carrying case. The carrying case shall be fabricated from a 16-gauge vinyl laminated cotton drill, vinyl laminated paperboard, or molded linear polyethylene.

* 3.2 Style, design, and dimensions. The unit shall consist of a hand operated resuscitator complete with face mask, non-rebreathing valve, and compressible bag with inlet valve, an aspirator consisting of a foot operated bellows assembly, an aspirator jar including necessary valving and tubing, and a carrying case which shall provide for storage and transport of both devices. Valving system shall include a by-pass or self-actuating device to prevent application of excess pressure to the patient, from either the oxygen supply or other gas source. The hand operated resuscitator shall conform to the requirements of FSN 6515-018-9494 except that the plastic storage bag shall not be supplied. The unit shall be in accordance with DPSC Drawing 20703.

* 3.2.1 Bellows assembly. The aspirator bellows assembly shall consist of top and bottom plates of aluminum alloy to which are attached the ends of a tubular, accordion pleated neoprene bellows. The two plates shall be held apart by an austenitic stainless steel spring or springs inside the bellows. The maximum separation of the plates shall be limited by an austenitic stainless steel chain. The upper plate shall contain a suitable neoprene disc valve and austenitic stainless steel valve disc retainer. The lower plate shall have a suitable nipple for attachment of the suction hose to the aspirator jar. As an alternate, the aspirator bellows shall be closed at the top end and contain an aluminum plate with a groove for accommodating the bellows spring. At the bottom of the bellows, there shall be the base assembly. The bellows spring shall be mounted over a central protrusion of the base assembly. The bellows spring shall be mounted over a central protrusion of the base assembly. An elastomeric check valve shall be located in the center of the base assembly and there shall be an intake hole leading to an elbow. The aspirator bellows-to-aspirator cup tubing shall be attached to this elbow.

* 3.2.2 Aspirator tubing. The aspirator tubing shall be made of a suitable neoprene or silicone elastomer and the section from the cup to the bellows assembly shall be approximately 10 inches long. The section from the suction nipple on the aspirator jar cap (patient hose) shall be approximately 48 inches long. A short length of polyethylene, polyvinyl chloride or similar suitable tubing of plastic shall be located near the patient end

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:28Z
Check the source to verify that this is the current version before use.

common carrier which will insure safe

of the aspirator tubing so that the operator may see the contents. The patient end shall be made of a suitable aluminum alloy or plastic such as polypropylene and shall be tapered to accept endotracheal tubes and suction catheters, and shall be finished smooth with no sharp edges to permit its use in the human pharynx without danger to soft tissue.

3.3 Performance.

3.3.1 Capacity. Bag shall deliver up to 1200 cc. of air at 20 cm. H_2O upon one compression cycle, when tested in accordance with 4.4.2.

3.3.2 Cycles. Valve action and self-inflating property of bag shall be such that unit will at least operate up to 30 cycles per minute.

3.3.3 Vacuum. The bellows assembly shall be capable of creating a partial vacuum of 12 inches of Hg on each expansion of the bellows, and shall also be capable of developing suction even though the jar or trap is overflowing when tested as specified in 4.4.3.

3.3.4 Overflow. The overflow valves and tubing, together with the bellows assembly, shall function so that suction will be obtained regardless of whether the aspirator jar is full or overflowing when tested as specified in 4.4.3.

3.4 Carrying Case. The carrying case shall be made of material as specified in 3.1.7 and shall be suitable for storage and transport of resuscitator and aspirator. Case shall be readily opened by snap fasteners.

3.5 Instructions sheet. Each unit shall be supplied with an instruction sheet indicating proper use of unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

3.6 Finish. All aluminum or aluminum alloy parts shall be suitably anodized. Other metal parts shall be suitably finished and treated to be corrosion-resistant.

3.7 Service Data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance and lists of component parts. The manual shall comply with the requirements of MIL-M-7298 MANUAL, TECHNICAL, Commercial Equipment. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer, via the cognizant Government inspector, two preliminary copies of his proposed technical manual, for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of contract and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the Contracting Officer. Waiver of submission of 15 additional copies shall be at the option of the Government.

3.8 Identification marking. The apparatus shall be legible and permanently marked with the contractor's name or registered trademark and with apparatus serial number; in addition, resuscitators and aspirators of foreign (other than U.S.) manufacture shall be marked with the name of the country of orgin.

3.9 Workmanship. Workmanship shall be first class throughout. Resuscitator aspirator units shall be free from defects which detract from their appearance or may impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in Table I:

Table I. Sampling for Tests

Component	Characteristic	Rqmt.	Test procedure	Unit of product	Insp. level	AQL
Jar	Autoclaving	3.1.6	4.4.1	1-Jar	S-1	*
Resuscitator	Capacity	3.3.1	4.4.2	1-Resus.	S-1	*
Aspirator	Vacuum	3.3.3	4.4.3	1-Asptr.	S-1	*
	Overflow	3.3.4	4.4.3	1-Asptr.	S-1	*

* Acceptance number shall be zero for all sample sizes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

4.2.3 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The resuscitator-aspirator shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Autoclaving. Jars shall be steam autoclaved for 20 minutes at 250° Fahrenheit (F.) and 15 pounds per square inch gage (p.s.i.), then allowed to cool to room temperature (68° F. to 72° F.).

4.4.2 Capacity test. Capacity shall be tested with a calibrated gasometer or similar apparatus.

4.4.3 Vacuum test. The aspirator shall be tested with a calibrated vacuum gage or manometer. Aspirating jar shall be filled with water during test procedure.

5. PREPARATION FOR DELVIERY

* 5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. One resuscitator and aspirator, as specified, constitute one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table II shall be packaged in each unit and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table II. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	4 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units will be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:28Z
Check the source to verify that this is the current version before use.

5.1.3 Level A.

* 5.1.3.1 Unit package. Unit package shall be a box of appropriate size constructed in accordance with PPP-B-636, type CF, class domestic. Closure shall conform to method I as specified in PPP-B-636.

5.1.4 Level C. Units shall be packaged in standard commerical containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B, or C as specified (see 6.2).

5.2.1 Level A.

* 5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Fiberboard boxes shall conform to the special requirements as specified in PPP-B-636.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

* 5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4 of PPP-B-636.

* 5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

* 5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure of wood boxes shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II of PPP-B-636. Fiberboard boxes shall conform to the special requirements as specified in PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:28Z
Check the source to verify that this is the currentversion before use.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

* 5.3.1 Unit packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number, title and date of this specification.
(b) Applicable levels of packaging and packing (see 5.1 and 5.2).
(c) Federal stock number.

6.2 This specification covers the following item appearing in Federal Supply Catalog:

FSN 6515-890-1818 RESUSCITATOR AND ASPIRATOR, Intermittent Positive Pressure, Manual Cycling.

It does not include all types as indicated by the title nor all types commercially available.

6.3 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1643

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-R-36360C

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

Defense Supply Agency

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.