MIL-S-36315A
23 August 1978
SUPERSEDING
MIL-S-36315(DM)
28 October 1964

MILITARY SPECIFICATION

SPLINT, HAND, MASON-ALLEN, ALUMINUM

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers an aluminum hand splint of the Mason-Allen type for use in the treatment of injuries to the hand.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of the specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

QQ-A-250/1	- Aluminum Alloy 1100, Plate and Sheet.
PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-B-676	- Boxes, Setup.

STANDARDS

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, CODE ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:23Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Personnel Support Center

1960 - Splint, Hand, Mason-Allen, Aluminum.

(Copies of specifications, standards, and drawing required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer).

3. REQUIREMENTS

3.1 Material. The splint shall be fabricated of aluminum alloy 1100-0 in conformance with QQ-A-250/1.

* 3.2 Construction. The splint shall consist of a forearm portion rounded to conform to the convexity of the forearm and terminating in a cup shaped portion, shaped to accurately fit the palmar concavity and the heel of the hand.

3.2.1 Dimensions, style and design. Dimensions, style and design shall be in accordance with Defense Personnel Support Center Drawing 1960.

3.3 Finish. The splint shall be given a mirror or "Alumilite" finish.

3.4 Identification marking. The splint shall be permanently and legibly marked with the name or registered trademark of the manufacturer.

* 3.5 Workmanship. Workmanship shall be first class throughout. Splints shall be free from defects which detract from their appearance or impairs their serviceability. All edges of the splint shall be smoothly rounded. All surfaces shall be free from burrs and scratches.

* 4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:23Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the aluminum may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one splint.

TABLE I. Sampling for examination.

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	1.0
Minor defects	II	2.5
For dimensional examination	S-2	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105. The inspection level shall be S-2 and the acceptance number shall be zero. The unit of product for sampling purposes shall be one splint.

4.3 Examination. The hand splint shall be examined to determine compliance with all requirements contained in this specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:23Z
Check the source to verify that this is the current version before use.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table II. Examination shall not be restricted to the classified possible defects listed below.

TABLE II. Classification of defects.

Categories	Defects
Major	
101	Splint not free of burrs, scratches or sharp edges.
102	Splint warped or deformed.
103	Surface finish not uniform, corroded, tarnished, or pitted.
Minor	
201	Splint not free of dirt, grease or foreign material.
202	Splint discolored or stained.
203	Marking incomplete, illegible or missing.

4.3.2 Dimensional examination. The splint shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Packaging inspection. The inspection of the unit package, intermediate package, packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.5 Tests. Tests shall be conducted to determine compliance with specification requirements.

* 5. PACKING

5.1 Packaging. Packaging shall be level A or C, as specified (see 6.2).

5.1.1 Unit of issue. One splint constitutes a unit of issue.

5.1.2 Unit packaging. Each splint shall be wrapped in kraft paper or other suitable barrier material. The wrap shall be secured with tape or by other suitable means. As an alternate, each splint may be shrink wrapped with plastic film.

5.1.2.1 Intermediate package. Twelve unit packages shall be packaged in a box conforming to PPP-B-566, PPP-B-636, class domestic or PPP-B-676. Closure shall be as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:23Z
Check the source to verify that this is the current version before use.

5.1.3 Level C. Splints shall be clean, dry and packaged in a manner that will afford adequate protection against corrosion, deterioration and physical damage during direct shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Six intermediate packages shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

5.2.2 Level B. Six intermediate packages shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The packaged splints shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the carriers' rules and regulations.

5.3 Marking. Each unit package, intermediate package and shipping container shall be marked as specified in MIL-STD-129.

* 6. NOTES

6.1 Intended use. The hand splint covered by this specification is intended for use in immobilizing the hand in the treatment of hand injuries.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) National Stock Number (NSN).
(c) Selection of applicable levels of packing and packaging (see 5.1 and 5.2).

6.3 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6515-00-372-2150	SPLINT, HAND, MASON-ALLEN, ALUMINUM

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:23Z
Check the source to verify that this is the currentversion before use.

6.4 This specification does not include all types, classes, grades or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the type normally procured to meet military requirements.

6.5 Changes from previous issue. The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assume no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DoD - MB

Preparing activity:
Defense Logisitcs Agency - DM

Project number: 6515-2131

☆ U.S. GOVERNMENT PRINTING OFFICE: 1978—703-122/4228

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:23Z
Check the source to verify that this is the currenversion before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

POSTAGE AND FEES PAID

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS
DEFENSE PERSONNEL SUPPORT CENTER
DIRECTORATE OF MEDICAL MATERIEL CODE: ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:23Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS: This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.

NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER *(Number)* AND TITLE
MIL-S-36315A SPLINT, HAND, MASON-ALLEN, ALUMINUM

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1. ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*

TELEPHONE NO.

DATE

DD FORM 1 OCT 76 1426 EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:23Z
Check the source to verify that this is the current version before use.