MIL-S-36362(DM)
18 January 1965

MILITARY SPECIFICATION

STETHOSCOPE, DeLee-Hillis

1. SCOPE

1.1 This specification covers stethoscopes used to convey ear sounds produced by the body.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
ZZ-T-831	Tubing, Rubber.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed And Lock Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up, Paperboard.

Military

MIL-L-10547	Liners, Case and Sheet, Overwrap, Water-Vaporproof, or Waterproof, Flexible.

STANDARDS

Federal

Fed. Test Method Std No. 601 Rubber; Sampling and Testing.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129	Marking for Shipment and Storage.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z.
Check the source to verify that this is the current version before use.

DRAWING

Defense Medical Supply Center
20375 Stethoscope, DeLee-Hillis

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Brass. Brass shall be suitable for the purpose intended and shall comply with plating and corrosion tests of GG-I-526 for class 4, composition N material.

3.1.2 Carbon steel. Carbon steel shall be tool-steel quality, heat-treated to spring temper, and shall comply with the plating and corrosion tests of GG-I-526 for class 1, composition A material.

3.1.3 Rubber tubing. Rubber tubing shall be black commercial gum rubber tubing conforming to the applicable requirements of ZZ-T-831 for grade A tubing except that the specific gravity shall be a maximum of 1.40. The breaking strength and elongation of the tubing shall not decrease more than 25 percent nor shall the tubing crack, or become tacky or brittle when subjected to the accelerated aging test specified in 4.4.1.

3.1.4 Plastic ear tips. Plastic shall be black, non-shattering and shall be suitable for the purpose intended. Plastic shall be non-toxic when used for the purpose intended, and shall be free from objectionable odor.

3.2 Style, design and construction. Stethoscope shall be an obstetric type of the DeLee-Hillis design with adjustable metal headband suitable for determining foetal heart beats without use of operator's hand. Style, design and construction shall be as shown in Drawing 20375.

3.3 Identification marking. The stethoscope shall be permanently and legibly marked with the manufacturer's name or registered trademark.

3.4 Workmanship. Workmanship shall be first class throughout. stethoscopes shall be free from defects which detract from their appearance or may impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commerical laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

4.2.2 For tests.

4.2.2.1 For rubber tubing. Sufficient samples shall be selected from each lot to perform all tests. Failure of any sample to comply with any test shall be cause for rejection of the lot represented.

4.2.2.2 For other tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, inspection level S-1, except that the acceptance number shall be zero (0) for all sample sizes.

4.2.3 Raw materials. Four ounces of each lot of each metal used in the construction of the stethoscopes shall be selected for chemical analysis.

4.2.4 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The stethoscopes shall be examined to determine compliance with all requirements contained in this specification.

Source: https://assist.dla.mil - Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Accelerated aging test. The rubber tubing shall be subjected to the air heat test in accordance with method 7221 of Standard No. 601.

PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) stethoscope, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit and exterior container, as applicable, for the required level of protection specified in procurement document.

Table I. Packaging Quantities

Unit package	Exterior container
1 unit	24 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in a carton of appropriate size constructed in accordance with [redacted] PPP-B-566, or PPP-B-676, except that commercial colors shall be acceptable; or PPP-B-636. Closure shall be adequate to prevent spillage of contents under normal handling.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with [redacted] PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

3.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix of PPP-B-636. In addition, waterproofing shall be as specified in 30.4.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic. Closure of fiberboard boxes shall be in accordance with Method I of appendix to box specification.

5.3 Marking.

5.3.1 Interior packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Custodians:
Army. - MD
Navy. - MS
Air Force. - 03

Preparing activity
Defense Supply Agency-DM

Review activities:
Army. - Md
Navy. - MS
Air Force. - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.