MIL-T-36331A
25 June 1971
SUPERSEDING
MIL-T-36331(DM)
13 November 1964

MILITARY SPECIFICATION

TUBE, DRAINAGE, SURGICAL, PLASTIC, UROLOGICAL, DISPOSABLE, 9/32 INCH

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a urinary drainage tube suitable for use as a connecting tube between a urethral or urinary catheter and a drainage container.

2. APPLICABLE DOCUMENTS

* 2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Box, Set-Up.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:24Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

20777	Tube, Drainage, Surgical, Plastic Urological, Disposable, 9/32 inch.

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

U. S. PHARMACOPEIAL CONVENTION, INC.

Pharmacopeia of the United States.

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pa. 18042.)

3. REQUIREMENTS

3.1 Material.

* 3.1.1 Plastic tubing. Plastic tubing shall be transparent, medical-grade polyvinyl, and sterile.

3.1.2 Plastic connector. Plastic connector (Type II), when used, shall be fabricated of a material suitable for the purpose intended, and shall be sterile when tested as specified in section 4.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:24Z
Check the source to verify that this is the current version before use.

3.1.3 Rubber cap. Rubber cap shall be of the highest grade latex rubber, be suitable in all respects for the purpose intended, and shall be sterile when tested as specified in section 4.

* 3.2 Style, design and dimensions. The drainage tube shall be either with integral connector (Type I), or, with an attached connector (Type II), in accordance with Drawing 20777 of the Defense Personnel Support Center, Directorate of Medical Materiel, Phila., Pa. and as hereinafter specified.

* 3.2.1 Type I. Type I drainage tube shall be fabricated with a connector which is an integral part of the plastic tubing (formed from this tubing). The tube shall comply with the requirements of 3.1.1, 3.1.3, 3.2, 3.3 and 3.7.

* 3.2.2 Type II. Type II drainage tube shall be fabricated with an attached connector. The tube shall comply with the requirements of 3.1.1, 3.1.2, 3.1.3, 3.2, 3.3, 3.6 and 3.7.

3.3 Sterility. Tubes shall be sterile when tested for sterility as specified in 4.4.1.

* 3.4 Certificates of sterility, listing all sterility lot numbers, stating that samples from each sterility lot have been tested for sterility and found to be sterile, shall be furnished by the contractor. Certificates shall specify the tests performed and the results obtained. The certificates of sterility shall be submitted by the contractor directly to Technical Operations Division, Directorate of Medical Materiel, ATTN: Quality Assurance Branch, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101.

* 3.5 Manufacturer's record. The manufacturer shall place a code or lot number on the paper label, or on the printed carton, which can be verified against the manufacturer's records of all tests made on sterile tubes. These records shall be made available for inspection by any duly authorized Government representative. No shipment shall be made until all tests have been completed, verified, and recorded in the permanent records of the manufacturer.

* 3.6 Security of assembly. The plastic connector (Type II) and tubing connection shall withstand a 3-pound load for 10 minutes when tested as specified in 4.4.2.

3.7 Instruction sheet. Each tube shall be supplied with an insert providing instruction for proper connection of tube to catheter and drainage container.

Source: https://assist.dla.mil -- Downloaded: 20261O-21T08:24Z
Check the source to verify that this is the current version before use.

3.8 Workmanship. Workmanship shall be first class throughout. Tubes shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

* 4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

* 4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I. The unit of product for purposes of examination shall be one drainage tube.

Table I. Sampling for Examination

Type of defect	Inspection level	AQL (percent defective)*
Major	I	1.0
Minor	S-3	2.5

* If acceptance number is not exceeded, lot shall be acceptable. If rejection number is equaled or exceeded, inspection at level II may be performed to determine final acceptance or rejection.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:24Z
Check the source to verify that this is the current version before use.

* 4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in Table II. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Table II. Sampling for Tests

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Tubing-connector	Security of assembly (Type II)	3.6	4.4.2	S-2	1.0
Unit of product	Sterility	3.3	4.4.1	S-2	*

* Acceptance number shall be zero for all sample sizes.

* 4.2.2.1 For Sterility. Sampling procedures for sterility shall be conducted as indicated in Table II.

4.2.2.2 For security of assembly. Sample size for test shall be as indicated in Table II. The unit of product for inspection purposes shall be one tube.

* 4.3 Examination. The tubes shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted in accordance with MIL-STD-105 and the classification of defects in Table III. Examination shall include but shall not be limited to the defects listed in Table III.

Table III. Classification of Defects

Categories	Defects
Major	
101	Defective connector.
102	Dimensions not as specified.
103	Permanent kinks in tubing restricting flow.
Minor	
201	Presence of foreign material, loose particles.
202	Instruction sheet not present.

* 4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:24Z
Check the source to verify that this is the current version before use.

* 4.4.1 Sterility. Testing for sterility shall be in accordance with the method specified in the U.S.P. Procedural Details for Sterility Tests of Class B, Type II, Inoculated Products, and for Parenteral Devices. As an alternate, testing for sterility shall be in accordance with the method specified in the U.S.P. Procedural Details for Sterility Tests, Class B, Type II, Inoculated Carrier, and Parenteral Devices, except that the inoculated carrier shall contain sterilization test strips (MIL-S-36586) carrying spores of Bacillus subtilis var. globigii (niger) and Bacillus stearothermophilus.

* 4.4.2 Security of assembly. The tubing shall be cut 10 inches from the connector. The security of assembly of the connector shall then be tested by gradually applying a 3-pound weight to the severed end of the tubing while the connector is held in a clamping device so that the tubing hangs vertically downward. The load shall remain applied for 10 minutes without the tubing separating from the connector.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. Fifty (50) tubes, as specified, constitute one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 Unit	Not required	4 Units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:24Z
Check the source to verify that this is the current version before use.

5.1.3.1 Unit packaging.

5.1.3.1.1 Individual package. Each tube shall be formed into a coil of such diameter that it will not kink or be permanently deformed, and shall be packaged in a bag fabricated from seamless polyethylene tubing having an average wall thickness of not less than 0.002 inch, with no wall thickness at any point below 0.00125 inch. Excess air shall be expelled or evacuated, and closure effected by heat sealing. The completed package shall be adequate to maintain sterility in accordance with the requirements of 3.3.

5.1.3.1.2 Unit package. Unit package shall be a box of appropriate size and design constructed in accordance with PPP-B-566 or PPP-B-676, except commercial colors will be acceptable, or PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent accidental opening under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

* 5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed in accordance with paragraph 30.4 of PPP-B-636.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:24Z
Check the source to verify that this is the current version before use.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure of wood boxes shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II of PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

* 5.3.1 Individual package. Each individual package shall be marked with lot (control) number, the name of the manufacturer, commercial description of contents, the legend "Sterility not guaranteed if package is broken or damaged", (or similar statement).

5.3.2 Unit package. Each unit package shall be marked in accordance with MIL-STD-129. Lot (control) number, contract or purchase order number and name of contractor shall be shown. In addition, the method of sterilization (G-Gas, R-Radiation, S-Steam, D-Dry Heat sterilization) shall be shown following the level of packaging and the date of packaging. Example: "A-9/71-G". Final sterilization shall be on or after indicated packaging date.

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Lot (control) number and the method of sterilization as specified in 5.3.2 shall also be shown.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(c) FSN number.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:24Z
Check the source to verify that this is the current version before use.

6.2 This specification covers the following item as it appears in the Federal Supply Catalog as FSN 6515-042-3850 TUBE, DRAINAGE, SURGICAL, Plastic, Urological, Disposable, 9/32 inch, 50s. It does not cover all types, classes, sizes or grades as indicated by the title nor all types which are commercially available.

* 6.3 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the previous issue.

6.4 Supersession data. This specification includes requirements of Defense Medical Purchase Description No. 4, dated 27 March 1969 and Modification No. 1, dated 29 March 1969.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1597

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:24Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:24Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

S/N 0102-014-1802

Source: https://assist.dla.mil -- Downloaded: 2026-10-24T00:24Z
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

Commander
Defense Personnel Support Center
Directorate of Medical Material
Attn: DPSC-ATTS
2800 South 20 th St.
Philadelphia, Pa. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026l0-21T08:24Z
Check the source to verify that this is the currentversion before use.