MIL-T-36348(DM)
December 8, 1964

MILITARY SPECIFICATION

TABLE, PATIENT EXAMINING, ADJUSTABLE, MOTOR DRIVEN

1. SCOPE

1.1 This specification covers an electrically-hydraulically powered patient examining and treatment table suitable for eye, ear, nose and throat, varicose vein, proctological, obstetrical, urological, gynecological, and routine horizontal examinations in military medical treatment facilities.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids, or request for proposal form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

UU-P-271	Paper, Wrapping, Waterproofed Kraft.
CCC-A-70	Artificial Leather, Cloth, Coated, Vinyl Resin (Upholstery).
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.

Military

MIL-R-17252	Rubber, Cellular (Bonded Shredded).

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:26Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING MATERIALS

ASTM Designation D1564-62T	Specification and Methods of Tests for Flexible Urethane Foam.

(Copies of ASTM Standards may be obtained upon application to the American Society for Testing Materials, 1916 Race Street, Philadelphia 3, Pa).

American Standards Association, Inc.

C73.1-1961	Outlet Receptacles, Attachment Plug Caps, and Appliance Plugs.

(Application for copies should be addressed to American Standards Association, Inc., 10 East 40th Street, New York, N.Y. 10016).

UNDERWRITERS' LABORATORIES, INC.

Underwriters' Laboratories Standards.

(Underwriters' Laboratories Standards are available from the Underwriters' Laboratories, Inc., Walt Whitman Road, Melville, L. I., N. Y. 11749; 207 E. Ohio Street, Chicago 11, Illinois; and 1655 Scott Lane, Santa Clara, Calif.)

3. REQUIREMENTS

3.1 Fire and casualty hazards.

3.1.1 The contractor shall submit to the contracting agency prior to delivery proof that the equipment he will supply under this specification conforms to the applicable requirements of Underwriters Laboratories Inc. standards. Such proof shall consist of a laboratory report or listing by the Underwriters' Laboratories, Inc. The label or listing of the Underwriters' Laboratories shall be acceptable as evidence that the equipment conforms to the above. A copy of each laboratory report shall be furnished to the contracting officer.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:26Z
Check the source to verify that this is the current version before use.

3.1.2 In lieu of the label or listing by the Underwriters' Laboratories Inc., the contractor may submit a laboratory report by a nationally recognized testing agency adequately equipped and competent to perform such services as shall be evidenced by the attachment of its seal, stamp or label to the article, or the issuance of a certificate. The contractor shall submit to the laboratory a report which shall include test results and methods used to determine compliance, to the contracting officer for approval. The Government reserves the right to require a facilities report showing the qualifications of the testing laboratory and to conduct a facilities survey to determine its acceptability.

3.1.3 Compliance with the above requirements regarding fire and casualty hazards, does not absolve the contractor from complete compliance with the requirements of this specification in order to secure the acceptance of his commodity.

3.2 Material.

3.2.1 Plastic foam. Plastic foam used in the fabrication of the patient examining table shall be uncored virgin or shredded and bonded polyurethane having the following properties:

Cell count	40 to 55 cells per inch
Density	2.00 ± 10% pounds per cubic foot
Indentation load (25%)	30 to 50 pounds
Deflection (65%)	60 to 100 pounds
Compression load (25%)	0.40 to 0.60 pounds per square inch
Deflection (65%)	0.55 to 0.80 pounds per square inch
Resilience (ball rebound)	25 - 35 percent
Tear strength	4 to 6 pounds per inch
Tensile strength	25 to 40 pounds per square inch
Tensile elongation	300 to 500 percent

Tests to determine the physical properties of plastic foam shall be conducted as specified in 4.4.4.

3.2.2 Cellular rubber. Cellular rubber shall confrom to MIL-R-17252 for type I or II, class 2, medium condition.

3.2.3 Artificial leather. Artificial leather used for upholstery shall be regular fire resistant coated cloth conforming to class 2 of CCC-A-700.

3.3 Construction. Table shall consist of a base and pedestal, elevating column, 5-section adjustable table top, the necessary mechanisms for positioning and control, and the accessories specified herein. The table shall be designed for ease of maintenance and cleaning, and to prevent seepage of oil to the table

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:26Z
Check the source to verify that this is the current version before use.

exterior, or the entrance of liquids or dirt into any of the internal mechanisms.

3.3.1 Base. Base shall be of suitable metal and rectangular in shape. Base shall be designed so as not to obstruct or interfere with the physician's toes when standing in a normal working position on either side of the table. Shall be of sufficient size and weight to assure that the table is stable throughout the entire range of positioning and adjustment as specified herein, when supporting a uniformly distributed 300 pound load. The base and pedestal mounted thereon shall enclose an electric motor driven hydraulic pump which, with necessary power cylinders, solenoid valves, linkages, and controls, shall provide for continuously controlled power raising, lowering and tilting of the table top.

3.3.2 Elevating column. The elevating column enclosed in the pedestal shall satisfactorily raise and lower the table top and shall be actuated by means of a foot pedal controlled electro-hydraulic system. The range of adjustment of the table shall be from a minimum of not more than 27 inches to a maximum of not less than 42 inches. Pedals shall be suitably mounted in one or both sides of the base and sufficiently durable to satisfactorily withstand normal use. When in the fully raised position and supporting the test load, shall not settle more than 1/4 inch in 8 hours when tested as specified in 4.4.2.

3.3.3 Table top. Table top shall be suitably and securely mounted on or attached to the elevating column and shall include a head rest section, back section, seat section, leg section, and foot and knee section. All sections shall remain firmly in position throughout the entire range of positioning and shall be of durable material, covered with not less than 1-1/2 inches of padding made of plastic foam or cellular rubber. All sections shall be upholstered with artificial leather. With all sections parallel to the floor and the foot and knee and head rest sections fully extended and in the same plane as the table top, the overall length of the table top shall be at least 78 inches. Shall be capable of 360° rotation clockwise and counter clockwise around the vertical axis of the elevating column. The table top shall be free to rotate to any position within the 360° rotation at no greater than 15° increments.

3.3.3.1 Headrest section. Headrest section shall be nominally 17 inches wide and 8 inches long. Shall be securely attached to the head end of the back section by means of two hinged metal members provided with friction locks. The friction locks shall be positive acting and easily operated. The friction locks shall permit the head section to be adjusted vertically and horizontally through a wide range of positions, from perpendicular to and above the back section, to parallel with and below the plane of the back section. There shall be no evidence of slippage, weakness or strain on the friction locks when tested as specified in 4.4.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:26Z
Check the source to verify that this is the current version before use.

3.3.3.2 Back section. Back section shall be nominally 23 inches wide and the combined total length of the back, seat, and leg sections in the horizontal position, shall be nominally 58 inches. Back section shall be suitably counterbalanced so that it can be easily adjusted upward (manually) to 90° above the plane of the seat section, and shall be capable of being securely positioned at any point throughout its range of adjustment. Shall be designed for use with side rails. Shall have dual lock and release handles easily accessible from both sides of the table and located so as not to extend beyond the sides and front of the back section.

3.3.3.3 Seat section. Seat section shall be nominally 23 inches wide, and pivotally attached so that the section may be tilted electro-hydraulically through the use of the foot control pedals. The range of adjustment shall be from foot end of the section 45° above horizontal to foot end of section 15° below horizontal. Shall have an easily removable perineal cutout section centered at the foot end of the seat section. Cut out section shall be padded and upholstered the same as other table top sections. Shall be supplied with a sliding, removable stainless steel or aluminum drain tray supported by stainless steel guides or tracks. Shall have two folding, adjustable heel stirrups of chrome plated malleable iron or brass attached at the foot end of the seat section. One stirrup shall be securely attached under each side of the seat section, and designed to permit folding out of the way and underneath the sides of the section when not in use. When extended for use the stirrups shall be easily adjustable to meet individual patient needs. Seat section shall be designed to permit the use of siderails.

3.3.3.4 Leg section. Leg section shall be nominally 20 inches wide, and counter-balanced for easy manual adjustment downward 90° in relation to the seat section. Shall be lowered and positioned at any point throughout its range of adjustment by means of dual release handles conveniently located underneath the leg section. Handles shall be easily accessible from both sides with the table in any position, and shall not protrude beyond the sides or ends of the leg section. A siderail shall be securely fastened to each side of the leg section to permit mounting of the foot and knee rest section. Siderails shall be of cold rolled carbon steel, chrome plated, and at least three-quarters the length of the leg section.

3.3.3.5 Foot and knee rest section. Foot and knee rest section shall be 20 inches by 8 inches, nominal dimensions, for use as a foot rest and a knee rest. Shall be designed to permit easy removal and attachment to the siderails of the leg section, and positioning at least at right angles to, and in a plane with the leg section. Shall be provided with a mechanical means to positively lock the foot and knee rest at right angles to the leg section at various positions along the leg section siderails. Positioning normal to the leg section shall provide for adjustment as may be required to match leg length of a patient.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:26Z
Check the source to verify that this is the currentversion before use.

Shall safely support the test load without tipping or distortion of any part of the table.

3.3.4 Controls. Control and position mechanisms provided except for foot controls, shall be easily operable from both sides of the table. The foot operated controls shall be positioned as previously described and shall actuate the electro-hydraulic mechanism and permit tilting the table top, returning the table top to the horizontal position, and raising and lowering the table top vertically. Hand operated controls as previously described shall permit manual rotation of the table top, and positioning of the head rest, back, seat, leg, and foot and knee rest sections. It shall be possible to operate the table easily, without binding, jerking, or evidence of stress when tested as specified in 4.4.2.

3.3.5 Electrical characteristics. Table shall operate satisfactorily on 110 volt, 60 cycle, AC, and shall be furnished with a three conductor power cord not less than 10 feet in length. The cord shall be provided with an attachment plug suitable for use with receptacles described in ASA Standard C73.1-1961 for Caps and Receptacles, 125 Volt, 15 Amperes, 2 Pole, 3 Wire Grounding Type (see 2.2).

3.3.5.1 Dielectric strength. There shall be no electrical breakdown when tested as specified in 4.4.3.

3.3.6 Accessories. Durable side rails of corrosion resisting steel for use with back and seat sections shall be supplied.

3.4 Finish. Corrosion resisting steel shall be given No. 4 commercial finish. All other metal parts shall be finished in two coats of enamel baked on between each coat in accordance with paint manufacturer's instructions. Color finish of enamel shall be silver lustre or charcoal.

3.5 Identification marking. Each table shall be fitted, in a conspicuous location, with a nameplate that shall contain the following data:

Name or registered trademark of manufacturer.
Manufacturer's type designation.
Serial number.
Electrical rating.
Stock number.
Contract or purchase order number.
U.S.

Markings secured by adhesives shall not be acceptable.

3.6 Service data. The contractor shall furnish, with each unit, two copies of service data containing complete instructions for installation, operation, maintenance, and a list of component parts. The service data shall include such detailed information which shall permit ready installation and maintenance of the item. All information contained in the manual shall be

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:26Z
Check the source to verify that this is the current version before use.

suitably indexed to permit ready reference to the required information. Wiring diagrams shall be included for electrical items or components. Wiring diagrams, drawings, and descriptive material shall include adequate information so that the replacement parts can be ordered by the manufacturer's part number, and without confusion as to the item identification. Identical or interchangeable replacement parts shall be identified by a single part number only. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the contracting officer, via the cognizant Government inspector, two preliminary copies of his proposed service data, for review and comment by the Defense Medical Supply Center. The preliminary service data shall be submitted sufficiently early to permit adequate review by the Defense Medical Supply Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of required revisions and corrections in the final service data without delaying delivery under the contract or purchase order. Upon approval, the contractor shall mail 15 copies of the revised service data to the contracting officer for distribution.

3.7 Workmanship. The workmanship, finish, safety, and detail of apparatus shall be first class in every respect. Tables shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:26Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105.

Component	Characteristic	Requirement	Test Procedure	Inspection Level	AQL
Headrest	Friction locks	3.3.3.1	4.4.1	S-4	*
Elevating column	Load Test	3.3.2	4.4.2	S-4	*
Foot & knee rest section	Load Test	3.3.3.5	4.4.2	S-4	*
Table (controls)	Load test	3.3.4	4.4.2	S-4	*
Table	Dielectric strength	3.3.5.1	4.4.3	100%	-

*Regardless of sample size, acceptance number shall be zero (0).

4.2.3 Components. Sufficient samples of the components shall be selected from each lot of material used in the manufacture of examining tables to perform tests as required by the applicable subsidiary specifications or by tests specified herein. When sample size for components covered by the subsidiary specifications is not specified, the sample size for each test shall be in accordance with inspection level S-2. No test result shall fall below the minimum or exceed the maximum values specified. The sample unit for testing shall be quantity necessary to perform all required tests one time each. Lot sizes are expressed as follows:

Polyurethane----------square feet
Cellular rubber----------yards
Artificial leather----------square yard

4.3 Examination. The table shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Headrest locks. With the head rest section in horizontal position, sand bags totaling 60 pounds in weight shall be evenly distributed and secured

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:26Z
Check the source to verify that this is the current version before use.

to the head rest section. The section shall be repeatedly adjusted to all possible positions and examined for evidence of slippage, weakness or strain on the friction locks.

4.4.2 Load test. The table shall be subjected to load tests as follows:

a. With the 5 section tabletop in the horizontal position and with an evenly distributed 300 pound load, electro-hydraulically raise and lower the tabletop from 27 inches to at least 42 inches and allow to stand for 8 hours. The tabletop shall then be measured from the top of the cushion to the floor to determine the amount of settling.

b. With the tabletop in the horizontal position and approximately 27 inches above the floor, supporting an evenly distributed 300 pound load, electro-hydraulically tilt the table to 45° trendelenberg and 15° reverse trendelenberg 10 times.

c. With the tabletop in the horizontal position and the stirrups in position, manually angulate the leg section downward 90° in relation to the seat section.

d. With the table in the proctological position, that is, with the back and seat sections horizontal, the leg section down 90° (the foot and knee rest normal to the leg section) and the headrest positioned to provide head and arm support, uniformly distribute a 300 pound load and electro-hydraulically tilt the table to 45° trandelenberg position.

e. With the table in the chair position that is, seat section horizontal, back section 90° upward with respect to the seat section, leg section 90° downward with respect to the seat section, the head, foot-and-knee rest section adjusted for a patient, a 300 pound load distributed on the seat and back sections, raise and lower the table, and then tilt the table to 45° trendelenberg electro-hydraulically.

4.4.3 Dielectric strength. Following the load test (4.4.2), the table shall be subjected to a gradually increasing 60-cycle potential of 900 volts, which shall be maintained at this value for a period of one minute.

4.4.4 Plastic foam. Plastic foam shall be tested in accordance with procedures outlined in ASTM D1564-62T.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:26Z
Check the source to verify that this is the currentversion before use.

5. PREPARATION FOR DELIVERY.

5.1 Packaging.

5.1.1 Unit of issue. One (1) table, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.3 Unit preparation. Each unit shall be disassembled as far as is practicable. All exterior finished surfaces shall be protected by wrapping with kraft paper and cushioning with excelsior pads, flexible corrugated fiberboard or other suitable materials. Wrapping and cushioning material shall be secured by tying with cotton cord or tape, or by sealing with kraft gummed paper tape. Work shall be accomplished in such a manner as to insure against marring, scratching or damage to the unit.. Disassembled parts and accessories shall be adequately protected and packaged in one or more cartons of suitable size and design. Cartons shall be secured to prevent accidental opening.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Each unit shall be overpacked in an exterior container designed for type 3 load, and constructed in accordance with PPP-B-601, table II, using type III, class 1, plywood or PPP-B-621, table IV. The unit shall be adequately secured within the exterior container by means of bolts, wood braces, or other interior fittings, to insure against shifting or breakage. Each box shall be lined with waterproof barrier material conforming to UU-P-271. The exterior container design shall include three skids of not less than 1-1/2 inches in thickness applied crosswise of base, so positioned as to facilitate handling with a forklift truck and to maintain unit in an upright position. Closure and strapping shall comply with appendix of applicable specification.

5.2.2 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each container shall be marked to identify contents.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Directional arrows in accordance with 5.6.7.3 of MIL-STD-129 shall be applied to indicate correct position for stowage.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:26Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

a. Number and date of this specification.

b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Custodians:
Army -- MD
Navy - MS
Air Force-03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:26Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:26Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

FORM APPROVED BUDGET BUREAU NO. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use of the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity.

SPECIFICATION MIL-T-36348(DM) TABLE, PATIENT EXAMINING, ADJUSTABLE, MOTOR DRIVEN
Dated 8 December 1964

ORGANIZATION | CITY AND STATE

CONTRACT NO | QUANTITY OF ITEMS PROCURED | DOLLAR AMOUNT $

MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?
   ☐ YES ☐ NO IF "YES" IN WHAT WAY?

4. REMARKS (ATTACH ANY PERTINENT DATA WHICH MAY BE OF USE IN IMPROVING THIS SPECIFICATION. IF THERE ARE ADDITIONAL PAPERS, ATTACH TO FORM AND PLACE BOTH IN AN ENVELOPE ADDRESSED TO PREPARING ACTIVITY)

SUBMITTED BY (PRINTED OR TYPED NAME AND ACTIVITY) | DATE

DD FORM 1426

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:26Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold Under

DEFENSE SUPPLY AGENCY
DEFENSE MEDICAL SUPPLY CENTER
3RD AVENUE AND 29TH STREET
BROOKLYN, N.Y., 11232

OFFICIAL BUSINESS

POSTAGE FEES PAID
DEFENSE SUPPLY AGENCY

Defense Supply Agency
Defense Medical Supply Center
ATTN: Directorate of Technical Operations (Code T-M5)
3rd Avenue and 29th Street
Brooklyn, N.Y., 11232

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:26Z
Check the source to verify that this is the current version before use.