MIL-E-36806
AMENDMENT-3
23 October 1974
SUPERSEDING
AMENDMENT-2
24 July 1972

MILITARY SPECIFICATION

ERYSIPHAKE, BELL

This amendment forms a part of Military Specification MIL-E-36806, dated 21 December 1971, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 2:

* 3.3 In second line delete "B75" and substitute "B65".

3.4 At end of paragraph delete period (.) and add the following: "except delete dimension "1-7/8 ± 1/8" and substitute "1-7/8 + 1/8 - 3/16", delete dimension "1-3/8 ± 1/8" and substitute "1-3/8 + 1/8 - 3/16", delete dimension "3/16 ± 1/32" and substitute "3/16 + 3/32 - 1/32", and as an alternate the cannula of the Bell erysiphake shall be threaded into the cup portion and then silver soldered or brazed."

* 3.6 Delete requirement for marking "stainless".

Add the following new paragraph:

* "6.5 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the previous issue."

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1889

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:22Z
Check the source to verify that this is the current version before use.