MIL-F-36801A
20 May 1987
SUPERSEDING
MIL-F-36801
2 November 1971

MILITARY SPECIFICATION

FORCEPS, INTESTINAL (Doyen)

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a Doyen model forceps suitable for use in gastro-intestinal tract surgery.

1.2 Classification. Forceps shall be of the following types as indicated below (see 6.2).

Type

1. Curved.
2. Straight.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. The following specifications and standards form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-RST, 2800 S. 20th Street, Philadelphia, PA 19101, by using self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:21Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

FEDERAL

GG-I-526	- Instruments, Dental and Surgical; General Specifications.
PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-601	- Boxes, Wood, Cleated-plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-corner.
PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-B-676	- Boxes, Set-up.

MILITARY

MIL-P-116	- Preservation-Packaging, Methods of.
MIL-L-10547	- Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.
MIL-STD-147	- Palletized Unit Loads.
MIL-STD-794	- Parts and Equipment, Procedures for Packaging of.

(Copies of specifications and standards required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of the documents which are DoD adopted shall be those listed in the issue of the DoDISS specified in the solicitation. Unless otherwise specified, the issues of documents not listed in the DoDISS shall be the issue of the nongovernment documents which is current on the date of the solicitation.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:21Z
Check the source to verify that this is the current version before use.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM) STANDARDS

E 18 - Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials, Tests for.

E 92 - Vickers Hardness of Metallic Materials, Tests for.

E 140 - Hardness Conversion Tables for Metals (Relationship Between Brinell Hardness, Vickers Hardness, Rockwell Hardness, Rockwell Superficial Hardness and Knoop Hardness).

F 899 - Stainless Steel Billet, Bar and Wire for Surgical Instruments.

F 921 - Standard Definition of Terms Relating to Hemostatic Forceps.

F 1026 - General Workmanship and Performance Measurements of Hemostatic Forceps.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race St., Philadelphia, PA 19103.)

(Nongovernment standards and other publications are normally available from the organizations which prepare or which distribute the documents. These documents also may be available in or through libraries or other informational services).

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence. Nothing in this specification, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 General requirements. The forceps shall be in accordance with the applicable requirements of ASTM F1026 and as specified herein. The terms used to describe the components and portions of the instruments described in this specification shall be in accordance with ASTM F921, and as specified herein.

3.2 Material. The forceps shall be fabricated from Class 4, martensitic stainless steel in accordance with ASTM F899.

3.3 Hardness. The forceps shall have a Rockwell hardness of C43 to C48 when tested as specified in 4.4.1. (with the exception of the screw or pin).

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:21Z
Check the source to verify that this is the current version before use.

3.4 Style, design and dimensions. The forceps shall be a Doyen model. The forceps shall be long, ring handled instruments with flexible, serrated blades, used for gastro intestinal tract surgery. Both jaws of the forceps shall be serrated throughout their entire length. The forceps shall be in accordance with Figures 1 and 2.

3.5 Locks. The forceps shall have a box type lock. Performance shall be as specified in ASTM F1026.

3.6 Finish. Finish shall be entirely mirror, satin or matte. Should the forceps be of mirror finish, the locks and ratchets shall be polished smooth, but need not possess a mirror finish. The matte finish shall exhibit a uniform nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure.

3.6.1 Passivation. The forceps shall be properly passivated.

3.7 Jaws and ratchet operation. Shall be as specified in ASTM F1026.

3.8 Identification markings. In addition to the requirements specified in ASTM F1026, forceps of foreign manufacture shall be marked "stainless."

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection shall be in accordance with GG-I-526.

4.1.2 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:21Z
Check the source to verify that this is the current version before use.

4.1.3 Certificates of quality. Certificates of quality supplied by the manufacturer of the corrosion-resistant steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificates shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number, the item identification, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicated in table I. Unit of product for examination purposes shall be one forceps.

TABLE I. Sampling Criteria for Examination

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in table II. The unit of product for test purpose shall be one forceps.

TABLE II. Sampling Criteria for Tests

Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Composition	3.2 and ASTM F899	GG-I-526	**	*
Hardness test	3.3 & ASTM F1026	4.4.1	S-2	1.5
Dimensions	3.4	---	S-2	2.5
Distortion test	ASTM F1026	---	S-2	2.5
Copper sulfate test	GG-I-526	GG-I-526	S-2	1.0
Boil test	GG-I-526	GG-I-526	S-2	1.0
Jaw and ratchet operation test	3.7	ASTM F1026	S-2	1.0

* Acceptance number shall be zero for all sample sizes.

** Sufficient quantity of each lot of metal used in the manufacture of the forceps shall be selected for chemical analysis.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:21Z
Check the source to verify that this is the currentversion before use.

4.3.1 Examination of forceps The forceps shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in tables I and III. Examination, however, shall not be restricted to defects listed in table III. Any forceps in the sample containing one or more defects shall be rejected and if the number of defective forceps in any sample exceeds the acceptance number for the sample, the lot represented by that sample shall be rejected.

TABLE III. Classification of defects

Categories	Defects
Critical	None defined
Major	
101	Forceps not of the specified design for appropriate type, stock number and item identification.
102	Both jaws not serrated throughout full length.
103	Jaw serrations not of uniform depth and not well defined.
104	Jaw serrations not chamfered through full length.
105	Ratchet teeth not well defined.
106	Forceps not free from cracks, pits, burrs, sharp edges scale, rough areas or corrosion.
107	Finish of forceps not uniform and not as specified.
108	Serrations of both haws do not completely mesh in fully closed position.
109	Box lock performance not as specified in 3.5.
110	Jaws possess sharp tips.
111	Ratchet teeth do not completely engage in all positions.
112	Pin for box lock not flush with surface and not free from crevices.
Minor	
201	Forceps not free from dirt, polishing material or foreign matter.
202	Identification markings not present, not correct, not complete, or not legible.

Source: https://assist.dla.mil -- Downloaded: 20261O-21T09:21Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Hardness test. Hardness shall be determined in accordance with ASTM Standards E 18 or E 92. Conversion of Vickers hardness values to Rockwell values shall be in accordance with ASTM Standard E 140.

4.5. Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Cleaning and drying. Forceps shall be cleaned and dried in accordance with MIL-P-116.

5.1.1.2 Preservative applications. Preservatives shall not be used.

5.1.1.3 Unit package. Each forceps shall be packaged in accordance with MIL-P-116, method 1C-1.

5.1.1.4 Intermediate package. Twelve unit packages containing forceps of one type only shall be packaged in a box conforming to PPP-B-566, PPP-B-676 or PPP-B-636, Type CF, class domestic. Closure shall be as specified in the applicable box specification.

5.1.2 Level C. The level C preservation for forceps shall conform to the MIL-STD-794 requirements for this level.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Six intermediate packages containing forceps of one type only shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type, PPP-B-621, class 2 or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:21Z
Check the source to verify that this is the current version before use.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Six intermediate packages containing forceps of one type only shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type, PPP-B-621, class 1 or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. The level C packing for forceps shall conform to the MIL-STD-794 requirements for this level.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in a exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 40 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in height, 48 inches in width, and 40 inches in length, and shall conform to Type IV or Type V.

5.2.5.1 Levels A and B. Forceps of one type only packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets conforming to MIL-STD-147.

5.2.5.2 Level C. Forceps of one type only packed as specified in 5.2.3 shall be unitized as specified in 5.2.5.

5.3 Marking.

5.3.1 Levels A, B and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:21Z
Check the source to verify that this is the current version before use.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.5.

6. NOTES

6.1 Intended use. These forceps of the Doyen design are to be used in grasping portions of the gastrointestinal tract during surgical procedures.

6.2 Ordering data.

6.2.1 Acquisition requirements. Acquisition documents should specify the following:

(a) Title, number, and date of specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(c) Applicable National Stock Number.

6.3 National Stock Number Coverage. This specification covers the following National Stock Numbers:

National Stock No.	Item Identification	Specification Designation Type
6515-00-335-1900	FORCEPS, INTESTINAL, CURVED, Doyen, 8-3/4 Inch	1
6515-00-335-3200	FORCEPS, INTESTINAL, STRAIGHT, Doyen, 9 Inch	2

6.4 Subject term (key word) listing:

Doyen
Forceps
Intestinal

6.5 This specification does not include all types, sizes, or grades as indicated by the title of all types, which are commercially available.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:21Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Project No. 6515-3507

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:21Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-LS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:21Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:21Z
Check the source to verify that this is the current version before use.