MIL-F-36825
Amendment-3
22 October 1974
SUPERSEDING
Amendment-2
2 May 1973

MILITARY SPECIFICATION

FORCEPS, ENDOTRACHEAL TUBE, MAGILL

This amendment forms a part of Military Specification MIL-F-36825, dated 7 April 1972, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 2:

3.4 Style, design, and dimensions. At end of paragraph delete period and add: "except delete distance from distal end to end of rings in end view of '4-3/4 to 7-1/4' and substitute '4-3/4 to 8-1/8', in front view delete width of rings dimension '2-7/8 ± 1/8' and substitute '2-3/4 to 3-5/64' and delete overall length '7-1/2 ± 1/2' and substitute '7 ± 1'; in top view delete angle offset of '18° to 25°' and substitute '18° to 40°'; and in Detail 'A' delete clearance '3/32' and substitute '0 to 7/64'."

5.1.3.2 Preservative. Delete "Not required" and substitute: "No preservative shall be applied."

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1879

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:24Z
Check the source to verify that this is the current version before use.