MILITARY SPECIFICATION

HOLDER, SUTURE NEEDLE, KALT, 5-3/8 INCH

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a suture needle holder of the Kalt design suitable for holding eye needles.

2. APPLICABLE DOCUMENTS

2.1 Specifications and standards. The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical, General Specifications.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Box, Wood, Wirebound.
PPP-B-601	Box, Wood, Cleated-Plywood.
PPP-B-621	Box, Wood, Nailed and Lock-corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Box, Set-up.

Military

MIL-P-116	Preservation, Methods of.
MIL-L-10547	Liners, Case, and Sheet Overwrap; Water-Vaporproof or Waterproof, Flexible.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards, required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 General Requirements. The needle holder shall be in accordance with the applicable requirements of GG-I-526 and the following option exercises under that specification:

3.2 Material. The needle holder shall be fabricated from Class 4, Type 410 or 420 material.

3.3 Hardness. After proper stress relieving, the needle holder shall have a Rockwell hardness of C40 to C45, with the exception of the screw or pin.

3.4 Style, design and dimensions. Shall be a suture needle holder of the Kalt design, with spring action and thumb rest. Shall have straight jaws, and one jaw shall be slotted and have straight serrations, the other jaw shall have cross serrations. Shall be suitable for securely holding eye needles. Shall be in accordance with Figure 1.

3.4.1 Locks. Shall be screw type lock.

3.5 Finish. Finish shall be mirror.

3.6 Identification marking. In addition to GG-I-526 requirements, needle holders of foreign manufacture shall also be marked "stainless".

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.




FIGURE 1

UNLESS OTHERWISE SPECIFIED, TOLERANCES SHALL BE AS FOLLOWS:

ALL DIMENSIONS 1/8 IN. AND UNDER	±1/64
ALL DIMENSIONS 1/8 IN. TO 1/2 IN.	±1/32
ALL DIMENSIONS OVER 1/2 IN.	±1/16
ALL ANGULAR DIMENSIONS	±2°
ALL DECIMAL DIMENSIONS	±.003

Source: https://assist.dla.mil -- Downloaded: 2026/10-21T09:20Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection shall be in accordance with GG-I-526.

4.1.2 Records. Records of examination and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.3 Certificates of quality. Certificates of quality supplied by the manufacturer of the corrosion resisting steel may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificates shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, the individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in Table I. The unit of product for examination purposes shall be one suture needle holder.

Table I. Sampling Criteria for Examination

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. Unit of product for test purposes shall be one suture needle holder.

Table II. Sampling Criteria for Tests

Component	Characteristic	Requirement	Test procedure	Insp. level	AQL
1-Holder	Composition analysis	3.2 and GG-I-526	GG-I-526	**	*
	Hardness	3.3 and GG-I-526	GG-I-526	S-2	*
	Dimensions	3.4	--------	S-2	4.0
	Boil test	GG-I-526	GG-I-526	S-2	*
	Copper sulfate	GG-I-526	GG-I-526	S-2	*

* Acceptance number shall be zero for all sample sizes.
** Four ounces of each lot of metal used in the construction of the suture needle holder shall be selected for chemical analysis.

4.3 Examination. The suture needle holder shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Table I and III. Inspection, however, shall not be restricted to defects listed in Table III. Any suture needle holder in the sample containing one or more defects shall be rejected, and if the number of defective items in any sample exceeds the acceptance number for test sample, the lot represented by the sample shall be rejected.

Table III. Classification of Defects

Categories	Defects
Critical	None Defined.
Major	
101	Style and design not as specified.
102	Holder not free from cracks, pits, dents, nicks, sharp edges and burrs.
103	Screw in lock not permanently set in a fixed position.
104	Serrations not as specified.
105	Ratchet teeth not uniform, or function poorly.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

Table III. Classification of Defects (cont'd.)

Categories	Defects
Minor	
201	Finish not mirror, and not uniform.
202	Identification marking not present, correct, complete or legible.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. One holder as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	6 units	144 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Not required.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive material.

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method IC 1.

5.1.3.5 Intermediate package. Intermediate package shall be a box of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except commercial colors will be acceptable, or PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent accidental opening under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3c.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4 of PPP-B-636.

5.2.1.4 Strapping Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:20Z
Check the source to verify that this is the currentversion before use.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The suture needle holder is to be used for holding eye needles during surgical procedures.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number, title and date of this specification.
(b) Applicable Federal Stock Number (FSN).
(c) Selection of applicable level of packaging and packing required (see 5.1 and 5.2).

6.3 Federal Stock Number (FSN) coverage. This specification covers the following item listed in the Federal Supply Catalog:

Federal Stock Number	Item Identification
6515-341-7800	HOLDER, SUTURE NEEDLE, Kalt, 5-3/8 inch

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

6.4 Supersession data. This specification includes the requirements of Purchase Description No. 11, dated 5 January 1966 for FSN 6515-341-7800 HOLDER, SUTURE NEEDLE, Kalt, 5-3/8 inch.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1606

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION

MIL-M-36702

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:20Z
Check the source to verify that this is the current version before use.

piece and what joint

FOLD

POSTAGE AND FEES PAID



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

Commander
Defense Personnel Support Center
Attn: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pa. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.