MIL-H-36794
3 September 1971

MILITARY SPECIFICATION

HANDPIECE, DENTAL, LABORATORY, STRAIGHT, HIGH SPEED, BELT DRIVEN

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a belt driven, round nose, straight handpiece for high speed dental laboratory use.

2. APPLICABLE DOCUMENTS

2.1 Specifications, standards, and drawing. The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

MMM-A-250	Adhesive, Water-resistant, (For Sealing Fiberboard Boxes).
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-sensitive Adhesive Paper, Water-Resistant (For Carton Sealing).

Military

MIL-P-116	Methods of Preservation.
MIL-M-7298	Manual, Technical, Commercial Equipment.

STANDARDS

Federal

FED-STD-151	Metal, Test Methods.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

24279 Handpiece, Dental Laboratory, Straight, High Speed, Belt Driven.

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM E 18 - Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials, Tests for.

ASTM E 92 - Vickers Hardness of Metallic Materials, Tests for.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, Pa. 19103).

3. REQUIREMENTS

3.1 Drawings. The drawings forming a part of this specification are specification control drawings. The supplier is responsible for preparing his own shop drawings. Where tolerances prescribed could cumulatively result in incorrect fits, the supplier shall provide tolerances within those prescribed on the drawings to insure correct fit, assembly, and operation of the item. No deviation from the prescribed dimensions on tolerances is permissible without prior approval from the contracting officer.

3.2 Materials. Materials shall be new and shall conform to the requirements specified herein. Materials not specifically designated in Table I shall be of high commercial quality throughout and shall be suitable for the intended purpose.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

Table I Materials, Platings, And Hardnesses.

Part Number	Part	Material	Plating or Coating	No. Rqd.	Rockwell * Hardness
1	Pin	AISI 420	------	1	B190-235
2	Nose Cone	Brass	Chromium	1	----
3	Chuck Holder	AISI 430F	------	1	B190-235
4	Chuck	AISI 420	------	1	C50-52
5	Back Cone	AISI 430F	------	1	B190-235
6	Ball Bearing	Steel (Corrosion Resisting)	------	1	----
7	Handpiece Shaft	AISI 430F	------	1	B190-235
8	Outer Sheath	Brass	Chromium	1	----
9	Washer	AISI 420	------	1	C51-54
10	Cap Nut	AISI 430F	------	1	----
11	Connecting Sheath	AISI 430F	------	1	----
12	Bushing	Brass	Nickel	1	----
13	Ball Bearing	Steel (Corrosion Resisting)	------	1	----
14	Wrist Joint Frame				
	A. Wrist Piece	Brass	Nickel-Chromium	1	----
	B. Shaft	AISI 430F	------	1	----
15	Pulley	Aluminum	Anodized	1	----
16	Nut	Aluminum	Anodized	1	----
17	Drive Rod	AISI 430F	------	1	B190-235
18	Washer	Spring Steel	Black Oxide	1	----
19	Washer	AISI 420	------	1	----
20	Bow	Steel	Nickel-Chromium	1	----
21	Washer	Spring Steel	Black Oxide	2	----
22	Screw	Steel	Nickel-Chromium	2	----
23	Washer	Steel	Black Oxide	2	----
24	Pulley Sheath	Brass	Nickel-Chromium	2	----
25	Ball Bearing Pulley	Steel	Nickel-Chromium (See 3.12)	2	B150-190
26	Washer	Brass	Nickel-Chromium	1	----
27	Pin	Steel	Black Oxide	2	----
28	Joint Plate	Brass	Nickel-Chromium	1	----
29	Shaft	Steel	Nickel-Chromium	1	----
30	Shield	Brass	Nickel-Chromium	1	----
31	Screw	Steel	Black Oxide	1	----
32	Connecting Arm	Steel	Nickel-Chromium	1	----
33	Screw	Steel	Black Oxide	1	----
34	Spring	Spring Steel	Black Oxide	1	C50-54
35	Pin	Steel	Black Oxide	1	----

* Hardness values designated for plated items are values before plating.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

3.3 Design and construction. The handpiece shall be a heavy duty, ball bearing, laboratory, straight handpiece equipped with a standard wrist joint having a swivel bow. The wrist joint shall be capable of being attached to the conventional type belt arm of a dental engine (See 6.3). The handpiece shall be suitable for use in grinding, contouring and finishing of dental porcelains and metals. The handpiece shall be a round nose, durable, precision-made instrument suitable for operation at handpiece speeds up to 40,000 revolutions per minute (r.p.m.). The design, construction and dimensions of the handpiece including wrist joint shall be in conformance with DPSC Drawing 24279 and shall essentially consist of the parts listed in Table I. The quantity of the components shall be as specified in Table I for the part number and name indicated. All handpiece parts having the same manufacturer's number, shall be interchangeable or replaceable as individual components or assemblies, as applicable.

3.3.1 Nose cone and sheath. The handpiece shall have a wide type, outer sheath with a thumb rest on the nose cone. The nose cone shall screw into the sheath, forming an integral combination. The sheath shall have interrupted knurling at the distal (nose) portion to provide gripping action.

3.3.2 Chuck. The handpiece chuck shall have three jaws and shall accept and hold standard straight handpiece burs and mandrel shanks of 0.0919 to 0.0929 inch diameter. The chuck shall hold the mandrel shank securely when the handpiece is tightened. Upon releasing chuck tension, the shank shall be free in the chuck.

3.3.3 Ball bearings. The handpiece shall have a total of four ball bearings; one each in the front and rear of the handpiece, and one each in the wrist joint pulleys. The ball bearings in the handpiece shall be wide type, double sealed, permanently lubricated ball bearings. The pulley ball bearings shall be single sealed, permanently lubricated ball bearings located in the wrist joint pulleys. All bearings shall provide smooth operation of the handpiece and wrist joint.

3.3.4 Wrist joint. The wrist joint shall have a bow to minimize interference, regardless of handpiece position. The wrist joint shall have two ball bearing "V" type pulleys. The wrist joint, including the belt arm attachment portion, shall conform to DPSC Drawing 24279 for design, construction and dimensions. The wrist joint shall be attached to the handpiece at one end and the opposing end shall be designed for attachment to a standard, dental engine, lower belt arm (See 6.3). The periphery of the pulleys shall not extend beyond the hinges of the joint.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

3.4 Accessories. One of each of the following accessories shall be included with each handpiece: wrench for removal of spindle tip (Cadmium plated steel), chuck cleaning brush, handpiece oil (1-1/4 ounce plastic squeeze bottle).

3.5 Hardness. Hardness of components shall be in accordance with Table I. Wherever a value is not specified, the hardness shall be suitable for the material used and purpose intended.

3.6 Plating. Components shall be plated as specified in Table I for the applicable components. Plating shall be uniform in thickness and appearance. Plating shall be free from blisters, visible porosity, or other defects.

3.7 End and Side Play. The end play of the spindle in relation to the nosepiece shall not exceed 0.001 inch when tested as specified in 4.4.3. Side play of the spindle at the nosepiece end shall not exceed 0.001 inch total indicator reading measured 1/2 inch from the nosepiece tip when tested as specified in 4.4.3.

3.8 Runout. Runout, measured 1/2 inch from the nosepiece tip shall not be greater than 0.001 inch total indicator reading when tested as specified in 4.4.4.

3.9 Operation Test. The handpiece shall run smooth and be free of excessive vibration when tested as specified in 4.4.5.

3.10 Temperature Rise. Handpiece temperature rise shall not exceed 42 degrees Fahrenheit (F^{o}) when tested in accordance with 4.4.6.

3.11 Endurance Test. The handpiece shall withstand one hour of operation without exhibiting excessive vibration, excessive wear of any surfaces, or excessive generation of heat, when tested as specified in 4.4.7.

3.12 Finish. Exposed surfaces shall have satin finish except for pulleys, pulley shield and knurled portions. The wrist joint pulleys and pulley shield shall have a bright chrome finish. The wrist joint pulleys shall have a black enamel finish on the outside faces of the center portion.

3.13 Identification Marking. Each handpiece shall be permanently and legibly marked with the name or registered trademark of the manufacturer. Handpieces of foreign (other than U.S.A.) manufacture shall be marked with the name of the country of origin.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

3.14 Service Data. The contractor shall furnish with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance, and lists of component parts. The manual shall comply with the requirements of MIL-M-7298. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer, via the cognizant Government inspector, two preliminary copies of his proposed technical manual for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of contract and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the Contracting Officer. Waiver of submission of 15 additional copies shall be at the option of the Government.

3.15 Workmanship. Workmanship shall be first class throughout. Handpieces shall be free from defects which detract from their appearance or which may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials used in the handpieces, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered are classified as specified in Tables II and IV (see 4.2.1 and 4.3) and Table III (see 4.2.2).

4.2.1 Sampling for examination. Sampling for examination for the handpiece shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL's) shall be as indicated in Table II. The unit of product for examination purposes shall be one handpiece.

Table II. Inspection and Acceptable Quality Levels

Type defect	Inspection level	AQL (percent defective)
Major A	II	1.0
Major B	S-1	2.5
Minor A	II	2.5
Minor B	S-1	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and Table III for the handpiece. The acceptance number shall be zero for all sample sizes. The unit of product for test purposes shall be one handpiece. Where feasible, the same sample unit may be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

Table III. Sampling criteria for Testing

Component	Characteristic	Requirement	Test procedure	Inspection level
All parts in table I	Chemical Analysis	3.2	4.4.1	*
Parts indicated in table I	Hardness	3.2	4.4.2	S-1
Assembled handpiece	End play and side play	3.7	4.4.3	S-1
Assembled handpiece	Runout	3.8	4.4.4	S-1
Assembled handpiece	Operation	3.9	4.4.5	100%
Assembled handpiece	Temperature rise	3.10	4.4.6	S-1
Assembled handpiece	Endurance	3.11	4.4.7	S-1

* A sufficient amount from each lot of raw materials used in the construction of the item shall be selected for chemical analysis.

4.3 Examination. The handpiece shall be examined to determine compliance with requirements contained in this specification. Examination shall be conducted in accordance with the following classification of defects. Examination shall include, but shall not be limited to, the defects listed in Table IV.

Table IV. Classification of Defects

Categories	Defects
Critical	None defined
Major A	
101	Belt arm attachment portion of wrist joint will not accept and lock-in to a standard dental engine lower belt arm.
102	Plated surfaces show signs of blisters, peeling and flaking.
103	Handpiece does not readily accept and securely hold standard shank mandrels.
104	Finish not as specified.
105	Pits, nicks and discoloration present.
106	Handpiece and pulleys bind or do not turn smoothly when turned manually by hand.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

Major B	
151	Dimensions affecting fit of belt arm attachment portion of wrist joint to standard dental engine lower belt arm not as specified.
152	Dimensions affecting fit of bur shank to bur socket not as specified.
Minor A	
201	Service data missing.
202	Identification marking missing.
203	Identification marking not permanent and legible.
204	Components missing.
Minor B	
251	Dimensions (other than 151 and 152) not as specified.

4.4 Tests.

4.4.1 Chemical analysis. Metal analysis shall be conducted in accordance with method 111 or 112 of FED-STD-151.

4.4.2 Hardness test. Hardness shall be determined in accordance with ASTM Designation E 18 (Rockwell) or E 92 (Vickers or Diamond Pyramid).

4.4.3 End play and side play. The test shall be conducted with a true running test shank in the handpiece. A dial indicator sensitive to 0.0002 inch equipped with a knife edge foot or other suitable equipment of equal accuracy, shall be used. The indicator shall be positioned such that end play (movement of the test shank parallel to the axis of the spindle) can be measured. The test shank shall be moved toward and away from the nosepiece (in line with axis of the spindle) and the end play measured. The difference between the maximum and minimum readings shall be reported as end play. The indicator shall then be positioned such that side play (movement of the test shank perpendicular to the axis of the spindle) can be measured (1/2 inch from the nosepiece tip). The test shank shall be moved side to side (in a plane perpendicular to the spindle axis) and the side play measured. The difference between the maximum and minimum readings shall be reported as side play.

4.4.4 Runout. The test shall be conducted with a true running test shank in the handpiece. A dial indicator sensitive to 0.0002 inch equipped with a knife edge foot or other suitable equipment of equal accuracy shall be employed. Runout shall be measured at 1/2 inch from the nosepiece. The test shank shall be rotated slowly by the handpiece through 360 degrees and maximum and minimum readings recorded. The difference between the maximum and minimum readings for one full rotation of the shank shall be reported as the runout.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

4.4.5 Operation test. The assembled and lubricated handpiece shall be tested by running the handpiece under no load for 3 minutes at 39,500 to 40,000 r.p.m. The handpiece shall run smooth and shall be free of excessive vibration when hand held.

4.4.6 Temperature rise. A properly calibrated copper-constantan type thermocouple shall be attached to the external portion of the handpiece at both bearing areas. Readings shall be taken which shall represent the starting or ambient temperatures. The handpiece, properly assembled and lubricated, shall be operated under no load for 5 minutes at 39,500 to 40,000 r.p.m. The temperatures being monitored by the thermocouples shall not rise higher than 42°F. above the starting temperatures.

4.4.7 Endurance test. The lubricated handpiece shall be tested by cutting glass with a No. 2 Tungsten Carbide Bur in the handpiece with the handpiece operating at a speed from 34,000 r.p.m. ± 5 percent by means of a belt engine. The handpiece shall be held stationary and a movable carriage shall move the glass block into contact with the side of the bur and shall maintain a one pound ± 1/2 ounce load against the side of the bur (perpendicular to the glass surface being cut). The test shall be a one hour test consisting of fifteen 4-minute cycles of 3 minutes engaged (cutting glass) and 1 minute disengaged (not cutting glass). The glass shall be plate glass having a minimum thickness of 1/8 inch and having a Knoop hardness of 450 to 510 with a one kilogram load. At the end of the one hour test, the handpiece shall be retested for operation (smoothness and excessive vibration) and heat rise as specified in 4.4.5 and 4.4.6. The handpiece shall then be disassembled and examined under a 15 power (15X) magnification for excessive wear of components.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One handpiece, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table V shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table V. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	6 units	24 units

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Not required.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive materials.

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method IA-15.

5.1.3.5 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flaps with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure of wood boxes shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II of PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Type required.
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 Federal Stock Number (FSN) coverage. This specification covers the following item appearing in the Federal Supply Catalog: 6520-484-9495 - HANDPIECE, DENTAL Straight, Round Nose, High Speed: 40,000 rpm speed range.

6.3 The handpiece covered by this specification is for use with the following item appearing in the Federal Supply Catalog: 6520-484-9494 - DENTAL ENGINE, Bench Type, High Speed, 110 volt, 60 cycle, AC.

6.4 Limitations. This specification does not cover all types, classes, grades, styles, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover those items which are normally procured to meet Military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity
Defense Supply Agency - DM
Project No. 6520-1298

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the [illegible] and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to [illegible] comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion [illegible] contractual requirements.

[illegible]:

ORGANIZATION

CITY AND STATE	CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*	DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026l0-21T09:20Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:20Z
Check the source to verify that this is the current version before use.