MIL-O-36811A
21 September 1978
SUPERSEDING
MIL-O-36811
26 April 1972

MILITARY SPECIFICATION

OTOSCOPE AND OPHTHALMOSCOPE SET, BATTERY TYPE, POCKET SIZE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a miniature otoscope with fiber optics illumination, and a miniature ophthalmoscope, both contained in a single pocket size carrying case.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bid or request for proposal form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-585	- Boxes, Wood, Wirebound.
PPP-B-601	- Boxes, Wood, Cleated-Plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-B-676	- Boxes, Setup.

MILITARY

MIL-B-18	- Batteries, Dry.
MIL-P-116	- Preservation, Methods of.
MIL-L-10547	- Liners, Case and Sheet Overwrap, Water-Vaporproof or Waterproof, Flexible.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:22Z
Check the source to verify that this is the current version before use.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

(Copies of specifications, standards, drawings and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. All glass parts shall be fabricated from high quality optical glass. All other materials shall be as specified herein and of high quality throughout, suitable in all respects for the purpose intended.

3.2 Design and construction. The set shall be a pocket size ophthalmoscope, and fiber optics illuminated otoscope set consisting of the following components:

(a) Fiber optics otoscope head with lamp.
(b) Battery handle for otoscope head.
(c) Ophthalmoscope head with lamp.
(d) Battery handle for ophthalmoscope head.
(e) Spare lamp for otoscope.
(f) Spare lamp for ophthalmoscope.
(g) Four reusable specula; 2, 3, 4 and 5 millimeters (mm).
(h) Ten adult size disposable specula.
(i) Cushioned and compartmented carrying case.

3.3 Components.

3.3.1 Otoscope.

3.3.1.1 Otoscope head. The otoscope head shall be of the fiber optics type. The otoscope head shall be fabricated from nylon or other suitable high-impact plastic. The head shall have a permanently attached metal speculum slotted to accept and hold the specula. The glass fiber bundle shall be encased between the funnel-shaped speculum holder and funnel-shaped inner part. The speculum holder shall be fabricated of metal, its surface shall be highly polished and chromium plated. The interior funnel shall be fabricated of suitable black plastic with matte or dull surface on the interior, to prevent undesired light reflections. The distal end of the body (metal speculum holder) shall extend beyond the distal end of the fiber optics light transmitting bundle for a distance

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:22Z
Check the source to verify that this is the current version before use.

of not less than 2.5 mm. The source of illumination shall be a lamp as specified in 3.3.1.2. The axial illumination, provided by the otoscope head with an annular fiber optics light transmitting bundle, shall be not less than 300 lumens per square foot at the distal end of the longest speculum at a lamp voltage of not more than 2.5 volts when tested as specified in 4.4.1. The head shall have a threaded male connector which screws into the battery holder and handle. The head shall be equipped with a magnifier that shall be capable of being moved out of the line of vision to permit oblique observation. The magnifier shall have a focal length of not more than 85 mm for obtaining sufficient magnification. The inner diameter of the viewing aperture shall not be less than 4 mm.

3.3.1.2 Lamp. The otoscope lamp shall be rated at a maximum of 280 milliamperes (ma) at 2.2 volts. The lamp shall be a plug-in type with center contact. The base of the lamp shall contain a spring element to exert a pressure on the wall of the lamp socket to prevent the lamp from falling from the socket. The lamp shall meet the illumination requirements specified in 3.3.1.1. A lamp shall be furnished in position in the otoscope head. The lamp shall have a minimum life of 12 hours when tested as specified in 4.4.2.

3.3.1.3 Specula.

3.3.1.3.1 Reusable specula. The reusable aural specula shall be supplied in four sizes; 2, 3, 4 and 5 mm inside diameter (ID) at the distal end. These specula shall be fabricated of a suitable plastic material, black in color, and nonreflecting to prevent interfering glare to the operator. The specula shall be designed to permit ready attachment to, or removal from, the otoscope head. The specula shall not be distorted or the finish marred when subjected to the boil test specified in 4.4.3. The specula shall not chip, fracture or become distorted when subjected to the drop test specified in 4.4.4. There shall be no discoloration or dulling of the surface of the specula or other defects when subjected to the alcohol immersion test specified in 4.4.5.

3.3.1.3.2 Disposable specula. Each set shall be equipped with ten (10) disposable specula of adult size. The specula shall be fabricated from a suitable plastic. The specula shall be contained in the case in a transparent plastic container.

3.3.2 Ophthalmoscope.

3.3.2.1 Ophthalmoscope head. The ophthalmoscope shall be of the mirror type with single rekoss disk (lens disc). The head or housing shall be fabricated from a suitable plastic material. The exposed surfaces shall be black. The viewing lens shall be enclosed on the viewer's side with an optically polished face plate. The light emitting aperture shall be capable of being closed with a sliding cover. The ophthalmoscope head shall be designed for ready attachment to, or removal from, the battery handle specified in 3.3.3. The connector for attaching the ophthalmoscope head to the battery handle shall be fabricated of a suitable metal. The optical reflecting system, the device for selecting the filter and apertures, and the rekoss disc shall be incorporated into the head.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:22Z
Check the source to verify that this is the current version before use.

3.3.2.2 Optical reflecting system. The optical reflecting system shall consist of a lamp, a condenser lens, a device for selecting the aperture, an illumination objective for projecting the apertures at a suitable distance from the ophthalmoscope, and a mirror for deflecting the beam of light to the lens system. The aperture shall include a full circle, a half circle, a vertical slit, and a green filter. The device for selecting the apertures and filters shall be designed to permit convenient operation without interrupting the examination. The lamp shall be as specified in 3.3.2.4. The lamp shall be located in the connector of the ophthalmoscope head. Indexing of the lamp receptacle shall be provided so that the lamp filament is always properly centered and focused in relation to the optical system. No adjustment shall be required when replacing a lamp. No special tool shall be required to replace a lamp. A spring element shall be provided in the lamp housing to provide a frictional force on the lamp to prevent the lamp from sliding from its receptacle. The illumination provided by the optical system, at a point 7-3/4 inches from the exit part of the ophthalmoscope, shall be not less than 3.5 lumens per square foot when the lamp is operated at 2.5 volts and tested as specified in 4.4.1. The illuminated spot produced by the ophthalmoscope head shall be uniform.

3.3.2.3 Rekoss disc. The rekoss disc shall consist of 16 lenses, 0 to +20 diopters and 0 to -20 diopters, with an adequate working assortment of powers. Numerical markings shall be provided in the rekoss disc to identify the lens selected. These numerals shall be clearly visible in a magnified and illuminated window in the ophthalmoscope head. Magnification and illumination shall be sufficient to enable the operator to read the markings in a darkened room. Plus diopters shall be marked in black, minus diopters shall be marked in red. The lenses shall be fabricated of suitable polished, optical glass. The lenses shall be securely mounted to the rekoss disc. A click stop shall be provided for holding each lens in position.

3.3.2.4 Lamp, ophthalmic. The lamp shall be the plug in type with a center contact. The base of the lamp shall be equipped with an indexing device for positioning the lamp in relation to the optical system as specified in 3.3.2.2. The lamp shall have a close coiled filament. The axis of the filament shall be straight. The lamp shall draw a maximum current of 280 milliamperes (ma) when operated at 2.2 volts The lamp shall have a minimum life of 12 hours when tested as specified in 4.4.2.

3.3.3 Battery handles. The battery handles shall be cylindrical in shape and of a size to hold two type AA batteries per MIL-B-18. The battery handle shall be fabricated from nylon or other suitable high impact plastic. The handle shall have a maximum length of 12 centimeters (cm), excluding the heads, and shall be 20 ± 2 mm outside diameter (OD) at the top and 18 ± 2 mm OD at the bottom. The upper end of the handle shall have a threaded female connector for easy attachment of the otoscope and ophthalmoscope heads. The other end of the battery handles shall be provided with a threaded cap. The cap shall be provided with a contact spring designed to hold the batteries securely in position while making electrical contact with the cap. Each battery handle shall be equipped with a metal pocket clip suitable for securely holding the handles in the pocket with the otoscope and ophthalmoscope attached. The pocket clip shall include, as an integral part, a sliding switch, which switches off the handle when clipped into a pocket. The battery handles shall be supplied without batteries.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:22Z
Check the source to verify that this is the current version before use.

3.3.4 Carrying case. The carrying case shall be fabricated of a high impact material. There shall be a form fitting insert compartment to accommodate all components of the set. The case shall be provided with a continuous hinge with suitable springs which shall automatically close the cover under spring pressure. The top and bottom sections of the case shall have metal trimmed edges. The overall dimensions of the carrying case shall be 6-5/8 ± 1/4 inches long by 3-1/4 ± 1/4 inches wide by 1-1/8 ± 1/4 inches thick.

3.4 Identification marking. Each instrument, the ophthalmoscope head, the otoscope head, the battery handles and the carrying case, shall be permanently and legibly marked with the name or registered trademark of the manufacturer. The markings of the carrying case shall also include the manufacturer's model number for the set, if such number is used. The markings shall be of an appropriate size for easy reading and shall be either black or gold in color. As an alternate the markings on the carrying case shall be accomplished by use of an adhesive backed aluminum plate. As an alternate, the markings on the case may include the National Stock Number (NSN) for the item and the letters "US".

3.5 Instructions. Each set shall be furnished with two copies of clear instructions for assembly and use of the instrument. The instructions shall contain maintenance information and a list of component parts including the manufacturer's part or catalog number.

3.6 Workmanship. Workmanship shall be first class throughout. The otoscope and ophthalmoscope set shall be free from defects which detract from its appearance or impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:22Z
Check the source to verify that this is the current version before use.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one otoscope and ophthalmoscope set.

TABLE I. Sampling for examination.

Type defect	Inspection level	AQL (percent defective)
Major A	S-2	1.0
Major B	II	1.0
Minor	II	2.5
Dimensional	S-1	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II. The unit of product for test purposes shall be one otoscope and ophthalmoscope set.

TABLE II. Sampling for tests.

Component	Characteristics	Requirement	Test procedure	Inspection level	AQL (percent defective)
Otoscope head	axial illumination	3.3.1.1	4.4.1	S-1	1.5
Otoscope lamp	life test	3.3.1.2	4.4.2	S-1	1.5
Specula	boil test	3.3.1.3.1	4.4.3	S-1	2.5
	drop test	3.3.1.3.1	4.4.4	S-1	2.5
	alcohol immersion	3.3.1.3.1	4.4.5	S-1	2.5
Ophthalmoscope head	axial illumination	3.3.2.2	4.4.1	S-1	1.5
Ophthalmoscope lamp	life test	3.3.2.4	4.4.2	S-1	1.5
Lamp					
Otoscope	current draw	3.3.1.2	4.4.6	S-1	1.5
Ophthalmoscope	current draw	3.3.2.4	4.4.6	S-1	1.5

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:22Z
Check the source to verify that this is the current version before use.

4.3 Examination.

4.3.1 Set. The components of the set shall be examined to determine compliance with all the requirements specified herein. Examination shall be conducted in accordance with tables I and III. Examination however shall not be restricted to defects listed in table III.

TABLE III. Classification of defects.

Categories	Defects
	Otoscope head
Major A:	
101	Clear, undistorted view and uniform light intensity not obtained without adjustment of lamp.
102	Magnifier distorted, scratched, chipped, not free of bubbles or striae, or otherwise defective so that image is distorted.
Major B:	
151	Unit inoperative.
152	Does not provide uniform light concentrated and focused as specified, resulting in uneven illumination.
153	Illuminated spot not uniform.
154	Speculum holder does not hold specula securely.
155	Interior of specula not black.
156	Distal ends of specula not rounded, not free of sharpness.
157	One, or more, specula missing.
158	Not easily and smoothly attachable to, or detachable from, battery handle.
159	Not of fiber optics type.
160	Exposed parts not smooth and free of sharpness, burrs, or flash.
161	Finish is porous, peeling, or incomplete.
162	Part or parts missing.
Minor:	
201	Exposed surfaces contain scratches, chips, or discoloration.
202	Identification markings missing or not legible.
203	Service data missing or not complete.
204	Replacement lamps missing or inoperative.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:22Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects. (Cont'd)

Categories	Defects
	Ophthalmoscope head
Major A:	
101	Glass not suitable optical glass.
102	Not detachable, mirror type, with single rokoss disc.
103	Exposed plastic surfaces not finished in black as specified (except rekoss disc and sliding dust cover).
104	Not designed for ready attachment to, or removal from battery handle.
105	Reflecting lens, aperture, and filter systems not situated or do not operate as specified.
106	Mirror and condensing lens not located and secured as specified.
107	Optimum illumination not obtained without manipulation of ophthalmoscope head.
108	Lens system does not consist of lenses and aperture as specified.
109	Not provided with illuminated dial-reading aperture with magnifier as specified.
110	Aperture and filter system not as specified.
111	Lamp housing not as specified.
112	Lamp not as specified.
113	Rekoss disc lenses distorted, scratched, chipped, not of high luster and polish, not free of bubbles or striae, or otherwise defective so that images are distorted.
Major B:	
151	Unit inoperative.
152	Condensing lens and mirror not situated, not adjusted as specified, resulting in shadows and back reflections.
153	Illuminated spot not uniform or free of distortions.
154	Identification markings for lenses (and zero aperture) not as specified, not legible, or missing.
155	Rekoss disc not operating properly (i.e., lenses not centered at stop position, dial not smooth in operation, binding, etc.)
156	Aperture and filter system selector not operating properly.
157	Not easily and smoothly attachable to, or detachable from, battery handle.
158	Exposed parts not smooth and free of sharpness, burrs, or flash.
159	Finish is porous, peeling, or incomplete.
160	Part or parts missing.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:22Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects. (Cont'd)

Categories	Defects
Minor	
201	Exposed surfaces contain scratches, chips, or discoloration.
202	Identification markings missing or not legible.
203	Service data missing or not complete.
204	Replacement lamps missing or inoperative.
	Battery handle
Major A:	
101	Not cylindrical and of a size to hold two type AA batteries as specified.
102	Not fabricated as specified.
103	Supplied with batteries.
Minor:	
201	Exposed surfaces contains scratches, chips or discoloration.
202	Identification marking missing or illegible.
	Carrying case
Major A:	
101	Not fabricated as specified.
102	Components do not fit properly in compartments.
Minor:	
201	Exposed surfaces contains scratches, chips or discoloration.
202	Identification marking missing or illegible.

4.3.2 Packaging inspection. The inspection of the packaging, packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Axial illumination. The axial illumination shall be determined with the use of the MacBeth Illuminometer, or an equivalent apparatus. A photoelectric photometer may be used providing that the MacBeth Illuminometer is used for a verification test or for calibration purposes. The otoscope head and the ophthalmoscope head shall meet the requirements of 3.3.1.1 and 3.3.2.4 respectively.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:22Z
Check the source to verify that this is the current version before use.

4.4.2 Lamp life. The lamp shall be supplied with 2.2 ± 0.125 volts direct current. The lamp shall be operated in cycles of 5 minutes on and 2 minutes off. The minimum on-time shall be 12 hours. The lamp shall meet the requirements of 3.3.1.1.

4.4.3 Boil test. The specula shall be boiled in distilled water for four successive 15 minute periods. After each 15 minute period, the specula shall be removed and allowed to air dry for 10 minutes. The specula shall meet the requirements of 3.3.1.3.1.

4.4.4 Drop test. The specula shall be dropped upon a cement floor from a height of 5 feet. The specula shall meet the requirements of 3.3.1.3.1.

4.4.5 Alcohol immersion test. The specula shall be completely immersed in 95 percent ethyl alcohol for 30 minutes. After removal, the specula shall be rinsed thoroughly in distilled water and dried. The specula shall meet the requirements of 3.3.1.3.1.

4.4.6 Current draw. The otoscope lamp and the ophthalmoscope lamp shall be tested for current draw.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit packaging. Each otoscope and ophthalmoscope set, packaged in the case, shall be further packaged into a box of appropriate size and design constructed in accordance with PPP-B-566 or PPP-B-676. Closure of box shall be in accordance with the appendix of the applicable box specification.

5.1.1.2 Intermediate package. Twelve (12) otoscope and ophthalmoscope sets, shall be placed into a box of appropriate size constructed in accordance with PPP-B-636, type CF, class domestic. Closure of box shall be in accordance with the appendix of PPP-B-636.

5.1.2 Level C. Otoscope and ophthalmoscope set shall be individually packaged in a manner that will afford adequate protection against corrosion, deterioration and physical damage during shipment from supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Four (4) intermediate packages shall be packed into an exterior container designed for a type 2 load constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather resistant, grade V3c. Closure and strapping shall be in accordance with the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:22Z
Check the source to verify that this is the currenversion before use.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof as specified in the appendix of PPP-B-636.

5.2.2 Level B. Four (4) intermediate packages shall be packed into an exterior container designed for a type 2 load constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1, or PPP-B-636, class domestic. Closure shall be in accordance with the appendix of the applicable box specification.

5.2.3 Level C. The packaged otoscope and ophthalmoscope sets shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carrier rules and regulations.

5.3 Marking. Each unit, intermediate package and exterior container shall be marked as specified in MIL-STD-129.

6. NOTES

6.1 Intended use. The otoscope and ophthalmoscope set is intended for use as a pocket sized diagnostic set for examination of the ears and the eyes.

6.2 Ordering data. Procurement documents should specify the following:

a) Title, number, and date of this specification.
b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
c) The National Stock Number (NSN).

6.3 This specification is intended to cover the following item listed in the Federal Supply Catalog:

National Stock Number	Item Identification
6515-00-165-6544	OTOSCOPE AND OPHTHALMOSCOPE SET, BATTERY TYPE, POCKET SIZE

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DoD - MB

Preparing activity:
DLA-DM

Project No. 6515-2124

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:22Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:22Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS. This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.

NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER *(Number)* AND TITLE

MIL-O-36811A OTOSCOPE AND OPHTHALMOSCOPE SET, BATTERY TYPE, POCKET SIZE

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1. ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*

TELEPHONE NO.

DATE

DD FORM 1 OCT 76 1426 EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:22Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

POSTAGE AND FEES PAID



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

DEFENSE PERSONNEL SUPPORT CENTER
DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:22Z
Check the source to verify that this is the current version before use.