MIL-P-36792
Amendment-2
21 August 1972
SUPERSEDING
Amendment-1
24 February 1972

MILITARY SPECIFICATION

PAD, ISOPROPYL ALCOHOL IMPREGNATED, STERILE

This Amendment forms a part of Military Specification MIL-P-36792, dated 31 August 1971, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 1:

2.1 Under SPECIFICATIONS, Military, add: "MIL-S-36586 STERILIZATION Test Strip Set, Bacterial Spore.

Page 2:

2.2 Other publications. Add the following:

"Code of Federal Regulations
Title 21, Part 147 to End
Revised as of January 1, 1971

For sale by Superintendent of Documents, U.S. Government Printing Office, Washington, D. C. 20402".

Page 3:

3.2.2 Cotton or rayon fabric. Delete last sentence in first and second paragraphs. After second paragraph, add: The cotton or rayon fabric (See 4.5) shall have a thickness range of 0.035 to 0.065 inches. The dry weight of single thick fabric shall be as follows: 1. For Single thickness pad - 120 to 230 gms. 2. For Double thickness pad - 84 to 86 gms."

3.3 Construction. Delete first sentence and substitute: "Material of the types described in 3.2 shall be cut into pads, either flat, single thickness of folded to double thickness and the finished pad shall have dimensions of 1-3/8 $\pm$ 3/8 wide by 2 $\pm$ 5/8 inches long."

Page 4:

3.4 and 3.5 Delete entirely and substitute:

3.4 Sterility. The isopropyl alcohol impregnated pads shall be sterile when tested with a biological indicator as specified in 4.4.5.

FSC 6510

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

3.4.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item identification.
b. Federal stock number.
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.).
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or a designated qualified alternate.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pa., 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

3.4.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and simultaneously sterilized in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.1. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

Renumber paragraph 3.6 to "3.5" and delete last sentence.

Renumber paragraph 3.7 to "3.6".

Page 7:

In Table II, under "Requirement Paragraph" delete "3.6" and substitute "3.5".

4.4.4 Alcohol content of the complete pad. Line 2, delete "pads" and substitute "envelope".

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:20Z
Check the source to verify that this is the current version before use.

4.4.5 Sterility test. The isopropyl alcohol impregnated pads shall be tested for sterility in accordance with the method specified in the USP for purified cotton, gauze, surgical dressings, and related material using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in paragraph 4.4.5.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for Type I Sterilization and 30 for Type II Sterilization including inoculated carrier and product test units. The inspection level shall be as indicated in paragraph 4.2.3.

4.4.5.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65°C
Dry Heat	B. stearothermophilus, or	55° - 65°C
	B. subtilis var.niger (or globigii)	30° - 35°C
Gas	B. subtilis var.niger (or globigii)	30° - 35°C
Radiation	B. pumilus	30° - 35°C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

4.4.5.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.5 Sterility test and 4.4.5.1 Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.4.5.1. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

4.4.5.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

4.4.5.4 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the U.S.P.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

Page 8:

4.4.6 Dermatology test. Dermatological reaction tests shall be performed by the methods described in the Code of Federal Regulations, Title 21 - Food and Drugs, Part 147 to End, Revised as of January 1, 1971, Paragraphs 191.10 and 191.11 page 248, with values from table for skin reactions of 1 or less, or that method as approved by the contracting officer.

Page 9:

5.2.1.1 Exterior container. End of paragraph add: "Special requirements of applicable Table shall apply for PPP-B-636 box."

Page 10:

5.2.2.1 Exterior container. End of paragraph, add: "Special requirements of applicable Table shall apply for PPP-B-636 box."

5.3.1 Immediate container. Line 4, delete the word "sterile".

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6510-0583

Review activities:

Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination, draft circulation should be based on information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.