MILITARY SPECIFICATION

PAD, ISOPROPYL ALCOHOL IMPREGNATED, STERILE

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a sterile pad impregnated with alcohol, suitable for use as a pre-injection skin preparatory.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Setup.

Military

MIL-L-10547	Liners, Case; and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

UNITED STATES PHARMACOPEIAL CONVENTION, INC.

Pharmacopeia of the United States

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pennsylvania 18042.)

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE

Food and Drug Administration

Federal Food, Drug, and Cosmetic Act and Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, D. C. 20204.)

AMERICAN PHARMACEUTICAL ASSOCIATION

National Formulary

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pa. 18042.)

3. REQUIREMENTS

3.1 Component parts. The pad shall be a nonwoven cotton or rayon fabric, impregnated with isopropyl alcohol.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

3.2 Materials.

3.2.1 Isopropyl alcohol. Isopropyl alcohol shall be Isopropyl Rubbing Alcohol, NF. A residual trace of sterilizing chemical, used in sterilizing the pads, shall be permitted as an additive. The alcohol shall meet all the standards of Isopropyl Rubbing Alcohol, NF.

3.2.2 Cotton or rayon fabric. The nonwoven cotton felt used in the fabrication of the pads shall be made from fully bleached purified, new or new reworked cotton such as comber or card strips; shall be of fine filaments and shall be substantially free of oil spots, streaks of dirt, all other foreign matter, and lumps. In fabrication the cotton fibers of the cotton pad material shall be physically locked and entangled to form a curled and kinked configuration. Material (see 4.5) shall have a minimum dry weight of 120 grams per square yard and a minimum thickness of 0.035 inch.

The rayon pads shall be made from fully bleached viscose rayon. Fibers shall be substantially free of oil spots, streaks, dirt, foreign matter, and lumps. The rayon pad material shall be made by the process of needle-looming (see 4.5). The material shall have a minimum dry weight of 140 grams per square yard.

When tested as specified in 4.3, the nonwoven material shall have the following characteristics:

3.2.2.1 Absorbency. Not more than 30 seconds shall be required for complete submersion when tested as specified in 4.4.1.

3.2.2.2 Water extract (residue weight prior to ignition). Not more than 0.75 percent of the weight of the cotton fabric, nor more than 1.0 percent of the rayon fabric shall remain when tested as specified in 4.4.2.

3.2.2.3 Fatty matter. Not more than 0.70 percent of weight of cotton or rayon fabric shall remain when tested as specified in 4.4.2.

3.2.2.4 Residue on ignition. Not more than 0.5 percent of the weight of the cotton fabric, nor more than 2 percent of the rayon fabric shall remain when tested as specified in 4.4.2.

3.2.2.5 Uncombined dyes. The percolate may show a yellowish color as a result of the test, but neither a blue nor a green tint shall be apparent when tested as specified in 4.4.2.

3.2.2.6 Acid or alkali. The water extract shall not show a pink color when tested as specified in 4.4.2.

3.2.2.7 Dextrin or starch. Test results shall not show a red, violet, or blue color when tested as specified in 4.4.2.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:20Z
Check the source to verify that this is the current version before use.

3.3 Construction. Material of the types described in 3.2 shall be cut into pads measuring 1-1/16 to 1-3/4 inch in width, and 1-7/16 to 2-1/16 inches in length. The pads shall be impregnated with 70% isopropyl alcohol solution in accordance with 3.2.1. The pads shall then be sealed in a foil plastic laminate. The final product shall be a sterile, nonwoven cotton or rayon pad, containing not less than 0.87 gm. of 70 percent isopropyl alcohol, and sealed in a rectangular foil plastic laminate.

3.4 Sterility. Finished pads shall be sterile, when tested as specified in 4.4.5.

3.5 Certificate of sterility. A certificate of sterility, stating that samples from each lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

(a) Item identification.
(b) Federal stock number.
(c) Contract or purchase order number.
(d) Sterilization lot or control number.
(e) Date of sterilization.
(f) Date sterility test was started.
(g) Specific U.S.P. test method used and results.
(h) Statement that samples comply with U.S.P. sterility test.
(i) Signature of head of laboratory or a designated alternate qualified in the testing performed.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to Directorate of Medical Materiel, ATTN: ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, and a copy forwarded with each shipment to consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government quality assurance representative for submission to the procuring activity. Complete records, which shall include the gauge pressure and temperature readings, shall be maintained by the contractor for a period not less than two years from the date of delivery of contract material to the Government and shall be made available upon request. The Government reserves the right to test such product material for sterility.

3.6 Certificate of Dermatological reactions. The contractor shall submit to the procuring activity, a copy of the original certification, that the lots or batches offered to the Government were composed of the same formula and manufactured identically as the production lot or batch originally tested for dermatological reactions to the impregnated fabric. When tested as specified in 4.4.6 the results of tests shall show a sensitizaiton rate of one (see 4.4.6).

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:20Z
Check the source to verify that this is the current version before use.

3.7 Workmanship. The impregnated pad shall be free from defects which may affect its appearance or serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Manufacturer's protocol. A manufacturer's protocol that lists the end item testing characteristics with specification requirements and which state that test results comply with said requirements, is acceptable providing the actual test results are made readily available for review to Government representatives upon request.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables I and II.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated below. Unit of product for examination purposes shall be one pad.

AQL - Major Defects - 1.0
AQL - Minor Defects - 2.5
Sampling plan - Level I

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:20Z
Check the source to verify that this is the currentversion before use.

4.2.2 Examination. The pads shall be examined for defects including, but not limited to, those listed in table I.

4.2.3 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, Level S-1, except that the acceptance number shall be zero for all sample sizes. Unit of product for test purposes shall be 1 pad. Minimum sample for variables shall be three.

4.2.4 Classification of defects. The unit shall be examined or tested in accordance with the following classification of defects:

Table I. Categories and defects

Categories	Defects
Major A	
101	Finished pad not free of ravelings.
102	Individual envelope not completely sealed.
103	Pad not porous.
104	Finished pad not free of dirt or foreign matter.
105	Envelope (laminate) not easily opened.
106	Envelope damaged.
107	Envelope leaking.
Major B	
151	Labeling not complete or correct.
152	Pad dimensions not within tolerances.
Minor	
201	Labeling not legible.

* Examination shall not be limited to the defects stated.

4.2.5 Sample for tests. Where possible, the same sample unit shall be used for the determination of two or more tests.

4.3 Testing of end item. The requirements and method of testing shall be as listed in table II. The sample level shall be S-1 with an acceptance number of zero. Minimum sample shall be three.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

Table II. Tests

Characteristic	Requirement Paragraph	Test Method
Isopropyl Alcohol	3.2.1	4.4.3
Cotton or Rayon Fabric	3.2.2	
Absorbency	3.2.2.1	4.4.1
Water Extract	3.2.2.2	4.4.2
Fatty Matter	3.2.2.3	4.4.2
Residue on Ignition	3.2.2.4	4.4.2
Uncombined Dyes	3.2.2.5	4.4.2
Acid or Alkali	3.2.2.6	4.4.2
Dextrin or starch	3.2.2.7	4.4.2
Alcohol Content of the complete pad	3.3	4.4.4
Sterility	3.4	4.4.5
Dermatology	3.6	4.4.6

4.4 Tests.

4.4.1 Absorbency. A sample pad before impregnation shall be completely submerged within the time specified in 3.2.2.1, when held horizontal to and almost in contact with the surface of distilled water at 25° Centigrade, and allowed to drop lightly upon the water.

4.4.2 Foreign matter. Cotton or rayon fabric shall be tested for compliance with foreign matter requirements as listed in 3.2.2.2 through 3.2.2.7 in accordance with test procedures for absorbent gauze as outlined in the current edition of the Pharmacopeia of the United States.

4.4.3 Isopropyl alcohol. Isopropyl alcohol shall be determined in accordance with The National Formulary for Isopropyl Rubbing Alcohol, without additives.

4.4.4 Alcohol content of the complete pad. Samples of the complete pads shall be weighed. The pads shall then be opened, and the alcohol saturated pads exposed. Both the pad and the envelope shall be conditioned for at least 4 hours in a standard atmosphere of 65 ± 2 percent relative humidity at 70 ± 2° F. The pad and envelope shall then be weighed.

4.4.5 Sterility. Sterility shall be determined in accordance with the current edition of the U.S.P.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:20Z
Check the source to verify that this is the current version before use.

4.4.6 Dermatology test. Dermatological reaction tests shall be performed by the method shown in the Journal of Investigating Dermatology, Volume 47, page 393, November 1966. The test shall be performed on a minimum of 25 adults. Sensitization by means of the standard aqueous solution shall not be done.

4.5 Certificates of compliance. Certificates of compliance shall be supplied as follows:

The manufacturer of the nonwoven cotton fabric shall supply a certificate stating that the cotton fibers, in the process of manufacture of the fabric, have been physically locked and entangled to form a curled and kinked configuration.

The manufacturer of the rayon nonwoven fabric shall supply a certificate stating the material has been made by needle looming.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One package containing 100 pads, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table III shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table III. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	10 units	40 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

5.1.3 Level A.

5.1.3.1 Individual package. Each impregnated pad shall be sealed in a foil-plastic laminate material that will maintain sterility of contents and preclude evaporation of the isopropyl alcohol solution unless the package is damaged or opened. The exterior of the package shall be either reflective or non-reflective.

5.1.3.2 Unit package. Unit package shall be a box of appropriate size and design constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors are acceptable. Closure shall be adequate to prevent spilling of contents under normal handling.

5.1.3.3 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class domestic. Closure of box shall be adequate to prevent spilling of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load, and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, grade V3c.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4 of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

title, number and date of this specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Fiberboard box shall be closed by method II of PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Immediate container. In addition to any commercial markings, each immediate container shall bear the manufacturer's name, the item identification or "Pad Alcohol Impregnated, or Trade Name of Manufacturer's commercially available product, Sterile", and the legend (or similar) - "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED".

5.3.2 Unit package. Each unit package shall be marked in accordance with MIL-STD-129. In addition, labeling shall comply with the requirements of the Federal Food, Drug and Cosmetic Act, and shall bear the name and address of the manufacturer; when the manufacturer is not the contractor, the name and address of the contractor shall also appear; contract or purchase order number; lot (control) number; and the method of sterilization (G-Gas, R-Radiation, S-Steam, D-Dry Heat sterilization) shall be shown following the level of packaging and date of packaging. Example: "A-3/71-G". Final sterilization shall be on or after indicated packaging date.

5.3.3 Intermediate package. Each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Lot (control) number, contract or purchase order number and name of contractor shall be shown. The method of sterilization as specified in 5.3.2 shall also be shown.

5.3.4 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Lot (control) number and the method of sterilization as specified in 5.3.2 shall also be shown.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use. The pad covered by this specification is intended for use on patients as a pre-injection skin preparatory.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) Selection of applicable level of packaging and packing (see 5.2.1 and 5.2.2).

6.3 The pad covered by this specification, is listed in the Federal Supply Catalog under the following stock number:

Federal Stock Number	Item Identification
6510-786-3736	PAD, ISOPROPYL ALCOHOL IMPREGNATED

6.4 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MS
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6510-0572

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION

MIL-P-36792

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426

REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:20Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS; DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATS
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:20Z
Check the source to verify that this is the current version before use.