MIL-R-0036837A (DSA-DM)
25 October 1972
USED IN LIEU OF
MIL-R-36837
31 May 1972

MILITARY SPECIFICATION

# RACK, TEST TUBE, LABORATORY, POLYPROPYLENE

This limited coordination Military Specification has been prepared by the Defense Personnel Support Center based upon currently available technical information, but has not been approved for promulgation as a coordinated revision of Military Specification MIL-R-36837 (DSA-DM). It is subject to modification. However, pending its promulgation as a coordinated Military Specification, it may be used in procurement.

## 1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers autoclavable plastic test tube racks that are suitable for use in medical laboratories.

1.2 Classification. Racks covered by this specification shall be of the following capacities:

Size 1 - 40 test tube capacity to accommodate test tubes 16 to 20 mm. in diameter.

Size 2 - 72 test tube capacity to accommodate test tubes 10 to 13 mm. in diameter.

## 2. APPLICABLE DOCUMENTS

2.1 Specifications and standards. The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

| | |
|---|---|
| PPP-B-585 | Boxes, Wood, Wirebound. |
| PPP-B-601 | Boxes, Wood, Cleated-Plywood. |
| PPP-B-621 | Boxes, Wood, Nailed and Lock-Corner. |
| PPP-B-636 | Box, Fiberboard. |

Military

| | |
|---|---|
| MIL-P-116 | Preservation, Methods of. |
| MIL-L-10547 | Liners, Case, and Sheet, Overwrap; Water Vaporproof or Waterproof, Flexible. |



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

| | |
|---|---|
| MIL-STD-105 | Sampling Procedures and Tables for Inspection by Attributes. |
| MIL-STD-129 | Marking for Shipment and Storage. |

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. The racks shall be fabricated from suitable polypropylene material that is capable of meeting the requirements of 3.3. If welding is used to bond components of the rack, the bonding material shall be the same material as the rack.

3.2 Design and construction. The racks shall be constructed in accordance with either Figure 1 (for size 1) or Figure 2 (for size 2), as applicable. Variations in contours of the rack from that indicated in the applicable figure are acceptable when not limited by requirements or dimensions. Each rack shall have a rectangular base of stiffened design and two upper tiers. The base and tiers shall be firmly secured to end support by means of either mechanical joints or welding. The upper and lower tiers shall have an equal number of properly aligned holes; and the base shall have the same number of countersunk holes to accommodate the test tube bottoms. The holes for the size 1 racks shall conform to the requirements of 3.2.1, and the holes for the size 2 racks shall conform to the requirements of 3.2.2. The center lines of corresponding holes in the base and tiers of the rack shall be in proper alignment to assure that the inserted tubes rest vertically in the rack. The cutout for each end support shall serve as a carrying handle. The top and bottom surfaces of the end support shall be designed to permit secure stacking of the same size test tube racks. The underside of the rack shall have cutouts to accommodate projecting tabs from end supports of bottom racks during stacking.


3.2.1 Size 1 racks. The size 1 racks shall have 40 holes which are arranged in 4 rows with 10 holes per row equidistant from each other. Each hole shall be 7/8 inch in diameter with a tolerance of +0 and -1/32 inch.

3.2.2 Size 2 racks. The size 2 racks shall have 72 holes which are arranged in 6 rows with 12 holes per row equidistant from each other. Each hole shall be 9/16 inch in diameter with a tolerance of +0 and -1/32 inch.

3.3 Performance. Each rack shall be capable of withstanding sterilization by autoclaving. When tested in accordance with 4.4.1, each rack shall not disassemble, become tacky, or functionally distort. When tested in accordance with 4.4.2, the base and tiers shall remain secured to end support without evidence of cracks or fractures to the plastic material.




Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.






Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.




NOTE:
AS AN ALTERNATE TO A CLEAR HOLE INDICATED IN SECTION A-A, A COUNTERSUNK HOLE CONTAINING A CROSS MEMBER SHALL BE ACCEPTABLE.

SIZE 2

ALL DIMENSIONS IN INCHES

FIGURE 2



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.

3.4 Workmanship. Workmanship shall be first class throughout, and each rack shall be free from defects which detract from its appearance or may impair its serviceability. All surfaces (including cutouts) shall be smooth and free of sharp edges, surface craze, mold flash, or projections and burrs (except for mold marks due to gating). Mold lines shall not be cause for rejection provided there are no flash protrusions. Burrs due to gating shall not be cause for rejection, provided there is no evidence of sharpness or projections that protrude more than 1/64 inch from the surrounding area. Each rack shall not wobble when stacked on top of another rack.

## 4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, or auditory, investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and Table I.

Table I. For examination

| | Inspection level | AQL (percent defective) |
|---|---|---|
| For visual examination: | | |
| Major defects | II | 1.0 |
| Minor defects | II | 2.5 |
| For dimensional examination | S-3 | 2.5 |



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and Table II. Acceptance number shall be zero.

Table II. For tests

| Characteristic | Requirement | Test procedure | Sample size |
|---|---|---|---|
| Autoclavability | 3.3 | 4.4.1 | * |
| Drop test | 3.3 | 4.4.2 | * |

* Four racks per lot of polypropylene material.

4.3 Examination. The racks shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with Table III. Examination shall not be restricted to the defects listed in Table III.

Table III. Classification of defects

| Categories | Defects |
|---|---|
| Major | |
| 101 | Rack not free from deformation or distortion. |
| 102 | Rack base and tiers not firmly secured to end supports. |
| 103 | Holes on shelves not properly aligned. |
| 104 | Rack not free of cracks, cuts, or depressions. |
| 105 | Rack and cutouts not free of sharp edges or sharp protrusions. |
| 106 | Rack not free of wobbling when stacked on top of another rack. |
| 107 | Rack not free from embedded foreign particles. |
| 108 | End supports not properly constructed to allow for proper stacking of racks. |
| Minor | |
| 201 | Holes on upper or lower tier not round. |
| 202 | Rack and cutouts not free of rough surfaces (except for mold marks per 3.4). |
| 203 | Rack not free from dirt, grease, or removable foreign particles. |
| 204 | Rack not free of flash. |
| 205 | Rack not free of surface craze. |



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.

4.3.2 Dimensional examination. The racks shall be examined for defects in dimensions. Any dimension not within the limits specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Autoclavability. The racks shall be subjected to steam sterilization in a suitable autoclave at 250° Fahrenheit (°F.) for a period of 30 minutes. After the sterilization period, the racks shall be removed from the autoclave and examined for evidence of damage. Slight distortion is permissible as long as it does not detrimentally affect the secure stacking of the test tube racks or the vertical insertion of the maximum diameter test tube into the rack. The test on the racks shall be repeated two additional times without any further distortion.

4.4.2 Drop test. A sample rack shall be dropped on each corner of the base from a height of 36 inches to a concrete floor. The center of gravity of the rack shall be directly over the corner which strikes the floor.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One rack, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging quantities

| Unit package | Intermediate package | Exterior container |
|---|---|---|
| 1 unit | 4 units | 16 units |

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method III.

5.1.3.2 Intermediate package. Intermediate package shall be a box of appropriate size constructed in accordance with PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent spilling of contents under normal handling.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Special requirements as specified in the applicable table in PPP-B-636 shall apply for fiberboard boxes.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of appicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4 of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure of wood boxes shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II of PPP-B-636. Special requirements as specified in the applicable table in PPP-B-636 shall apply for fiberboard boxes.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The racks are intended for holding test tubes for use in medical laboratories.

6.2 Ordering data. Purchasers should exercise any desired options offered herein, and the procurement documents should specify the following:

a. Title, number, and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
c. Federal Stock Number (FSN).
d. Size (see 1.2).

6.3 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Federal Government requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6640-0765

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

Defense Supply Agency

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-R-0036837A(DSA-DM)

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

**DD FORM 1 JAN 72 1426** REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.