MIL-T-36831A
16 June 1975
SUPERSEDING
MIL-T-36831
19 April 1972

MILITARY SPECIFICATION

TRAY, INSTRUMENT, PLASTIC

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers two sizes of plastic trays for storing and carrying medical instruments and supplies.

1.2 Classification.

Sizes

I. 16 inches long by 15 inches wide by 8 inches deep.
II. 16 inches long by 15 inches wide by 4 inches deep.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein.

SPECIFICATIONS

Federal

PPP-B-585	Box, Wood, Wirebound.
PPP-B-601	Box, Wood, Cleated-Plywood.
PPP-B-621	Box, Wood, Nailed and Lock-Corner.
PPP-B-636	Boxes, Shipping, Fiberboard.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWINGS

Defense Personnel Support Center

6061-3074	Tray, instrument, plastic; 2 inches deep.
6061-3075	Tray, instrument, plastic; 4 inches deep.

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS.

3.1 Material. Shall be a polycarbonate plastic, non-reactive with disinfectants and detergents. Shall be of the highest quality throughout and suitable in all respects for the purpose intended (see drawing and 6.5).

* 3.2 Design and construction. The tray shall be a one-piece lightweight unit. The trays shall be capable of holding water and of being stacked. The tray shall be capable of withstanding [illegible] sterilizing system without any sign of cracking, crazing, distortion or other visible deterioration when tested as specified in 4.4.1.

3.2.1 Size I tray. The size I tray shall be fabricated in accordance with DPSC Drawing No. 6061-3074.

3.2.2 Size II tray. The size II tray shall be fabricated in accordance with DPSC Drawing No. 6061-3075.

* 3.3 Workmanship. Workmanship shall be first class in every respect. The trays shall be free from defects which detract from their appearance or impair their serviceability. All corners and edges of the tray shall be smooth and free from sharp edges. Minor mold check marks on the exterior of the tray shall not be cause for rejection.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of [illegible] years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and Table I.

Table I. Sampling for Examination

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and Minor combined)	II	4.0
For dimensional examination	S-2	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 for an inspection level of S-1. The acceptance number shall be zero.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.

4.3 Examination. The trays shall be examined to determine compliance with all requirements contained in this specification.

* 4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects, as shown in Table II.

Table II. Classification of Defects

Categories	Defects*
Major	
101	Tray not as specified.
102	Tray not free of cracks, holes, crazing; or distortion.
103	Tray not free of sharp edges.
Minor	
201	Finish not free of scratches, mars, chips.
202	Not clean, free of grease, oil, dirt, other foreign matter.
203	Tray does not rest evenly on flat surface.

* Examination shall not be restricted to the classified possible defects listed above.

4.3.2 Dimensional examination. The Tray shall be examined for defects in dimensions.

* 4.4 Tests.

4.4.1 Cantining. Tray shall be placed in boiling water for 30 minutes. At the end of this period the tray shall be removed and allowed to dry. Cycle shall be repeated 5 times. Tray shall be examined for compliance with 3.2.

* 4.5 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table III. Sampling shall be in accordance with MIL-STD-105. The unit of product shall be an individual fully prepared for delivery. The number of units of the sample shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Source: https://assist.dla.mil -- Downloaded: 20261O-21T09:25Z
Check the source to verify that this is the current version before use.

Table III. Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and packing Materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size location, sequence, or method of application.

* 5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit package. Tray without further packaging, shall be the unit package.

5.1.2 Level C. Trays shall be prepared for shipment which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Twelve (12) trays of size 1 or 2 shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636b, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 3; or PPP-B-636, water-resistant, front vie. Closure and strapping shall be as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.

5.2.1.1 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-I-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof in accordance with the appendix of PPP-B-636.

5.2.2 Level B. Twelve (12) trays of size 1 or 2 shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, type CF, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. Trays shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, to point of delivery called for in the contract or purchase order.

5.3 Marking. Each unit package (tray) and exterior container shall be marked in accordance with MIL-STD-129. Marking of tray shall be accomplished by a pressure sensitive label.

6. NOTES

6.1 Intended use. For storing and transporting medical instruments and supplies in medical facilities.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(c) Size required.
(d) Applicable revision date and the required drawing number.

6.3 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.

6.4 This specification covers the following sizes of surgical instrument and dressing trays listed in the Federal Supply Catalog:

Stock No.	Item Identification	Specification Designation
6530-00-113-1401	TRAY, INSTRUMENT, Plastic, Green, 16 inches long by 15 inches wide by 8 inches deep.	I
6530-00-113-1402	TRAY, INSTRUMENT, Plastic, Green, 16 inches long by 15 inches wide by 4 inches deep.	II

6.5 Mobay Chemical Co.'s M50-4010 polycarbonate shall be acceptable as an alternate to the material specified in the drawing.

6.6 NSN 6530-00-141-9294, Cover, Surgical Instrument and Dressing Tray is for use with the Trays covered in this specification.

6.7 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 68

Preparing activity:
Defense Supply Agency - DH

Project No. 6530-1113

is

Source: https://assist.dla.mil -- Downloaded: 20260-21T09:25Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS. The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers, vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-T-36831A TRAY, INSTRUMENT, PLASTIC

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*	TELEPHONE NO
	DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.

FOLD

[illegible]

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, [illegible]
2800 SOUTH 20TH STREET
PHILADELPHIA, PA [illegible]

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:25Z
Check the source to verify that this is the current version before use.