MILITARY SPECIFICATION

PAD, POVIDONE-IODINE, IMPREGNATED

This specification is approved for use by the Defense Personnel Support Center, Defense Supply Agency, and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers Pad, Povidone-Iodine, Impregnated (see 6.2).

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-C-186	- Containers, Packaging and Packing for Drugs, Chemicals and Pharmaceuticals. (See 6.3)

STANDARDS

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.

(Copies of specifications and standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Hq, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document, or by letter.

FSC 6510

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE, FOOD AND DRUG ADMINSITRATION

Federal Food, Drug, and Cosmetic Act and Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, DC 20204.)

U. S. PHARMACOPEIAL CONVENTION, INC.

The United States Pharmacopeia.

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

National Formulary.

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

3. REQUIREMENTS

3.1 Material. Shall be a single-use pad impregnated with povidone-iodine solution.

3.1.1 Fabric pad. Shall be 1-3/8 ± 1/8 inches by 2 ± 1/8 inches in size and shall be made of a suitable grade of cotton as specified in 3.3.2.

3.1.2 Impregnating solution. Shall be Povidone-Iodine Solution, [illegible] of solution when tested as specified in 4.5.1.

3.1.3 Suitability. Shall be suitable for use in preparing small skin areas prior to injections, catheterizations and surgery, and in treating cuts, minor wounds, burns, lacerations, and abrasions.

Source: https://assist.dla.mil -- Downloaded: 20260-21T10:52Z
Check the source to verify that this is the current version before use.

3.2 Finished product.

3.2.1 pH. The povidone-iodine solution expressed from the pads shall have a pH of not more than 6.0 when determined potentiometrically at 25°C, using the USP method.

3.2.2 Available iodine. The povidone-iodine solution expressed from the pads shall contain not less than 8.5 mg and not more than 12.0 mg of available iodine per ml when assayed as specified in 4.5.2.

3.2.3 Identification. The infrared absorption spectra of a dried sample of povidone-iodine solution expressed from the pads in a potassium bromide dispersion at a concentration of 1 in 200 shall exhibit maxima and minima only at the same wavelengths as that of a previously submitted standard infrared curve.

3.2.3.1 A copy of the standard infrared curve (graph) shall be submitted by the bidder or offeror, and shall be kept on file in the Technical and Quality Assurance Division, Directorate of Medical Materiel, of the Defense Personnel Support Center for use if samples are submitted.

3.3 Ingredients.

3.3.1 Povidone-iodine solution. The povidone-iodine solution used in the impregnation of the pads shall comply with the tests, standards, and requirements of the USP, including any supplements or revisions thereto.

3.3.2 Cotton. The cotton used in the manufacture of the pads shall comply with the tests, standards, and requirements of the USP for purified cotton, including any supplements or revisions thereto, with the following exceptions. The requirements for sterility and fiber length shall be deleted. Not more than 0.5 percent of the residue shall remain from the Residue on Ignition test. The weight of the residue from the Water-soluble Substances test shall not exceed 75 mg (0.75 percent).

3.3.3 Other ingredients. All other ingredients used in the manufacture of the product shall be of NF or USP quality, or, if not included in either of these compendia, the ingredients shall be suitable for use in this preparation.

3.4 Delivery. Not more than 6 months shall have elapsed from the date of manufacture of the product to the date of delivery to the Government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.

3.5 Workmanship. The material and its containers shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and, (i) as to any expiration dated item, for the expiration dating period specified by the contractor for such item, or such longer period as may be required by regulation of any federal agency; or (ii) as to non-dated items, for not less than 3 years after delivery of the item to the Government.

4.1.2 No company supplying any ingredient(s) to the contractor will be considered an acceptable facility for the performance of any inspection requirements specified herein.

4.2 Lot. For purposes in this specification, a lot, batch, or control is that single, uniform, and homogeneous quantity of impregnated pads produced from one formulation, subjected to the same compounding and manufacturing operation, and filled into final containers.

4.3 Sampling. Sampling shall be conducted in accordance with MIL-STD-105.

TABLE I. Sampling.

For visual examination	Inspection level	AQL (percent defective)	Unit of product
Major	S-2	1.0	Unit of issue
Minor	II	2.5	Unit of issue

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.

4.3.1 For raw material and end item testing. Noncompliance with any of the following is cause for rejection.

TABLE II. For raw material and end item testing.

Characteristic	Requirement paragraph	Inspection level 1/	Minimum sample (individual pad)
Impregnating solution (volume)	3.1.2	-	10
pH	3.2.1	S-2	100 composite
Available iodine	3.2.2	-	20 composite
Identification	3.2.3	S-2	10 composite

1/ Where possible, the sample shall be used for two or more tests.

4.4 Examination. The finished impregnated pad shall be examined to determine compliance with all requirements of this specification. Nonconformance will be permitted to the extent indicated in 4.3 and 4.3.1.

4.4.1 Classification of defects. Examination will be conducted in accordance with the following classification of defects:

TABLE III. Classification of defects.

Categories	Defects 1/
Major	
101	Pad not impregnated with solution.
102	Dimensions not as specified in 3.1.1.
103	Foil packette incompletely sealed or torn.
Minor	
201	Exterior of immediate container is not free from contents or foreign matter.
202	Label not free from defacing marks and tears.

1/ Inspection is not restricted to the classified possible defects listed above.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.

4.5 Tests.

4.5.1 Impregnating solution. Open one finished foil packette and carefully pour any povidone-iodine solution into a 5 ml graduated cylinder that is marked in tenths of a milliliter. Express the solution from the impregnated pad into the same graduated cylinder. The total volume of the povidone-iodine solution in the cylinder shall be not less than 0.8 ml. A minimum of ten (10) packettes shall be tested.

4.5.2 Available iodine. Squeeze the povidone-iodine solution out of 20 packettes into a beaker and take a 5 ml aliquot of the solution and assay by the method specified in the U.S.P. for Povidone-Iodine Solution.

4.6 Packaging inspection. The inspection of the unit, intermediate package, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.7 Labeling inspection. Sampling shall be in accordance with MIL-STD-105 and the unit of product shall be the unit of issue. The inspection lot shall be the number of units of issue of each batch, lot, control number, etc., offered for inspection at one time. The inspection level for labeling shall be S-2 and the acceptance number shall be zero.

5. PACKAGING

5.1 Packaging requirements. The requirements for packaging shall be in accordance with all applicable requirements of PPP-C-186 (See 6.3) and as specified herein.

5.2 Packaging. Packaging shall be level A or C, as specified (See 6.1).

5.2.1 Unit of issue. One box containing 100 pads, as specified, constitutes one unit of issue.

5.2.2 Level A.

5.2.2.1 Immediate container. The immediate container shall be a commercially available hermatically-sealed foil packette considered acceptable by the Food and Drug Administration (FDA) for the product contained therein and shall be of such compatibility that neither the contents nor the container are altered in any way by each other.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.

5.2.2.2 Unit packaging. Each unit shall be packaged as specified in 5.2.5 of PPP-C-186.

5.2.2.3 Intermediate packaging. Ten (10) units of issue shall be packaged as specified in 5.2.7.1 of PPP-C-186.

5.2.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.3 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.3.1 Level A. Sixty (6) units of issue shall be packed as specified in 5.4.4 of PPP-C-186.

5.3.2 Level B. Sixty (60) units of issue shall be packed as specified in 5.4.3.1.1 of PPP-C-186.

5.3.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.4 Marking.

5.4.1 Labeling. Labeling shall be in accordance with the requirements of the Federal Food, Drug and Cosmetic Act and shall include the information required below:

5.4.1.1 Immediate container. Each immediate container label shall bear the following information:

(a) label information in accordance with commercial practice.

5.4.1.2 Unit package. Each unit package label shall bear the following information:

(a) label information in accordance with commercial practice.
(b) the National Stock Number.
(c) the date of manufacture prefixed by "MFD".

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.

NOTES: (1) The National Stock Number shall not be preceded by any prefix.
(2) Strip labeling may be used.
(3) Label adhesion test shall not apply.

5.4.2 Intermediate package. Each intermediate package shall be marked as specified in paragraph 5.5.3 of PPP-C-186. Date of manufacture shall be shown in lieu of date packed.

5.4.3 Exterior container. Exterior containers shall be marked as specified in 5.5.4 of PPP-C-186. Date of manufacture shall be shown in lieu of date packed.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) National Stock Number.
(c) Selection of applicable levels of packaging and packing (See 5.2 and 5.3).

6.2 This specification covers the following item(s):

National Stock Number	Item Identification
6510-01-010-0307	PAD, POVIDONE-IODINE, IMPREGNATED, Cotton, 1-3/8 by 2 inches, 100s

6.3 All references to PPP-C-186 refer to Interim Federal Specification PPP-C-00186B which is mandatory for use until superseded by a revision.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6510-0617

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-P-37503(DSA-DM) PAD, POVIDONE-IODINE, IMPREGNATED

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:52Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.