MILITARY SPECIFICATION

SPECIMEN KIT, URINE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers an integrated set of compatible, non-sterile components suitable for collecting, transporting, and testing of urine specimens.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-585	- Boxes, Wood, Wirebound.
PPP-B-601	- Boxes, Wood, Cleated-Plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	- Boxes, Shipping, Fiberboard.

MILITARY

MIL-L-10547	- Liners, Case, and Sheet, Overwrap; Water-Vaporproof, or Waterproof, Flexible.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, Code ATT, 2800 South 20th Street, Philadelphia, Pa. 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:51Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. Materials shall be as specified herein. Materials which are not specifically described herein shall be of the highest quality throughout and suitable in all respects for the purpose intended.

3.1.1 Tube. The tube shall be made of clear plastic capable of passing the centrifuge test specified in 4.4.1 without exhibiting any evidence of leakage or breakage.

3.1.2 Cap. The cap shall be made of opaque plastic.

3.2 Design and construction. The kit shall consist of 4 containers, each containing the following:

125 - Urine specimen tubes
125 - Snap caps
125 - Adhesive labels
125 - Urine specimen collection cups
1 - Holder for urine specimen tubes

The average count shall be not less than 1 holder and 125 of each other item per container, with not less than 1 holder and 122 of each other item in any individual container.

3.2.1 Tube. The urine specimen tube shall be made of clear plastic and shall have a capacity of 15 ± 1 milliliters (ml). The top of the tube shall be circular and it shall have a lip to accommodate the snap cap. The tube shall terminate in a point (may be slightly rounded) at the bottom to facilitate the collection of centrifuged precipitate. The walls of the tube shall be straight or shall taper inward toward the bottom. The tube shall have the maximum dimensions specified in Table I.

Table I. Maximum dimension of tube.

Dimension	Millimeters (mm)
Diameter of lip	25.0
Diameter of tube below lip	17.5
Length of tube	120.0

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:51Z
Check the source to verify that this is the current version before use.

The tube shall have graduated markings at the 1/4 or 1/2 ml. level and the 12 ml. level. Additional graduation marks shall be acceptable. The tube shall have no breaks, holes, cracks, or sharp edges.

3.2.2 Cap. The cap shall be fabricated of opaque plastic and shall have an integral lift tab or rim. It shall snap on and off the lip of the tube without undue effort and shall fit securely over the opening of the tube. The tube and cap shall show no evidence of leakage when tested as specified in 4.4.2. Cap shall have no breaks, tears, or sharp edges.

3.2.3 Label. The label shall be white and shall measure 2-1/2 + 1/4 inches by 1-1/4 + 1/4 inches. The label shall be imprinted with at least three lines for recording data. It shall accept and retain writing from a standard ball-point pen. The label shall be coated on one side with an all-purpose, permanent pressure-sensitive adhesive. The label shall not lift or separate from the tube when tested as specified in 4.4.5. The adhesive surface shall be protected by a strip of suitable removable backing-material. The label shall peel freely from the backing and shall adhere firmly to the specimen tube.

3.2.4 Cup. The cup shall have a capacity of not less than 3 ounces and shall be fabricated of a material suitable for the collection of urine. The cup shall have a flat bottom so that it can stand on a level surface when filled. Cups shall be constructed so that they can be nested to conserve storage space. The cup shall show no evidence of leakage when tested as specified in 4.4.3. The cup shall contain no tears or holes.

3.2.5 Tube holder. The tube holder shall be fabricated of cardboard, plastic, or other suitable material, free of tears or other deformations. It shall have inserts, approximately 1-1/2 inches deep, or at least two levels of concentric holes, to hold 10 or more specimen tubes in an upright position. It shall be supplied fully assembled, or shall be capable of being assembled, without the use of tools, fasteners, or adhesive not supplied, in three minutes or less. The holder shall show no evidence of spillage, deformation, or other failure when tested as specified in 4.4.4.

3.3 Workmanship. Workmanship shall be first class throughout. All components of the kit shall be free of dirt, oil, and grease and shall be free of defects which detract from their appearance or which may affect their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:51Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract,the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request. at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the raw materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component or material, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and Table II. The unit of product for examination shall be the individual tube, cup, cap, label and holder, No more than 10 of each component shall be selected from any unit package.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:51Z
Check the source to verify that this is the currentversion before use.

TABLE II. Sampling for examination.

	Inspection level	AQL (percent defective)
For visual examination		
Major A (cap)	I	1.0
Major B (Tube)	I	1.0
Major C (label)	I	1.0
Major D (cup)	I	1.0
Major E (holder)	I	1.0
Major F (complete kit)	S-3	2.5
Minor defects	I	4.0
For dimensional examination	S-1	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in table III. Inspection level shall be S-1 and an acceptance number of zero shall apply for all sample sizes.

TABLE III. Sampling criteria for tests.

Component (Unit of product)	Characteristic	Requirement	Test procedure
Tube	Centrifuge capability	3.1.1	4.4.1
Tube and cap	Leakage	3.2.2	4.4.2
Cup	Leakage	3.2.4	4.4.3
Holder	Durability	3.2.5	4.4.4
Label	Adhesion	3.2.3	4.4.5

4.3 Examination. The kits shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with table IV. Examination shall not be restricted to the classified possible defects listed below.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:51Z
Check the source to verify that this is the current version before use.

TABLE IV. Classification of defects

Categories	Defects
Major A	CAP
101	Not free of grease, oil, and dirt.
102	Not free of sharp edges.
103	Not free of breaks and tears.
104	Does not fit securely on specimen tube.
Major B	TUBE
111	Not free of grease, oil, and dirt.
112	Not free of sharp edges.
113	Not free of breaks, holes, and cracks.
114	Plastic cloudy; not clear.
115	Tube lacks specified graduation marks.
Major C	LABEL
121	Backing material cannot be peeled freely or is otherwise inadequate.
122	Does not adhere firmly to tube.
Major D	CUP
131	Not free of grease, oil, and dirt.
132	Cup has tears or holes.
133	Cup crushed or deformed.
Major E	HOLDER
141	Does not support tubes in an upright position.
142	Not free of tears.
143	Not capable of being assembled.
144	Does not accept tubes.
Major F	COMPLETE KIT
151	Any component not provided.
152	Count less than specified.
Minor	
201	Tubes not free of waviness, water marks, excessive scratches, surface cracks greater than 1/4 inch long.
202	Marking of components and packages not complete or not present.

4.3.2 Dimensional examination. All components shall be examined for defects in dimensions. Examination shall include measurements for the volumetric accuracy of the tube and the cup. Any dimension not within the tolerances specified herein shall be classified as a defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:51Z
Check the source to verify that this is the current version before use.

4.3.3 Packaging inspection. The inspection of the unit package, packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Centrifuge test. Tubes shall be filled with water to the 12 ml. graduation and placed in a centrifuge. The centrifuge shall be operated at 2200 revolutions per minute for a 10 minute period. The tubes shall then be removed from the centrifuge and examined for evidence of leakage or breakage.

4.4.2 Tube leakage test. Tubes shall be filled with water to the 12 ml. level. The caps shall be fastened on and all traces of water removed from the exterior of the tubes and caps. The tubes shall be held upside down for two minutes, and then placed lengthwise on a horizontal surface in a vacuum chamber and subjected to a negative pressure of 6 inches of mercury for one hour at normal room temperature (65° F to 90° F). At the end of the hour the tubes shall be examined for any evidence of leakage.

4.4.3 Cup leakage test. Cups shall be filled with 2 1/2 fluid ounces of water and left standing upon a level surface for 2 hours at room temperature. Cups shall then be examined for any evidence of leakage.

4.4.4 Tube holder test. Ten tubes shall be filled to the 12 ml. level and placed in the holder. Each tube shall be removed and replaced twice, and allowed to remain in the holder for 2 hours. After 2 hours, the tube holder shall be examined for any evidence of failure to support the tubes upright without spillage.

4.4.5 Label adhesion. Specimen tubes shall be filled with 12 ml. of water at room temperature and capped. A label shall be applied firmly to each tube. The tubes shall be refrigerated at 40° F or less for a period of at least 10 hours. Tubes shall then be centrifuged at 2200 revolutions per minute for a 10 minute period. The tubes shall then be removed from the centrifuge and examined for any evidence of labels lifting from the tube surface.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:51Z
Check the source to verify that this is the current version before use.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. One package consisting of four containers each containing 125 caps, 125 cups, 125 labels, 125 tubes and one holder, as specified, constitutes one unit of issue.

5.1.1.1 Level A.

5.1.1.1.1 Unit packaging. One hundred twenty-five of each component (caps, cups, labels and tubes) and one tube holder shall be neatly and orderly packaged in an interior box of appropriate size and constructed in accordance with PPP-B-566. Four such cartons shall be packaged in a box of appropriate size and constructed in accordance with PPP-B-636, type CF, class domestic. Closure of the box shall conform to the appendix of PPP-B-636.

5.1.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Two (2) units of issue shall be packed in a box designed for a type 2 load, and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2 or PPP-B-636, class weather resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof in accordance with the appendix of PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:51Z
Check the source to verify that this is the current version before use.

5.2.2 Level B. Two (2) units of issue shall be packed in a box designed for a type 2 load, and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, type CF, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior box, unit package and exterior container. Each of the four interior boxes of each unit shall be marked, in addition to any commercial marking, with the identity and quantity of the contents. Each unit package and exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The kits are intended for use in military medical facilities.

6.2 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification.
b. Selection and applicable levels of packaging and packing.

6.3 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6530-00-075-6636	SPECIMEN KIT, URINE

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6530-1387

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:51Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:51Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-S-37498 SPECIMEN KIT, URINE

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:51Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:51Z
Check the source to verify that this is the current version before use.