MILITARY SPECIFICATION

SINK SERVICE UNIT, CLINICAL LABORATORY

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one type of sink service unit which contains two each cup-sinks, water faucets, water hoses, and a drain line adapted for the connection of two mating sinks (see 6.1).

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bid or request for proposal form a part of this specification to the extent specified herein.

SPECIFICATIONS

MILITARY

MIL-M-7298	- Manual, Technical, Commercial Equipment.
MIL-A-8625	- Anodic Coatings, for Aluminum and Aluminum Alloys.

STANDARDS

FEDERAL

FED-STD-595	- Colors.

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.

MIL-STD-401 - Sandwich Constructions and Core Materials General Test Methods.
MIL-STD-810 - Environmental Test Methods.

DRAWINGS

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

DPSC 5061-3540 - Sink Service Unit and Bill of Materials.

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.1.2.1 and 6.2).

3.2 Description. The sink service unit shall be equipped to connect to the drain and water supply in the expandable, Medical Unit, Self-contained Transportable (MUST) shelter. The sink shall be equipped to distribute water to two integral cup sinks and two mating double-basin sink units; and through two integral retractable hose units to any location within the shelter. Two swivel gooseneck spouts shall provide water to the basins of the mating sinks and the integral cup sinks. Flow check valves shall be provided to prevent back flow or siphoning of water. Two flexible connection's in the drain line shall be provided to service the drains of the two mating sink units. A one-inch, flare type fitting shall be provided for connection to the shelter sump-pump.

3.3 Materials. The sink bowls and drain line shall be constructed of durable, chemical-resistant plastic. Water tubing and connectors shall be constructed of aluminum alloy. The component items of sink service unit shall conform to the requirements of the bill of materials to DPSC Drawing 5061-3540.

3.4 Design and construction. The design and construction shall be in accordance with DPSC Drawing 5061-3540. The outside dimensions of the basic unit, excluding plumbing items, shall be a nominal 58 inches wide by 36 inches high by 9 inches deep. Exact dimensions shall be in accordance with the drawings. The net weight of the sink service unit shall not exceed 105 pounds.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.

3.5 Plumbing. All pipes, valves, fixtures, and fittings shall be suitably secured in position in order to reduce strain at the connections to the sheet metal, and to prevent vibration of the plumbing. The fixture handles shall be marked to indicate either hot or cold water. The complete plumbing system shall be capable of being drained to prevent water from remaining in any portion of the system while in transit. All plumbing valves and fixtures shall be fabricated of bronze or brass, and shall be finished with a plating of chromium over nickel.

3.5.1 Drain line leakage. There shall be no leakage when unit is tested as specified in 4.4.3.

3.5.2 Supply line leakage. There shall be no leakage when unit is tested as specified in 4.4.2.

3.6 Bond strength. When the honeycomb panels are tested as specified in 4.4.1, the average peeling torque per inch of specimen width shall not be less than 9 inch-pounds prior to the humidity test, and not less than 8 inch-pounds after the humidity test. These peeling torque values shall include the torque required to bend the panel facing.

3.7 Finish. Exposed surfaces of corrosion-resisting steel, excepting hardware items, shall have a Commercial Standard Number 4 finish. When chrome plating is specified, chromium shall be plated over nickel. Nickel plating shall be a minimum 0.0004 inch in thickness and chromium plating shall be a minimum 0.00005 inch in thickness. When anodizing is specified, anodizing shall conform to Type II, Class 1 of MIL-A-8625; anodized parts shall not be painted. The underside of the sink service unit shall be painted. Prior to painting, steel parts shall be phosphatized; aluminum, zinc and cast-iron parts shall be phosphatized or washprimed (see 6.4). Primer coat shall be applied to a dry film thickness of 0.0004 to 0.0007 inch (see 6.5). A color coat shall be applied to a wet thickness of 0.004 to 0.006 inch (see 6.6). A final, clear texture coat shall then be applied to achieve a standard texture, and the unit allowed to air dry (see 6.7). Color shall be 24516 of FED-STD-595. Paints shall be applied in accordance with the manufacturer's directions.

3.8 Identification marking. Each unit shall be marked permanently and legibly with the following:

Contract number.
Name or registered trademark of the manufacturer.
Manufacturer's model number.
National Stock Number (NSN).
Serial number.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.

3.9 Welding. Welding shall be durable, sound in every detail and in accordance with the latest practices recommended by the American Welding Society. All welds shall be produced by qualified welding operators with satisfactory welding equipment using satisfactory filler metal or electrode where applicable to provide the best possible weld. Surfaces to be welded shall be free from grease, oil, oxides or any other foreign matter. Any cleaning method which will remove these substances and leave the surface clean is satisfactory. All weld joints shall blend into the adjacent metal in gradual, smooth curves. Welds shall be smooth and free of slag, excessive undercut, or excessive splatter. Sufficient metal shall be added to provide a suitable fillet or reinforcement unless otherwise specified. Welds shall not be oxidized. Welds shall have no visible voids, pin holes, or cracks to the unaided eye. Welds, except spot welding shall be continuous.

3.10 Service data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance and lists of component parts. The manual shall comply with the requirements of MIL-M-7298. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer, via the cognizant Government inspector, two preliminary copies of his proposed technical manual for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract, or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of the contract and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies shall be at the option of the Government.

3.11 Workmanship. Workmanship shall be first class throughout. Item shall be free from defects which detract from its appearance or impair its serviceability. All burrs and sharp edges shall be removed from the item.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.

unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2.1 Classification of inspection. The inspection requirements specified herein are classified as follows:

a. First article inspection.
b. Quality conformance inspection.

4.1.2.1.1 First article inspection. First article inspection shall be performed on the sink service unit when a first article sample is required (see 3.1). This inspection shall include examination of 4.3 and tests of 4.4.

4.1.2.1.2 Quality conformance inspection. Quality conformance inspection shall include the examination of 4.3 and tests of 4.4.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the component/materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government, and provided also that the test methods and the results obtained comply with those described in this specification. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for examination shall be one sink service unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.

TABLE I. Sampling for examination.

	Inspection level	AQL (defect per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and Minor combined)	II	4.0
For dimensional defects	S-3	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with table II. The unit of product for test purposes shall be one sink. Acceptance number shall be zero. Where practicable, one or more test characteristics shall be conducted on the same sample unit.

TABLE II. Sampling for tests.

Characteristic	Requirement	Test Method	Inspection
Bond strength	3.6	4.4.1	(see 4.4.1)
Supply line leakage	3.5.2	4.4.2	Inspection level 100 percent, reject all defective units
Drain line leakage	3.5.1	4.4.3	Inspection level 100 percent, reject all defective units

4.3 Examination. The sink service unit shall be examined to determine compliance with all requirements specified herein. Examination shall be conducted as specified in table III. Examination however, shall not be restricted to defects listed in table III.

TABLE III. Classification of defects

Categories	Defects
Major	
101	Not style and design specified.
102	Not free of cracks, holes, tears.
103	Any component missing, broken or otherwise unuseable.
104	Welding not as specified.
105	Adhesive not properly applied.
106	Drain line joints not properly heat fused.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects. - Continued

Categories	Defects
Minor	
201	Identification marking not present, complete, legible, correct.
202	Finish not free of scratches, mars, chips.
203	Not clean, free of grease, oil, dirt, other foreign matter.
204	Movable parts do not operate properly.
205	Latches not as specified.
206	Color not as specified.
207	Service data not as specified.
208	Fasteners or rivets not tight, properly applied.
209	Not free of sharp edges.

Note: Periodic in-process inspections shall be made to insure conformance to adhesive requirements specified herein.

4.3.1 Dimensional examination. The sink service unit shall be examined defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.3.2 Packaging inspection. The inspection of the unit package, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and marking requirements of MiL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Bond strength. Six specimens which are representative of the honeycomb construction being used shall be selected and tested as specified in the Sandwich peel test of MIL-STD-401. Two specimens shall be taken from each of three bonded panels for each lot of adhesive. One half of the samples shall be tested in the "as received" condition and the other half shall be tested after subjecting to the humidity test in method 507 of MIL-STD-810, procedure I. In each testing sequence, the test results from the three specimens shall be averaged to detemine conformance to 3.6. The peeling torque values obtained in the test and specified herein shall include the torque required to bend the aluminum facing of the honeycomb panel. The pulling force shall be applied in the L or ribbon direction of the paper core.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.

4.4.2 Supply line leakage. The hot and cold water lines shall be connected to a water source at 30 psig. Water shall be run through the lines to purge any entrapped air. The faucet valves shall be closed and the 30 psig water pressure applied for 10 minutes. During this period, the lines and fixtures shall be observed, and any evidence of leakage shall constitute failure.

4.4.3 Drain line leakage. The drain line of the assembled sink service unit shall be capped or plugged. A stopper shall be placed in the "stop" position in each cup sink. Each cup sink shall be filled with water to within one inch of the top. The water shall remain for 5 minutes and any leakage shall be noted by observing the water level in the cup sinks. At the end of the 5 minute period the stoppers shall be removed. The one-inch drain outlet shall then be uncapped and the water permitted to drain, until the cup sinks are approximately one half full. The outlet shall be recapped for a 5 minute period. During this time the drain lines shall be observed, and any evidence of leakage shall constitute failure.

5. PACKAGING

5.1 Packaging. Packaging shall be level C (see 6.2).

5.1.1 Level C. Each sink service unit shall be clean, dry and packaged in a manner that will afford adequate protection against corrosion, deterioration and physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level C, as specified (see 6.2).

5.2.1 Level C. Each packaged sink service unit shall be packed in a shipping container in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the carrier's rules and regulations.

5.3 Marking.

5.3.1 Interior packages. Each package shall be marked to identify contents and quantity contained.

5.3.2 Shipping container. Each shipping container shall be marked as specified in MIL-STD-129.

6. NOTES

6.1 Intended use. This unit is intended for use in the Medical Unit Self-contained Transportable (MUST) Clinical Laboratory to dispense water and serve as a drain for liquid wastes. The sink service unit will also connect with and serve two clinical laboratory sinks, NSN 6545-00-027-5009, each of which contains double-basin sink and a glassware drying rack.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information on the procurement documents:

a. Title, number and date of this specification.
b. National Stock Number.
c. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
d. When a first article is required (see 3.1, 4.1.2.1 and 6.4).

6.3 This specification covers the following item listed in the Federal Supply Catalog:

National Stock Number	Item Identification
6545-00-019-9330	Sink Service Unit, Clinical Laboratory, Hot and Cold Water Supply.

6.4 An acceptable phosphate treatment is Parker Rust Proof Co. Bonderite 100, or equal; an acceptable wash primer is John L. Armitage and Co. P-147, or equal.

6.5 An acceptable primer coat is Armitage and Co. P-289, or equal.

6.6 An acceptable color coat is Armitage and Co. K-497, or equal.

6.7 An acceptable texture coat is Armitage and Co. M-516, or equal.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DOD - MB

Preparing activity:
DLA - DM

Project No. 6545-0029

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS: This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.

NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER *(Number)* AND TITLE

MIL-S-37505 SINK SERVICE UNIT, CLINICAL LABORATORY

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1. ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*

TELEPHONE NO.

DATE

DD FORM 1 OCT 76 1426 EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:52Z
Check the source to verify that this is the current version before use.