A-A-0051847(DM)
3 March 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

INDICATOR, STERILIZATION

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This interim commercial item description covers a package of 1000 ethylene oxide (ETO) sterilization indicators.

Salient characteristics:

The unit shall be a single use sterilization indicator for use in ethylene oxide (ETO) sterilizers.

The sterilization indicator shall be capable of monitoring the four major parameters of ETO sterilization of time, temperature, gas concentration, and moisture.

The sterilization indicator shall present the results in a graphic easy-to-read manner that requires no external processing of the indicators to obtain said results.

The sterilization indicator shall be designed to be left inside the sterilized package until the package is opened for use. Accurate results shall be obtained upon completion of a normal aeration cycle or when held at room temperature for a minimum of 24 hours after the sterilization cycle. Prolonged aeration or storage shall have no effect on the indicator or its reading.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

The indicator shall respond to the type of cycles described herein as specified:

Cycle Description	Concntrat. (mg/L)	Percent RH	Temp. (^{O}F)	Exposure Time (Min)	Reading
Standard cycle	600	50	130	60	Accept
Marginal time*	600	50	130	30	Reject
Insufficient time	600	50	130	20	Reject
Low concentration	200	50	130	60	Reject
Low temperature	600	50	100	60	Reject
Low percentage of relative humidity	600	20	130	60	Reject

* When compared to readings obtained by spore strips containing 1.3×10^{6} spores per strip, the indicator gage shall reject other marginal cycles at least as often.

The sterilization indicator shall be packaged in quantities of 1000 indicators per container.

Workmanship. The indicator shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing on thousand indicators, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-233-4404

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC

Project No. 6530-1909

Location: CID/INDICATOR/H1B-16

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:02Z
Check the source to verify that this is the currentversion before use.