A-A-0051849(DM)
11 March 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

ELEVATOR AND RETRACTOR, BONE (Bennett Design)

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a bone elevator and retractor of the Bennett design used during surgical procedures.

Salient characteristics.

General. The bone elevator and retractor shall be fabricated in accordance with the applicable requirements of GG-I-526, and as specified herein.

Material. The bone elevator shall be fabricated Class 3 corrosion-resistant steel in accordance with ASTM F899.

Hardness. The Rockwell hardness of the bone elevator shall be B75 to B95 when tested in accordance with ASTM Standards E18 or E92. Conversion of Vickers hardness values to Rockwell values shall be in accordance with ASTM Standard E140. The requirements of GG-I-526 shall apply.

Style, design and dimensions. The bone elevator and retractor, Bennett design, shall be in strict accordance with Defense Personnel Support Center (DPSC) Drawing #10056.

Finish. Finish shall be entirely mirror, satin or matte. The matte finish shall exhibit a uniform nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure.

Passivation. The instrument shall be properly passivated.

Corrosion-resistance. The instrument shall comply with the copper sulfate and boil test corrosion resistance requirements of Federal Specification GG-I-526.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Identification markings. In addition to the requirements specified in GG-I-526, bone elevator and retractors of foreign manufacture shall be marked "stainless."

Workmanship. The bone elevator and retractor shall be free from defects which detract from its appearance or impair its serviceability.

Product Quality Level

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One bone elevator and retractor, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one elevator and retractor. The inspection level shall be 1.0. The test procedure shall be in accordance with GG-I-526 and as specified herein.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. The inspection level for a major type of defect shall be II and the AQL (percent defective) shall be 1.0. The inspection level for a minor type defect shall be II and the AQL shall be 2.5.

Examination of bone elevator and retractor. The bone elevator and retractor shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Table I. Examination, however, shall not be restricted to defects listed in Table I. Any bone elevator and retractor in the sample containing one or more defect shall be rejected and if the number of defective bone elevators and retractors in any sample exceeds the acceptance number for the sample, the lot represented by that sample shall be rejected.

TABLE I. Classification of defects.

Categories	Defects
Critical	None defined.
Major	
101	Bone elevator and retractor not of the specified design for appropriate type, stock number and item identification.
102	Finish of bone elevator and retractor not uniform and not as specified.
Minor	
201	Bone elevator and retractor not free from dirt, polishing material or foreign matter.
202	Identification markings not present, not current, not complete, or not legible.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-361-8200	ELEVATOR AND RETRACTOR, Bone, Bennett

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:02Z
Check the source to verify that this is the currentversion before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4031

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.