A-A-0051851(DM)
19 March 1987

# INTERIM COMMERCIAL ITEM DESCRIPTION

## NEEDLE, HYPODERMIC, DISPOSABLE
## 30 Gage By 1/2 Inch (1.3 cm)

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a sterile 30 gage disposable hypodermic needle.

Salient characteristics:

The needle is a 30 gage by 1/2 inch (1.3 cm) hypodermic needle, the hub of which shall be in accordance with ANSI/HIMA MD70.1 and have a beige color. The cannula shall be made of Type 304 austenitic stainless steel in accordance with ASTM F899. The hub shall be made of polypropylene and be permanently affixed to the cannula. The hub and adhesive shall be free from all elutable toxic materials. The cannula exterior shall be coated with an inert, non-toxic silicone lubricant. Cannula and point dimensions shall be as follows:

| Gage | O.D. (inch) | I.D. minimum (inch) | Point Length (inch) |
|---|---|---|---|
| 30 | 0.0115 to 0.0125 | 0.006 near point | 0.047 to 0.063* |

* Measured from heel to tip parallel with longitudinal axis of cannula.

The cannula length shall be 1/2 + 1/32, -1/16 inch. The needle shall be free from foreign matter, chips and burrs. There shall be no separation of the cannula from the hub under a 5-lb. tensile load sustained for one minute. The needle shall be pyrogen-free and non-toxic in accordance with the USP. The needle shall be sterile and contained in an individual sealed, sterile package.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Stiffness test. The needle shall undergo a deflection of no more than 0.016 inch when tested under a 20-ounce load and a separation distance of 0.197 inch in accordance with 4.4.11 of MIL-N-36124B.

A. Sterility test. The needle shall be tested for sterility in accordance with the method specified in the USP using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it is clearly differentiated from product intended for distribution. The incubation medium for the paper spore strip (carrier) shall be soybean-casein digest. The incubation medium for the inoculated product, or inoculated simulated product shall be fluid thioglycollate or alternative thioglycollate at 30° to 35° C for one half of the specimens and soybean-casein digest at 20° to 25° C for the remaining half of the specimens. The minimum incubation time shall be ten days. Sampling shall be as follows:

| Medium | Paper spore strip | Inoc. Prod. or Inoc. Simulated Prod. |
|---|---|---|
| Soybean-casein digest | S-3* | S-3* |
| Fluid thioglycollate or Alternative thioglycollate | | S-3* |

*Use a minimum of 20 specimens. Acceptance number is zero.

B. Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

| Type of Sterilization | Organism |
|---|---|
| Steam | B. stearothermophilus |
| Dry Heat | B. stearothermophilus |
| | B. subtilis |
| Ethylene oxide | B. subtilis |
| Gamma radiation | B. pumilus |

The number of viable spores shall be sufficient to provide assurance that the product has been sterilized.



Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

C. Inoculated carrier. Inoculated carriers (paper spore strips) if used, shall be in accordance with MIL-S-36586A, Sterilization Test Strip Set, Bacterial Spore. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in A. Sterility test and B. Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in B. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP and paragraphs A and B.

D. Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

E. Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP. The inspection level shall be as stated in paragraph A above.

Workmanship. Workmanship shall be first class throughout. The needles shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing 100 individually packaged sterile needles, as specified, constitutes one unit. The individual package shall be a sealed, peel-open type container capable of maintaining sterility of contents unless package is damaged or opened and of protecting needle point from damage and preventing point from piercing container.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one needle.

Sampling criteria for tests

| Characteristic | Inspection level | AQL |
|---|---|---|
| Security of attachment of hub | S-1 | * |
| Non-pyrogenicity | S-1 | * |
| Non-toxicity | S-1 | * |
| Sterility | (See sterility clauses) | |
| Stiffness | S-1 | * |

*Acceptance number shall be zero for all sample sizes.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.



Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

| NSN | ITEM IDENTIFICATION |
|---|---|
| 6515-01-126-4320 | NEEDLE, HYPODERMIC, Disposable, 30 Gage by 1/2 Inch (1.3 cm), 100s. |

Ordering data (Intermediate/exterior packaging quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4028



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.