A-A-51851A
14 May 1990
SUPERSEDING
A-A-0051851(DM)
19 March 1987

COMMERCIAL ITEM DESCRIPTION

NEEDLE, HYPODERMIC, DISPOSABLE
30 Gage By 1/2 Inch (1.3 cm)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a sterile 30 gage disposable hypodermic needle.

Salient characteristics:

The needle is a 30 gage by 1/2 inch (1.3 cm) hypodermic needle, the hub of which shall be in accordance with ANSI/HIMA MD70.1 and have a beige color. The cannula shall be made of Type 304 austenitic stainless steel in accordance with ASTM Designation F899. The hub shall be made of polypropylene and be permanently affixed to the cannula. The hub and adhesive shall be free from all elutable toxic materials. The cannula exterior shall be coated with an inert,, non-toxic silicone lubricant. Cannula and point dimensions shall be as follows:

Gage	O.D. (inch)	I.D. minimum (inch)	Point Length (inch)
30	0.0115 to 0.0125	0.006 near point	0.047 to 0.63*

* Measured from heel to tip parallel with longitudinal axis of cannula.

The cannula length shall be 1/2 +/- 3/32 inch. The needle shall be free from foreign matter, chips and burrs. There shall be no separation of the cannula from the hub under a 5 lb. tensile load sustained for one minute (See Note 3). The needle shall be pryogen-free and non-toxic in accordance with the USP. The needle shall be sterile and contained in an individual sealed, sterile package.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Stiffness test. The needle shall undergo a deflection of no more than 0.016 inch when tested as specified. The cannula shall be placed across two test edges (0.002 inch radius maximum) separated by the distance of 0.197 inch. A load of 20 ounces shall be applied at the center of the span and the resulting maximum deflection of the cannula shall be measured using a gage with graduations of at least a maximum radius of 0.026 inch, and centered to within +/- 0.003 inch (See Note 2).

Sterility test. The needle shall be tested for sterility in accordance with the USP or AAMI requirements. Sterility assurance level shall be 0.000001.

Workmanship. Workmanship shall be first class throughout. The needles shall be free from defects which detract from their appearance or impair its serviceability,

Unit. Package (PG). One package containing 100 individually packaged sterile needles, as specified, constitutes one unit. The individual package shall be a sealed peel-open type container capable of maintaining sterility of contents unless package is damaged or opened and of protecting needle point from damage and preventing point piercing container.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more text characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 with an Inspection level of S-1 and an AQL of 1.0. The unit of product for test purposes shall be one needle.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers'own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and /or order.

NOTE 1: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-126-4320	NEEDLE, HYPODERMIC, Disposable 30 Gage by 1/2 Inch (1.3 cm), 100s.

NOTE 2: A maximum resistance (to bending) force test which can be corrected to give satisfactory stiffness test results is acceptable as an alternate.

NOTE 3: A straight-away pull to destruction test whose results can be corrected to show adequate suspended weight test results is acceptable as an alternate.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

Ordering data (Intermediate/exterior packaging quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4751

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.