A-A-0051852(DM)
4 March 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

ADAPTER, AUTOMATIC RETRACTOR HOLDER

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers an adapter used for attaching a retractor to an automatic retractor holder.

Salient characteristics.

General. The adapter shall be a universal connection for adapting standard retractors to automatic retractor holders (See Note 2).

Material. The adapter shall be fabricated Class 3 austenitic corrosion-resistant steel in accordance with ASTM F899.

Hardness. The Rockwell hardness of the adapter shall be B75 minimum. The requirements of GG-I-526 shall apply.

Style, design and dimensions. The adapter shall be in accordance with DPSC Drawing #24503. The adapter shall accommodate retractors having handles up to 5/8 inch. The wing nut shall expand four inches. The overall dimension shall be 4-3/4 inches.

Finish. The adapter shall possess a smooth mirror finish.

Passivation. The adapter shall be properly passivated.

Corrosion-resistance. The adapter shall comply to the boil test corrosion resistance requirements of Federal Specification GG-I-526.

Identification markings. In addition to the requirements specified in GG-I-526, adapters for foreign manufacture shall be marked "stainless."

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Workmanship. The adapter shall be free from defects which detract from its appearance or impair its serviceability.

Product Quality Level

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One adapter, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one clamp. The inspection level shall be 1.0. The test procedure shall be in accordance with GG-I-526 and as specified herein.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Inspection level shall be S-2 with an AQL of 1.0 (percent defective). The unit of product for tests purposes shall be one adapter.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

NOTE 1: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-100-5561	ADAPTER, AUTOMATIC RETRACTOR HOLDER, 3/1-2 by 1-7/8 by 3/4 Inch (8.89 by 4.76 by 1.90 cm)

NOTE 2: The adapter shall attach a retractor to the automatic retractor holder (NSN 6515-01-100-5560).

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4006

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.