A-A-0051860(DM)
20 March 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

PARTITION, CASE, MEDICAL, Aluminum

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a partition suitable for use with Case, Medical, Shipping and Storage, NSN 6545-01-238-5597 and NSN 6545-01-238-5598, for the division of the Case interior into adjustable sized compartments.

Salient characteristics:

Shall be a movable, rigid, aluminum partition, with polyethylene foam cushioning and spring loaded latches, able to be locked into the interior framework of the Medical Case without the use of tools.

The partition shall be constructed of 3/16 inch (0.48 cm) thick American Standard Construction grade aluminum in accordance with DPSC drawing number 24539 dated 21 October 1986. The aluminum shall have anodic coating in accordance with MIL-A-8625 Type II.

The partition shall be equipped with a layer of 1 inch (2.54 cm), flexible, 2 pound density, polyethylene foam cushioning, NSN 8135-01-250-0472, affixed to each side with pressure sensitive adhesive. The cushioning shall be 15 inches (38.10 cm) wide by 16 inches (40.64 cm) high for NSN 6515-01-241-7021, or 15 inches (38.10 cm) wide by 9-3/4 inches (24.77 cm) high for NSN 6515-01-241-7040. There shall be a cut-out for the spring loaded latch. The cushioning shall be centrally positioned on the partition surface such that the edges are not covered, to allow unhindered insertion into the slots of the interior frame.

Partition NSN 6515-01-241-7021 shall be for use in Medical Case NSN 6545-01-238-5597, and Partition NSN 6515-01-241-7040 shall be for use in Medical Case NSN 6545-01-238-5598.

AMSC/NA FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Workmanship. The partition shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One partition, as specified, constitutes one unit.

Reference standard. Where a standard or drawing is referenced herein, the issue of that standard or drawing in effect on the date of solicitation shall be used to determine compliance with stated requirements.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspections and tests. Sampling for inspection and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one partition.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6515-01-241-7021	Partition, Case, Medical, 15 by 16 Inch
6515-01-241-7040	Partition, Case, Medical, 15 by 9-3/4 Inch

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4035

Location: NCR Disk #2

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.