A-A-51862A
18 August 1988
SUPERSEDING
A-A-0051862(DM)
17 March 1987

COMMERCIAL ITEM DESCRIPTION

GUIDE WIRE, VEIN CATHETERIZATION

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a guide wire suitable for use in vein catheterization procedures.

Salient characteristics:

General. The guide wire shall be a sterile, straight, fixed core guide wire. The guide wire shall have dimensions of 0.025 plus zero, minus 0.001 inch diameter and 12 plus 1/4, minus 1/2 inches length.

Material. The inner fixed core and the outer spiraled wire shall be fabricated from type 302 austenitic corrosion-resisting steel. The composition of the corrosion-resisting steel shall be in accordance with American Society for Testing and Materials (ASTM) Designation F899.

Style and design. The guide wire shall be constructed of tightly spiraled wire over a fixed core surface. Approximately 1.5 inches of each end shall be flexible and the tips shall be rounded and smooth to reduce incidence of inadvertent vessel puncture.

Sterility. Each guide wire shall be individually wrapped in a suitable, tear-resistant peel-open package which will allow aseptic handling and maintain sterility of the content until opened or damaged. The guide wire shall meet the sterility requirements when tested in accordance with the USP.

Toxicity and pyrogenicity. The guide wire shall be non-toxic and pyrogen-free in accordance with the USP, and shall meet any Federal Food and Drug Administration (FDA) requirements for safety and toxicity applicable for invasive and implantable devices.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Workmanship. The guide wire shall be free from defects which detract from its appearance or impair its serviceability.

PRODUCT QUALITY LEVEL

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Package (PG). One package containing fifty individually packaged sterile guide wires, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:02Z
Check the source to verify that this is the currentversion before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for tests purposes shall be one guide wire.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-119-6386	GUIDE WIRE, VEIN CATHETERIZATION, Disposable, 0.025 Inch (0.64 mm) Diameter, 12 Inches (30.5 cm) Long.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4386

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.