A-A-51871
2 May 1988

COMMERCIAL ITEM DESCRIPTION

DRAINAGE SET, BLADDER, (Suprapubic, Reif, 8 Fr)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a sterile, disposable suprapubic bladder drainage set used in gynecological procedures.

Salient characteristics.

General. The suprapubic bladder drainage set shall include a catheter, body seal, 3 cc vial of medical adhesive, a trocar, a cannula, a 3-way stopcock and an adhesive applicator for use during prolonged bladder drainage and various gynecological procedures.

Material. The trocar and cannula shall be fabricated from martensitic corrosion-resisting steel. The catheter and body seal shall be fabricated from transparent silicone. Portions of the stopcock shall be fabricated from flexible and rigid plastic. The injection site shall be molded from natural rubber. The plastic, rubber and silicone components coming in contact with the fluid when the injection site is used to irrigate the bladder shall meet the requirements of the Pharmacopeia of the United States (USP) for Class II plastics. The silicone catheter shall meet the requirements of the USP for Class IV plastics.

These rubber, silicone and plastic components coming in contact with the fluid shall contain not more than 5 parts per million total of lead, barium and cadmium.

Style and design. The silicone catheter shall be size 8 French and shall be 20 + 4, -1 inches long. There shall be two eyes within 3/4 inch of the distal end of the catheter.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

A 10 gauge cannula and trocar shall be provided. The trocar shall be sharp at the distal end with a minimum tapered length of 0.29 inch and provide a means of securely gripping the trocar at the proximal end. The cannula shall be 4 inches ± 1/4 inch long. When the trocar is inserted into the cannula, it shall extend at least 9/16 inch past the distal opening of the cannula. When assembled, the trocar point shall be covered by an end protector.

The trocar shall be hollow with an opening near the distal point which shall be clear of the cannula when the trocar is inserted. The distal opening shall have a cross sectional area equal to at least 1/2 of the inside cross sectional area of the trocar. A second opening shall be at the proximal end of the trocar. The trocar and catheter shall be lubricated with pharmaceutical grade silicone lubricant.

The distal end of the 3-way valve shall form a leakproof seal when connected to the catheter. The valve shall operate in the following positions: to close the drainage system, to allow drainage to a collection bag and to shut off flow to the collection bag while allowing irrigation through the injection site. The proximal end of the valve shall have an enlarged section design to connect to tubing.

The molded rubber part of the injection site shall be securely attached to the 3-way stopcock.

The body seal shall be 2-7/8 inches ± 1/8 inches in diameter. The body seal shall have 5 holes, 2 on each side of the catheter retention grooves and one in the center. When attached to the patient, the body seal shall protect the catheter entry site. The body seal shall have two slotted grooves, designed to hold the catheter in place.

Three cubic centimeters of non-irritating, non-toxic, non-sensitizing, pressure sensitive silicone adhesive shall be provided. A brush applicator shall be provided.

Sterility and pyrogenicity. The drainage set shall be sterile and free from pyrogens. The drainage set shall be tested for sterility and pyrogenicity in accordance with the USP.

Clear passage of fluids. Occlude one open end of the stopcock and apply a water pressure of 4 psig to the opposite end. The stopcock shall then be opened. There shall be a clear passage of water within 2 seconds of the opening of the stopcock. Remove the rubber portion of the injection site and repeat the clean passage of fluids test.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Leakage test (injection site). The stopcock shall be closed and an internal water pressure of 4 psig applied. Open the stopcock. The rubber injection site shall reseal without one falling drop after puncture and withdrawal of an 18 gauge unlubricated needle. The resealing shall occur within 10 seconds of the needle withdrawal.

Leakage test (stopcock). Occlude one open end of the stopcock and apply a water pressure of 20 psig for at least 3 minutes. There shall be no leakage exceeding one falling drop from any portion of the stopcock.

Luer fitting. With the molded rubber portion of the injection site removed, the Luer fitting shall meet the dimensional and performance requirements of ANSI/HIMA MD70.1.

Heavy metals. The drainage set shall be tested for heavy metals in accordance with the Physio-Chemical Test of the USP.

Instructions. An instruction sheet detailing the proper use of the drainage set shall be provided with each device. As a minimum, the instruction sheet shall include detailed techniques for correct insertion and operation.

Workmanship. The drainage set shall be free from defects which detract from its appearance or impair its serviceability.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements. The lumens shall be void of obstructions.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Package (PG). One package containing ten drainage sets, sterile, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for visual examination shall be in accordance with MIL-STD-105. The unit of product for examination shall be one drainage set. Sampling shall be Level II with an AQL of 6.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one drainage set. AQL shall be 1.0 percent defective. Inspection level shall be S-2.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-125-3238	DRAINAGE SET, BLADDER, Suprapubic, Reif, 8 Fr

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

A-A-51871

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 47347 Project No. 6515-4053

☆U.S. GOVERNMENT PRINTING OFFICE: 1988 - 505-033/92217

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.