A-A-51872
18 March 1987

COMMERCIAL ITEM DESCRIPTION

SUTURE, ABSORBABLE, SURGICAL, (Polyglycolic Acid)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers absorbable surgical sutures fabricated of polyglycolic acid and suitable for use in various surgical procedures.

Salient characteristics:

Sutures shall be fabricated from braided, undyed filaments of polyglycolic acid. Sutures which are armed shall be swaged to eyeless needles and meet the requirements of the USP for security of needle attachment. Size and tensile strength shall also be in accordance with the USP.

Needles shall be fabricated from type 420 or 440 corrosion-resistant steel in accordance with ASTM F899. Needles shall be properly stress-relieved and passivated. Needles shall meet the requirements of ASTM E290 for ductility.

Sutures and needles shall have dimensions as specified in Table I.

TABLE I. Dimensions

NSN (6515)	Size (USP)	Suture Length (in)	No. of Arms	Needle Circle	Needle Point	Needle Length (in)	Wire Dia. (in)	No. of sut./ packet
00-139-5911	1	36	1	3/8	Blunt	2-1/2	0.032	1
00-139-5916	3-0	27	1	1/2	Taper	1-3/64	0.034	1
00-139-5938	4-0	36	1	1/2	Taper	1-7/16	0.034	1
00-139-5942	4-0	18	0					3
00-139-5943	5-0	27	1	1/2	Rev.Cut	13/16	0.017	1
00-139-5947	2-0	18	0					3
00-139-5948	3-0	36	1	1/2	Taper	1-7/16	0.034	1
00-262-6782	0	18	0					3
00-290-1938	3-0	18	0					3
00-482-2833	4-0	27	1	1/2	Rev.Cut	13/16	0.017	1
01-008-7107	0	27	1	1/2	Taper	1-7/16	0.039	1
01-140-5318	4-0	18	0					12
01-162-1835	3-0	27	1	1/2	Taper	1-7/16	0.034	1
00-551-2347	2-0	27	1	1/2	Taper	1-7/16	0.034	1

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

Sutures shall be sterile in accordance with the USP. Sutures shall have at least 50 months of the original shelf-life remaining at the time of delivery to the Government. Sutures shall be packaged sterile inside an inner container made of aluminum foil and/or plastic. The inner container shall be packaged sterile inside an outer container of the peel-back opening type as shown in Figure 1. The inner container shall show no evidence of leakage when tested as specified in ASTM D3078. Inner and outer containers shall maintain sterility throughout the shelf-life. Each inner container shall have a label identifying the suture material, size, length, number of arms, needle circle, needle point, and the legend "sterile."

Workmanship. Sutures shall be free from frays and other defects which detract from their appearance or impair their serviceability. Needles shall be bright and smooth, and free of feathered edges, cracks, burs, and other defects.

Unit. Package (PG). One package containing 36 individually packaged single-armed sutures, or 108 unarmed sutures packed 3 to a packet, or 432 unarmed sutures packed 12 to a packet, of one size only, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:02Z
Check the source to verify that this is the currentversion before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, inspection level S-1 and acceptance number zero. Sampling for sterility shall be in accordance with the USP.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirments of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the Contracting Officer to determine if the product is acceptable.

The Contracting Officer has the option of accepting or rejecting this product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:02Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-139-5911	SUTURE, ABSORBABLE, SURGICAL, Polyglycolic Acid, Size 1, Single-Armed, 36s
6515-00-139-5916	SUTURE, ABSORBABLE, SURGICAL, Polyglycolic Acid, Size 3-0, Single-Armed, 36s
6515-00-139-5938	SUTURE, ABSORBABLE, SURGICAL, Polyglycolic Acid, Size 4-0, Single-Armed, 36s
6515-00-139-5942	SUTURE, ABSORBABLE, SURGICAL, Polyglycolic Acid, Size 4-0, Unarmed, 36s
6515-00-139-5943	SUTURE, ABSORBABLE, SURGICAL, Polyglycolic Acid, Size 5-0, Single Armed, 36s
6515-00-139-5947	SUTURE, ABSORBABLE, SURGICAL, Polyglycolic Acid, Size 2-0, Unarmed, 36s
6515-00-139-5948	SUTURE, ABSORBABLE, SURGICAL, Polyglycolic Acid, Size 3-0, Single-Armed, 36s

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:02Z
Check the source to verify that this is the current version before use.

NSN	ITEM IDENTIFICATION
6515-00-262-6782	SUTURE, ABSORBABLE, SURGICAL, Polyglycolic Acid, Size 0, Unarmed, 36s
6515-00-290-1938	SUTURE, ABSORBABLE, SURGICAL, Polyglycolic Acid, Size 3-0, Unarmed, 36s
6515-00-482-2833	SUTURE, ABSORBABLE, SURGICAL, Polyglycolic Acid, Size 0, Single Armed, 36s
6515-01-008-7107	SUTURE, ABSORBABLE, SURGICAL, Polyglycolic Acid, Size 0, Single Armed, 36s
6515-01-140-5318	SUTURE, ABSORBABLE, SURGICAL, Polyglycolic Acid, Size 0, Single Armed, 36s
6515-01-162-1835	SUTURE, ABSORBABLE, SURGICAL, Polyglycolic Acid, Size 3-0, Single Armed, 36s
6515-00-551-2347	SUTURE, ABSORBABLE, SURGICAL, Polyglycolic Acid, Size 2-0, Single Armed, 36s

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:

ARMY - MD
NAVY - MS
AIR FORCE - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 05120

Project No. 6515-3466

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.




FIGURE 1

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51872	SUTURE, ABSORBABLE, SURGICAL. (Polyglycolic Acid)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*: _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* - Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* - Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* - Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426

PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.