A-A-51879A
2 February 1988
Superseding
A-A-005179(DM)
24 March 1987

COMMERCIAL ITEM DESCRIPTION

TUBE, CENTRIFUGAL HEMATOLOGY

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a tube kit suitable for use in microhematocrit procedures.

Salient characteristics:

Shall be a venous centrifugal hematology tube to perform 1000 hematocrits, total white blood cell counts, platelet counts and partial differential determinations. Shall be suitable for use with the Clay Adams QBC Centrifugal Hematology System. Shall contain 1000 tubes, floats and tube closures.

Tubes. Tubes shall be clean, transparent and fabricated of glass, suitable for the purpose intended. The length of the tube shall be 2.953 ± .010 in. and have an inside diameter of 0.066 ± .00035 inches. Tubes shall be straight and open at both ends without obstruction.

Tubes shall have a black line, the far end of which is 1.969 in. from one end of the tube. Two additional red lines whose ends are 0.664 and 0.875 in. from the aforementioned black line shall also be included. Tubes shall be internally coated with acridine orange, potassium oxalate, agglutinating agent and stabilizers.

Float. Shall be a solid plastic cylinder having a density of 1.080 g/cm^3 at 25°C. Shall have an outside diameter of 0.06285 ± .00035 inches.

Tube closure. Shall be a cap suitable for fitting over one end of the tube. Shall have a friction fit with the tube and form a leakproof seal to allow proper sampling for centrifugal hematology values.

AMSC N/A FSC 6630

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Workmanship. The tube shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 1000 centrifugal hematology tubes, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with commercial practice. The unit product for test purposes shall be one tube kit.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Aquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

6630-01-233-7593 VENOUS CENTRIFUGAL HEMATOLOGY TUBE

The following is the National Stock Number that subject NSN is intended for use with:

6640-01-234-6793 CENTRIFUGAL HEMATOLOGY ANALYZER

Ordering data. (Expiration dating period required, intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6630-0465

Location: Disk # 6

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.