A-A-51884
23 MARCH 1987

COMMERCIAL ITEM DESCRIPTION

CATHETER, CHOLANGIOGRAM (Plastic)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a cholangiography catheter to facilitate immediate postexploratory cholangiography.

Salient characteristics:

Shall be a sterile, disposable, 4 French, 40 cm ± 2 cm long catheter specifically designed for intra-operative cholangiography. The cholangiography catheter shall be flexible. The cholangiography catheter shall be made of a clear, transparent plastic for visualization of air bubbles. The radiopaque ring shall be 1.5 cm from the tip opening for placement determination of the catheter end near the common duct. The tube shall have a radiopaque stripe for its entire length except a portion of distal end. The cholangiography catheter shall have a beveled distal tip simplifying catheter insertion into the cystic duct.

The distal portion shall be cone-shaped with the flared proximal end of the cone being 0.5 cm from the catheter tip. It shall be designed to be held easily by an external suture and to prevent accidental withdrawal. The cone flange in the cystic duct shall be designed to minimize leakage of contrast media.

The cholangiography catheter shall be facilitated with two port Luer-lock fittings in accordance with ANSI/HIMA MD70.1 to allow dye and irrigation syringes to be attached simultaneously.

The French size shall be clearly marked on the catheter.

The cholangiography catheter shall meet any FDA requirements for safety and toxicity applicable for invasive and implantable devices in effect at the time of solicitation.

Workmanship. The cholangiography catheter shall be free from defects which detract from its appearance or impair its serviceability.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One catheter packaged sterile, uncoiled in a sealed peel-open type container capable of maintaining sterility of contents, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one cholangiography catheter.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Sterility test. The cholangiography catheter shall be tested for sterility in accordance with the method specified in the USP for parenteral devices using inoculated product, or paper spore stripe (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The minimum number of test units, regardless of lot size, shall be 10.

Alternate method. Suppliers conducting sterility tests shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the contracting officer, and approval obtained before employment of the alternate method on a government contract.

Radiopacity test. The catheter shall be in accordance with ASTM F640-79.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:02Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-165-6699	CATHETER, CHOLANGIOGRAM, Plastic, Radiopaque, Disposable, 4 French

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA -OSS
PHS
FDA-MPQAS

PGC 52674

Project No. 6515-3528

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:02Z
Check the source to verify that this is the currentversion before use.