A-A-51886
24 MARCH 1987

COMMERCIAL ITEM DESCRIPTION

INTRAVENOUS INJECTION SET, (With Flexible Drip Chamber, Non-Vented, Adult, Without Needle, Disposable)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a single type of non-vented intravenous injection set used in administering parenteral solutions from glass containers with airways or flexible bags.

Salient characteristics:

Materials. Rubber injection sites shall be of the best quality and suitable in all respects for injection and resealing after needle withdrawal. The tubing shall be clear, plasticized polyvinyl chloride with a Shore Durometer A hardness of 78 ± 5. The drip chamber shall be clear and flexible polyvinyl chloride with a Shore Durometer A hardness of 96 + 5, -10. All plastic and rubber components shall be sterile, pyrogen-free and, if in contact with the parenteral solution, shall be in compliance with USP Class IV plastic requirements. The spike and cap of the drip chamber, the roller clamp, the Y injection site, male Luer taper and end protectors shall be made of suitable rigid plastic.

Construction. The set shall be made without an air vent and include a spike, flexible drip chamber, a Y injection site, flash tube, needle adapter, roller clamp, tubing and end protectors. The overall length shall be 74 inches + 3 inches, - 1 inch. The Y injection site shall be six inches from the flash tube. The tubing shall have an inside diameter of 0.106 ± .005 inch and an outside diameter of 0.150 ± .005 inch.

The spike measured from the tip to the shoulder shall be 0.908 inch ± .03 inch and have a diameter of 0.210 ± .005 inch below the tip and at the level just above the shoulder 0.215 ± .005 inch. The spike shall be made with a sharp point and withstand penetration through rubber or plastic stoppers of parenteral solution containers without damage. The drip chamber shall have a volume of 6 ml + 1 ml. The drip orifice shall be calibrated to 15 ± 3 drops/ml. The adapter for the needle attachment shall be a slip fit Luer taper complying with ANSI/HIMA MD70.1. The complete set shall withstand a 20 psig water pressure for three minutes without demonstrating leakage of more than one falling drop.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

The roller shall not be able to be forced out of the clamp body under a 3.65 kg load applied axially and toward the open end of the clamp. The end protectors shall maintain the sterility of the fluid path and the exterior of the spike and Luer taper. The set shall be contained in a package that will maintain the sterility of the set unit the package is opened.

The Y and flash tube injection sites shall reseal without a falling drop after puncture and withdrawal of an 18-gage unlubricated needle. The resealing shall occur within 10 seconds of needle withdrawal. An internal water pressure of 4 psig shall be used in the conduct of this test.

A. Sterility test. The intravenous injection set shall be tested for sterility in accordance with the method specified in the USP using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it is clearly differentiated from product intended for distribution. The incubation medium for the paper spore strip (carrier) shall be soybean-casein digest. The incubation medium for the inoculated product, or inoculated simulated product shall be fluid thioglycollate or alternative thioglycollate at 30° to 35° C for one half of the specimens and soybean-casein digest at 20° to 25° C for the remaining half of the specimens. The minimum incubation time shall be ten days. Sampling shall be as follows:

Medium	Paper spore strip	Inoc. Prod. or Inoc. Simulated Prod.
Soybean-casein digest	S-3*	S-3*
Fluid thioglycollate or Alternative thioglycollate		S-3*

*Use a minimum of 20 specimens. Acceptance number is zero.

B. Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism
Steam	B. stearothermophilus
Dry Heat	B. stearothermophilus
	B. subtilis
Ethylene oxide	B. subtilis
Gamma radiation	B. pumilus

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

The number of viable spores shall be sufficient to provide assurance that the product has been sterilized.

C. Inoculated carrier. Inoculated carriers (paper spore strips) if used, shall be in accordance with MIL-S-36586A, Sterilization Test Strip Set, Bacterial Spore. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in A. Sterility test and B. Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in B. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP and paragraphs A and B.

D. Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

E. Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP. The inspection level shall be as stated in paragraph A above.

Workmanship. The set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 50 intravenous injection sets, each set individually packaged, sterile or with sterile fluid path in a sealed peel-open container, constitutes one unit of issue.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:02Z
Check the source to verify that this is the currentversion before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one intravenous injection set.

Sampling Criteria for Tests.

	Inspection	AQL
Pressure test for leakage	S-3	0.65
Security of roller in clamp body	S-3	0.65
Resealing of injection sites	S-3	0.65
Sterility	(See sterility clauses)	

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the unit/pound units, a request should be made to the contracting officer to determine if the product is aceptable.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-484-0484	INTRAVENOUS INJECTION SET, Non-Vented, With Flexible Drip Chamber, Disposable, 50s, Sterile, 76 Inches Long

Ordering data (Intermediate/exterior packaging quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION A-A-51886

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 60035 Project No. 6515-3857

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.