A-A-51887
24 MARCH 1987

COMMERCIAL ITEM DESCRIPTION

NEEDLE, HYPODERMIC, (Blood Collecting, Multiple Sample)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a sterile multiple-sample, thin-wall, double-pointed blood collecting needle with shut-off sleeve.

Salient characteristics:

Materials. The cannulae shall be made of Type 304 stainless steel in accordance with ASTM F899, Class 3 requirements. The exterior of the cannula shall be smooth and have on it an inert, non-toxic silicone compound suitable for the intended medical usage. The rubber shut-off sleeve shall be of the best quality and suitable for injection and resealing after needle withdrawal. The shut-off sleeve shall be securely attached to the hub. The hub shall be made of polypropylene or other suitable plastic and be capable of being tightly threaded into a tube holder without chipping.

Construction. Shall be a sterile, hypodermic, multiple-sample blood-collecting needle, packaged in an individual, sealed sterile package which shall maintain the sterility of the contents and be so designed as to permit ready dispensing of the needle by aseptic technique. The needle shall comply with the dimensions shown in Tables I and II below. Note: All dimensions are in inches and tolerances are ± 10% unless otherwise indicated.

Table I. Cannula and Point Dimensions

Gage	OD	ID	Injection Length	Overall Length	Point Length
22	0.028	0.018	1-1/2	2-19/32	0.128 ± .012

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Table II. Thread Dimensions

	Needle Hub	Tube Holder
Major Diameter	0.185	0.193
Minor Diameter	0.139	0.150
Pitch Diameter	0.164	0.175

The needle shall withstand loading without deflecting more than shown in inches in Table III below:

Table III. Stiffness

Gage	Span	Load (kg)	Maximum Deflection
22	0.62	1.42	0.023

The shut-off sleeve shall reseal against a 40 cm $\pm$ 5 cm hydrostatic head of water without more than one drop leakage. The test for effectiveness of the shut-off sleeve shall be accomplished by using a container with a bottom outlet to which is attached a gum rubber tube of the required length. The lower end of the tube shall be plugged. The patient end of the blood collecting needle shall then be inserted into the tubing at the specified distance below the water level in the container. Resealing quality is determined by using the thumb and finger to push the shut-off sleeve against the needle hub and observing its resealing action when it is released. As an alternate, a piece of bond paper can be held taut and pressed over the needle.

The cannula shall be straight and have a bevel length as shown in Table I. The point shall be ground flat or concave and may have a center ridge at the distal end to strengthen the tip. The point shall be ground at the outer edge to enhance the keeness. All edges of the point shall be free from burrs and slivers. The heel of the point shall be dull to prevent coring. The hub shall be securely attached on the cannula. The cannula exterior and interior shall be free from chips or any foreign matter.

The needle shall be free of pyrogens and be non-toxic in accordance with the USP.

All needles shall comply with the sterility requirements shown under Sterility test:

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:02Z
Check the source to verify that this is the currentversion before use.

A. Sterility test. The needles shall be tested for sterility in accordance with the method specified in the USP using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it is clearly differentiated from product intended for distribution. The incubation medium for the paper spore strip (carrier) shall be soybean-casein digest. The incubation medium for the inoculated product, or inoculated simulated product shall be fluid thioglycollate or alternative thioglycollate at 30° to 35° C for one half of the specimens and soybean-casein digest at 20° to 25° C for the remaining half of the specimens. The minimum incubation time shall be ten days. Sampling shall be as follows:

Medium	Paper spore strip	Inoc. Prod. or Inoc. Simulated Prod.
Soybean-casein digest	S-3*	S-3*
Fluid thioglycollate or Alternative thioglycollate		S-3*

*Use a minimum of 20 specimens. Acceptance number is zero.

B. Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism
Steam	B. stearothermophilus
Dry Heat	B. stearothermophilus B. subtilis
Ethylene oxide	B. subtilis
Gamma radiation	B. pumilus

The number of viable spores shall be sufficient to provide assurance that the product has been sterilized.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the current version before use.

C. Inoculated carrier. Inoculated carriers (paper spore strips) if used, shall be in accordance with MIL-S-36586A, Sterilization Test Strip Set, Bacterial Spore. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in A. Sterility test and B. Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in B. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP and paragraphs A and B.

D. Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

E. Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP. The inspection level shall be as stated in paragraph A above.

The needles shall be examined for dimensions, freedom from foreign matter and point sharpness at the following MIL-STD-105 inspection level:

Inspection level	AQL
S-3	1.0

Testing of the needles shall be at the inspection levels shown in Table IV.

Table IV. Sampling criteria for tests

Characteristic	Inspection level	AQL
Shut-off sleeve	S-2	*
Freedom from pyrogens	S-1	*
Toxicity	S-1	*
Sterility	(See sterility clauses above)	*
Stiffness	S-1	*

* Acceptance number shall be zero for all sample sizes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Workmanship. The blood collecting needle shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing one hundred (100) blood collecting needles, each needle individually packaged sterile as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one needle.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:02Z
Check the source to verify that this is the currentversion before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-153-5598	NEEDLE, HYPODERMIC, BLOOD COLLECTING, Disposable, Multi Sample, 22 Gage by 1-1/2 Inches, 100s. Sterile

Ordering data (Intermediate/exterior packaging quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 51277 Project No. 6515-3962

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.