A-A-51897
24 April 1987

COMMERCIAL ITEM DESCRIPTION

TRAY, INSTRUMENT

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a corrosion-resisting metal instrument tray with cover suitable for protecting microsurgical instruments during sterilization and transportation.

Salient characteristics:

The instrument tray shall be suitable for use in storing and sterilizing instruments used in opthalmic microsurgery. The tray shall be furnished in the following three styles and appropriate sizes:

Style 1 - 8 inches by 13 inches by 1.75 inches (20.3 by 33.0 by 4.5 cm) with a handle on each of three sides.

Style 2 - 8-1/4 inches by 8-1/4 inches by 1.75 inches (21.0 by 21.0 by 4.5 cm) with a 2.5 inches by 2.5 inches (6.4 by 6.4 cm) compartment for extremely small instruments, and at least one carrying handle.

Style 3 - 8-1/4 inches by 8-1/4 inches by 1.75 inches (21.0 by 21.0 by 4.5 cm) with at least one carrying handle.

The tray shall be constructed of American Iron and Steel Institute (AISI) 300 series corrosion-resisting steel, properly hardened and passivated, or of sturdy, long-wearing anodized aluminum.

The tray shall have multiple small perforations, or slots, on all surfaces to allow steam or ethylene oxide penetration and circulation during sterilization. Steam condensation shall be minimal.

The tray shall have a hinged type cover with a securing mechanism for locking. The cover shall rest evenly on the body of the tray. The tray shall rest evenly on a flat surface without rock or wobble.

DISTRIBUTION STATEMENT A. Approved for public release distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

The tray shall be furnished with a method of stabilizing and securing the microsurgical instruments, without excess rattle or vibration, to prevent damage to the instruments during sterilization or transportation.

If made of corrosion-resisting steel, the tray shall have a no. 4 commercial finish or better.

The tray shall be free of rough grind marks, trimming marks, pits, scales, and other defects or imperfections. The tray shall have rounded corners and shall have surfaces smooth and free of burrs or sharp edges.

Welding or soldering shall be durable, sound in every detail, free of cracks, porosity, and other defects.

The finished tray shall be thoroughly clean and free of dirt, grease, graphite, metal particles, and other foreign material.

Workmanship. The instrument tray shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA) One instrument tray, of size specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contact or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

First article inspection. First article inspection shall be performed on the tray when a first article sample is required. This inspection shall include tray set the examination and tests.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one instrument tray.

TABLE I. Sampling for examination.

	Inspection Level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total (major and minor defects combined)	II	4.0
For dimensional examination	S-1	2.5

Examination. The trays shall be examined to determine compliance with all requirements contained in this specification.

Classification of defects. Examination shall be conducted in accordance with the classification of defects listed on Table II. Examination shall not be restricted to the classified possible defects listed below.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

TABLE II. Classification of defects.

Categories	Defects
Major	
101	Tray not free of cracks, tears, holes, ripples.
102	Edges not free of burrs or sharpness.
103	Tray rocks, wobbles, or does not rest evenly on a flat surface.
104	Cover does not rest evenly on the tray body.
105	Tray not fabricated from a corrosion-resisting metal.
106	Welding or soldering not durable.
Minor	
201	Finish not free of scratches or marks.
202	Tray not free of grease, oil, dirt, chips, or other foreign matter.
203	Tray does not have a smooth and finish.

Dimensional examination. The trays shall be examined for defects in dimensions. Any dimension not within tolerances specified herein shall be classified as a defect.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered materials in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Numbers are covered by this document:

6530-01-169-0298 (Style 1)
6530-01-169-0299 (Style 2)
6530-01-169-0300 (Style 3)

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent.

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES.

VA-OSS
PHS
FDA-MPQAS

PGC 51099
51101
51113

Project No. 6530-1860

Location CID/T-INSTRUMENT/H13-5

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (*DO NOT STAPLE*), and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

Fold along this line

Fold along this line

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4956 WASHINGTON D.C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-51897	Tray, Instrument

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION (Mark one)
	☐ VENDOR
	☐ USER
b ADDRESS (Street, City, State, ZIP Code)	☐ MANUFACTURER
	☐ OTHER (Specify)

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER (Last, First, MI) – Optional	b WORK TELEPHONE NUMBER (Include Area Code) – Optional
c MAILING ADDRESS (Street, City, State, ZIP Code) – Optional	8 DATE OF SUBMISSION (YYMMDD)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.