MIL-C-36036B
20 December 1978
SUPERSEDING
MIL C-36036A (DSA-DM)
12 September 1975
MIL-C-36036
17 June 1960

MILITARY SPECIFICATION

CASE, MEDICAL INSTRUMENT AND SUPPLY SET, NYLON, NONRIGID, NO. 7

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers an empty case fabricated of nylon cloth and fitted with a slide fastener. Each case is equipped with a detachable carrying strap.

2. APPLICABLE DOCUMENTS.

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

V-F-106	- Fastener, Slide, Interlocking.
V-T-285	- Thread, Polyester.
PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-B-676	- Boxes, Set-Up, Paperboard.

MILITARY

MIL-P-116	- Preservation Packaging, Methods Of.
MIL-F-495	- Finish, Chemical, Black, For Copper Alloys.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, CODE ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6545

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:59Z
Check the source to verify that this is the current version before use.

MIL-W-530	- Webbing, Textile, Cotton, General Purpose, Natural Or In Colors.
MIL-R-3390	- Ring, Dee.
MIL-L-3851	- Loop, Slide (For Equipage).
MIL-W-4088	- Webbing, Textile, Woven Nylon.
MIL-L-11075	- Loop, Strap Fastener.
MIL-C-43128	- Cloth, Plain Weave, Nylon: Water Repellent, OG-106.
MIL-S-43770/1	- Snap Hook, Wire Body, Fixed Loop Eye, Flat Spring Closure, with Retainer.

STANDARDS

FEDERAL

Fed. Test Method Std. No. 151	- Metal, Test Methods.
FED-STD-595	- Color.
FED-STD-751	- Stitch, Seam, And Stitching.

MILITARY

MIL-STD-105	- Sampling Procedures And Tables For Inspection by Attributes.
MIL-STD-129	- Marking For Shipment And Storage.

DRAWINGS

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

1327	- Sling, Bag and Case, Carrying, Medical Field Kit Case.
1901	- Case, Medical Instrument and Supply Set, Nylon, Nonrigid No. 7.

(Copies of specifications, standards, and drawings required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer).

3. REQUIREMENTS

3.1 Materials. Except as specified herein, materials shall conform to the applicable portions of the specifications, standards and drawings specified herein, and shall be capable of meeting the requirements specified therein. In addition, the metals shall conform to the chemical requirements of their applicable specifications when tested as specified in Section 4. When materials to be used are not specifically designated, those used shall be of the highest commercial quality, and entirely suitable for the purpose intended.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:59Z
Check the source to verify that this is the current version before use.

3.1.1 Cloth, nylon. The nylon cloth shall be Olive Green No. 106, water repellent, conforming to MIL-C-43128.

3.1.2 Webbing.

3.1.2.1 Webbing, nylon. The 3/4 inch nylon webbing shall conform to type Ia of MIL-W-4088. The 1 inch nylon webbing shall conform to type II of MIL-W-4088. Color shall be Olive Drab No. 7.

3.1.2.2 Webbing, cotton. The 2 inch wide cotton webbing for the carrying strap shall conform to the requirements of MIL-W-530, type II, class 4, and shall be vat dyed Olive Drab No. 7.

3.1.3 Thread, polyester. The thread for all stitching, except overedge stitching, shall be type I, class 1, subclass B, size F, of V-T-285. Size B thread shall be used on overedge stitching. Thread shall be dyed Olive Drab Shade S-1, C.I. 66022, except that the colorfastness requirements and test methods shall be limited to weathering only.

3.1.4 "D" rings. The "D" (DEE) rings shall be style A, class 1, 25/32 inch by 3/8 inch, black chemical finish, of MIL-R-3390.

3.1.5 Fastener, slide. Except as specified herein, the slide fastener shall conform to V-F-106, type I, style 6, size M, single slider, with short tab pull fitted with a webbing thong. The metal parts shall be brass with a black chemical finish conforming to MIL-F-495.

3.1.5.1 Tape. The slide fastener tapes shall be vat dyed olive drab shade S, conforming to U.S. Army Color Card of Standard Shades for the Slide Fastener Tapes, and shall be water repellent and mildew resistant, in accordance with V-F-106. Mildew resistance shall be obtained by using copper-8-quinolinolate only.

3.1.5.1.1 Water repellence. The finished tape shall be tested for water resistance as specified in V-F-106.

3.1.5.1.2 Mildew resistance. The finished tape shall be tested for mildew resistance as specified in V-F-106.

3.1.5.2 Salt spray resistance. The slide fasteners shall meet the salt spray resistance requirements specified in V-F-106.

3.1.5 3 Webbing for thongs. Webbing for thongs shall conform to the requirements of V-F-106 for thongs (cotton webbing) except that mildew-resistance shall be obtained by using copper-8-quinolinolate only, and the finished length of the thong after assembly to the pull shall be 3-1/2 ± 1/2 inch

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:59Z
Check the source to verify that this is the current version before use.

3.1.6 Loops, slide. Slide loops shall conform to MIL-L-3851 and shall be class 1, style 2, 2 inches.

3.1.7 Loops, strap. Strap loops shall conform to MIL-L-11075 and shall be type 1, brass, style E, 2-1/16 by 1/4 inch, wire.

3.1.8 Snaphooks. Snaphooks shall conform to the requirements of MIL-S-43770/1, Material 1, Part No. M43770/1-CWBC3.

3.1.9 Stitching sealant. Exterior stitching shall be sealed by the application of one of the two sealants listed in 6.3, or equal, in accordance with the manufacturer's directions.

3.2 Component parts. This case shall consist of two major parts hinged together along one side, a slide fastener, two "D" rings, and a detachable carrying strap (sling). The sling shall be fitted with two snaphooks, a strap loop, and a slide loop.

3.3 Construction. The construction of the case (including detachable carrying strap) shall conform to the requirements specified herein and to those shown on DPSC Drawing 1901, and for the carrying strap, DPSC Drawing 1327

3.3.1 Stitching, machine. All stitching shall conform to FED-STD-751 as follows

For all stitching except overedge stitching; type 301, 8 to 10 stitches per inch.

For overedge stitching (single operation optional); type 503 or 504, 8 to 10 stitches per inch.

For joining seams and overedge stitching (simultaneous operation optional), type 301 in combination with 503, 504 or 502, 8 to 10 stitches per inch.

All edges of joining seams shall be stitched together. There shall be no lubrication of the thread by any means, prior to or during sewing. Any of the operations requiring box, box with cross stitch, W stitching, straight line tacking or any other design of stitching cited in this specification may be performed with an automatic stitching machine provided the requirements for stitch pattern, stitches per inch, size, and types of thread are met, and at least three or more tying, overlapping, or back stitches are used to secure the end of stitching.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:59Z
Check the source to verify that this is the current version before use.

3.3.1.1 Type 301 stitching. Thread breaks in stitching shall be overstitched not less than 1 inch at each break. Stitching shall be backstitched not less than 1 inch at the ends, except where ends are turned under in a hem or held down by other stitching. Thread tension shall be maintained so that there will be no loose stitching, and the lock will be imbedded in the materials sewed.

3.3.1.2 Overedge stitching. Thread breaks in overedge stitching shall be overstitched not less than 3/4 inch at each break. Thread tension shall be maintained so that there will be no loose stitching.

3.3.2 Fusing of nylon webbing. All ends of nylon webbing shall be fused after cutting.

3.4 Marking for identification. Except for the Geneva Red Cross on a white background, each case (including carrying strap), shall be legibly and permanently marked with black waterproof ink or paint in accordance with the drawings. The Geneva Red Cross and white background shall be applied to the case as shown on Drawing 1901. The Cross and white background shall conform to the following colors of Federal Standard No. 595: Red Cross-11105, white background-17886.

* 3.5 Workmanship. Workmanship shall be first class throughout. Complete cases (including carrying straps) shall be well made and free from defects which detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

* 4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:59Z
Check the source to verify that this is the current version before use.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the nylon cloth, webbing, thread, hardware, and fasteners, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contract, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and test required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in table I and II (see 4.2.1, 4.2.2 and 4.2.3).

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. Unit of product for examination purposes shall be one case.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per 100 units)
For visual examination		
Major	II	2.5
Total (Major and Minor combined)	II	6.5

4.2.2 Examination.

4.2.2.1 Classification of defects. The case shall be examined for defects including, but not limited to those listed in table II classification of defects.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:59Z
Check the source to verify that this is the current version before use.

TABLE II. Classification of defects.

Categories	Defects
Major	
101	Any component missing.
102	Any required operation omitted or improperly performed.
103	Any hole, cut, tear, mend, patch or unauthorized splice.
104	Any needle chew.
105	Any weakening defects such as smash, multiple float, or loose slub that may develop into a hole.
106	One or more spots or stains of any obviously permanent nature.
107	Any seam twisted or puckered as to impair appearance or serviceability.
108	Seam type other than specified.
109	Stitch type other than specified.
110	Less than 8 stitches per inch.
111	One or more required rows of stitching omitted.
112	Any open seam 1/4 inch or more.
113	Any raw edge.
114	Any webbing with frayed or scalloped edge.
115	Any webbing not tightly and firmly woven.
116	Exposed end of webbing not fused.
117	Any hardware broken or malformed.
118	Any burr or sharp edge on hardware.
119	Finish on hardware omitted.
120	Any marking omitted, incorrect, or illegible.
121	Any marking not size specified.
122	Not slide fastener specified.
123	Slide fastener does not operate smoothly.
Minor	
201	One or more removable noticeable spots or stains.
202	Thread ends not trimmed.
203	Any part of case pleated, twisted, or caught in any unrelated row of stitching.
204	Any broken or skipped stitch, not amounting to an open seam, occurring twice or more in the same row of stitching or an open seam less than 1/4 inch.
205	Any end of stitching backstitched less than specified.
206	Finish on hardware not uniform.
207	Colors not as specified.
208	Any markings not located as specified.

4.2.2.2 Dimensional defects. During this examination, any measurement deviating from the normal dimensions and tolerances of the finished case specified in the drawing shall be scored as a size measurement defect. The sample unit for this examination shall be one completely fabricated case. The lot size shall be expressed in units of one case. The sample size shall be selected in accordance with inspection level S-3 of MIL-STD-105. The acceptable quality level shall be 4.0 (defects per 100 units).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:59Z
Check the source to verify that this is the current version before use.

* 4.2.2.3 Packaging inspection. The inspection of the unit package shall be in accordance with the group A quality conformance inspection requirement of MIL-P-116. The inspection of the intermediate package, packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.2.3 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, inspection level S-1, except that the acceptance number shall be zero for all sample sizes. Unit of product for test purposes shall be one case.

4.2.3.1 Sampling for chemical testing of metals. A sufficient amount of each lot of each metal used in the construction of the cases shall be selected for the chemical analysis as specified in 4.3.1.

4.3 Tests. Tests shall be conducted to determine compliance with the requirements of this specification.

* 4.3.1 Chemical analysis of metals. The samples selected in accordance with 4.2.3.1 shall be analyzed in conformance with method 111 or 112 of Federal Test Method Standard No. 151. In case of dispute, method 111 shall be used for referee analysis.

* 5. PACKAGING

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit package. Each case shall be neatly folded and packaged in accordance with MIL-P-116, method III.

5.1.1.2 Intermediate package. Twelve unit packages shall be packaged in a snug fitting box constructed in accordance with PPP-B-566, PPP-B-676, or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.2).

5.2.1 Level A. Twelve intermediate packages (144 units) shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, class water-resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

5.2.2 Level B. Twelve intermediate packages (144 units) shall be packed in an exterior container designed for a type 2 load and constructed in accord-

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:59Z
Check the source to verify that this is the current version before use.

ance with PPP-B-636, class domestic. Closure shall be as specified in the appendix of PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose so constructed as to insure acceptance and safe delivery by common or other carriers to point of delivery called for in the contract or purchase order.

5.3 Marking. Each unit, intermediate package and exterior container shall be marked as specified in MIL-STD-129.

6. NOTES

6.1 Intended use. The case covered by this specification is intended for use as a container for NSN 6545-00-927-3000 Medical Instrument and Supply Set, Aeronautic Emergency.

6.2 Ordering data. Procurement documents should specify the following.

(a) Title, number and date of this specification.
(b) National Stock Number (NSN).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 Stitching sealants. The two sealants referenced in 3.1.9 are:

Bostik No. 7058	Tite Seal
Bostik Chemical Group, USM Corp.	Eureka Tent Corp.
Middleton, MA 01949	Binghamton, NY 13902

6.4 This specification covers the following item listed in the Federal Supply Catalog.

National Stock Number	Item Identification
6545-00-912-9830	CASE, MEDICAL INSTRUMENT AND SUPPLY SET, Nylon, Nonrigid, No. 7, 13 inches long by 5-3/4 inches wide by 4 inches high.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:59Z
Check the source to verify that this is the current version before use.

* 6.5 Changes from previous issue. The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DoD - MB

Preparing activity:
Defense Logistics Agency - DM

Project Number 6545-0042

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:59Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS This form is provided to solicit beneficial comments which may improve this document and enhance its use DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government Fold on lines on reverse side, staple in corner, and send to preparing activity Attach any pertinent data which may be of use in improving this document If there are additional papers, attach to form and place both in an envelope addressed to preparing activity A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed

NOTE This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s)-or to amend contractual requirements

DOCUMENT IDENTIFIER (Number) AND TITLE

MIL-C-36036B CASE, MEDICAL INSTRUMENT AND SUPPLY SET, NYLON NONRIGID, NO. 7

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1 ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID RESTRICTIVE LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW

A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDED WORDING CHANGE

C REASON FOR RECOMMENDED CHANGE(S)

2 REMARKS

SUBMITTED BY (Printed or typed name and address - Optional)

TELEPHONE NO

DATE

DD FORM 1 OCT 76 1426 EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:59Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

POSTAGE AND FEES PAID

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS
DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE: ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:59Z
Check the source to verify that this is the current version before use.