MIL-F-0036043A(BuMed)
17 December 1962
USED IN LIEU OF
MIL-F-36043
31 March 1961

MILITARY SPECIFICATION

FIRST AID KIT, EYE DRESSING

This limited coordination Military specification has been prepared by the Defense Medical Supply Center based upon currently available technical information, but it has not been approved for promulgation as a coordinated revision of Military Specification MIL-F-36043. It is subject to modification. However, pending its promulgation as a coordinated Military specification, it may be used in procurement.

1. SCOPE

1.1 This specification covers a kit suitable for first aid treatment of eye injuries.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids, or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

U-P-401	Plaster, Adhesive, Surgical.
TT-E-527	Enamel, Alkyd, Lustreless.
CCC-G-101	Gauze, Absorbent.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up, Paperboard.

FSC 6545

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

Military

MIL-L-10547 Liners, Case, Waterproof.
MIL-B-16152 Box, Folding (First Aid Packages).

STANDARDS

Federal

Fed. Std. No. 595 Colors.

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publication. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

U.S. PHARMACOPOEIAL CONVENTION, INC.

Pharmacopeia of the United States.

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pa.)

3. REQUIREMENTS

3.1 Component parts. The kit consists of two camouflaged eye pads, 8 camouflaged adhesive plaster strips (2 packets) and 2 tubes of Oxytetracycline-Polymyxin B Ophthalmic Ointment, 1/8 oz, arranged in a size 1 container bearing directions for use as specified in 3.5.

3.2 Materials.

3.2.1 Eye pad. Eye pad shall be 2 inches in diameter and shall be composed of a filler of a suitable quantity (approximately 0.8 gram) of U.S.P. grade absorbent cotton, covered top and bottom with single thicknesses of type I, 44-36 gauze, camouflaged, in accordance with the requirements of Specification CCC-G-101. The eye pad shall be free from loose threads and ravelings.

3.2.2 Adhesive plaster. Except as specified herein, adhesive plaster shall conform to type I of Specification U-P-401. The facing shall be crinoline as specified in 3.1.4, and the color shall be camouflaged (field brown) as specified in 3.4.2 of the same specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

3.2.3 Oxytetracycline-Polymyxin B Ophthalmic Ointment. Oxytetracycline-Polymyxin B Ophthalmic Ointment, shall conform to the requirements of the Military Medical Purchase Description, of the issue in effect on date of invitation for bids, for Stock No. 6505-299-8608, except that packaging shall be as specified in 3.3.3.1, and the marking shall be as specified in 3.4.3.

3.2.4 Container. Except as specified herein, container shall conform to style 1, size 1, of Specification MIL-B-16152.

3.2.5 Enamel. Enamel shall conform to the requirements of Specification TT-E-527. The color of the enamel for coating the external surfaces of the bodies of the tubes shall be olive drab, color 34087, as specified in Federal Standard No. 595 and the color of the enamel for the marking of the tubes, as specified in 3.4.3, shall be black, color 37038, of the same standard (see 6.3).

3.3 Kit contents. There shall be placed within the container (3.2.4), two eye pads (3.2.1), two adhesive plaster strip packets (3.3.2), and two tubes of Oxytetracycline-Polymyxin B Ophthalmic Ointment (3.2.3), arranged as specified in 3.5.

3.3.1 Eye pad. Each eye pad shall be individually packaged in an envelope fabricated of natural kraft-colored glassine or parchment paper, sealed and sterilized in accordance with the requirements of the Pharmacopeia of the United States (U.S.P.). The envelope shall be so constructed and sealed that sterility of the content is maintained until the envelope is opened.

3.3.1.1 Sterilization and sterility certificates. Compliance with 3.3.1.1.1 or 3.3.1.1.2 is required.

3.3.1.1.1 Certificate for steam (water vapor) sterilization. Actual accomplishment of sterilization shall be certified in writing for each sterilizer charge by sworn affidavit as having been accomplished by the manufacturer. Such certificate shall list the lot or control number for each sterilizer charge and shall state that the charge was subjected to saturated steam, at a minimum of 250° Fahrenheit (F.) for not less than 20 minutes after sterilizing conditions have been attained (when gauge pressure corresponds to that of saturated steam at the indicated temperature). Gauge pressure readings shall also be reported. Said readings shall correspond to the required pressure for saturated steam and temperature in the autoclave. "Sterilization detectors", of the type indicating duration of time to steam and temperature, shall be used to indicate that the desired sterilization conditions have been attained. Detectors shall be placed in the center, as well as distributed longitudinally, vertically, and transversely throughout the sterilizer load. The gauge pressure readings, and the sterilization detectors shall be maintained as a permanent record by the contractor and shall be made available upon request.

Note. Temperatures as low as 240° F. may be used, provided it is demonstrated that effective sterilization is obtained.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

3.3.1.1.2 Certificate of sterility for steam (water vapor) or other methods of sterilization. A certificate of sterility, listing all lot or control numbers, stating that samples from each sterilizer charge have been tested and found to be sterile, shall be submitted to the procuring activity. Sterility tests shall be conducted in accordance with U.S.P. requirements. Chemical sterilization is not permissible.

3.3.2 Adhesive plaster strip packets. Two strips of adhesive plaster (3.2.2), each not less than 3/8 inch wide, not more than 1/2 inch wide, and not less than 5 inches long, shall be faced in parallel by a single piece of crinoline (3.2.2) of size appropriate to protect the adhesive. Two assembled adhesive plaster strips shall be enclosed in a glassine paper envelope and sealed to form the adhesive plaster strip packet.

3.3.3 Ophthalmic ointment. Each tube of ophthalmic ointment shall be sealed in such a manner as to permit the unopened tube to be stored at a temperature of 140° F. for 8 hours without any evidence of leakage of the contents. Exterior metal surfaces shall be finished with one coat of enamel as specified in 3.2.5.

3.3.3.1 Cartoning of ophthalmic tubes. Each tube of ophthalmic ointment shall be individually packaged in a folding carton. Outside measurements shall not exceed 7/8 by 9/16 inch by 3 inches long.

3.4 Marking for identification.

3.4.1 Eye pad. Each envelope containing the eye pad shall bear the sterility clause and name or registered trade-mark of manufacturer. When the contractor is not the manufacturer, the contractor's name shall also appear.

3.4.2 Adhesive plaster strip packets. Each glassine envelope shall be marked "Adhesive Strips" and shall bear the name or registered trade-mark of the manufacturer.

3.4.3 Ophthalmic ointment tubes and individual cartons. Each ophthalmic ointment tube shall be marked with enamel as specified in 3.2.5, and each ointment carton shall be permanently and legibly marked in black ink as specified in Specification MIL-B-16152. Tube and individual carton shall show the stock number, item identification of the item (including name and quantity of contents), the name or registered trade-mark of the manufacturer; when the contractor is not the manufacturer, the name and address of the contractor shall also appear, the lot or control number ("Lot # _____" or "Control # _____"), the expiration date, and the following: "EYES - In the case of a foreign body which cannot be easily removed, squeeze a small amount of ophthalmic ointment on inner surface of lower lid and apply eye dressing. Seek medical aid at once.

CAUTION: Do not rub eye."

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

The quantity of contents on the tube and individual carton shall be designated as, "1/8 oz (3.5 Gm)". The lot or control number and the expiration date on the immediate container may be stamped on the crimp of the tube, in which case the label of the immediate container shall include the statement: "See crimp closure for lot or control number and expiration date". The lot number and expiration date when stamped in the crimp need not be preceded by the required prefixes.

3.4.4 Size 1 carton.

3.4.4.1 Front face. The front face of the container shall show the following:

FIRST AID KIT, EYE DRESSING

Stock No. 6545-853-6309

2 Eye Pads, STERILE.
2 Packets Adhesive Strips.
2 Tubes Oxytetracycline-Polymyxin B.
Ophthalmic Ointment FSN 6505-299-8608

Name of manufacturer
Name of contractor (if other than manufacturer)

INSTRUCTIONS d L'INTERIEUR DE LA TROUSSE

3.4.4.2 Back face. The back face of the container shall show the instructions for use, including illustrations as shown in figure 1.

3.4.4.3 Top edge and both ends. Top edge and both ends shall show stock number and item identification.

3.4.4.4 Bottom. The bottom shall show the legend:

"Open at Opposite End".

3.5 Arrangement of components. The specified components shall be arranged within the size 1 carton, as follows: The two cartons containing the ophthalmic ointments shall be placed against one end and front face of carton. The tube cartons shall be stacked on the 9/16 inch face two cartons high. One eye pad shall be placed between the back face of the size 1 carton and sides of the tube carton. The second eye pad shall be placed adjacent to, and slightly overlapping, the first pad. The instruction sheet (3.6) shall be placed between the cartons containing the ophthalmic ointment and the side of the size 1 carton. The packets containing the adhesive strips shall be neatly folded and placed at the void end of the carton.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

3.6 Instructions in French. Each unit shall contain illustrations shown on figure 1 and instructions printed in French, on a card of suitable quality and size, containing the information below, except, that in the illustrations, the words "BLESSURES OCULAIRES" shall be shown in C instead of "EYE INJURIES":

LISTE D'ELEMENTS		QUANTITE
No. d'Assortiment Fédéral		
Item Identification		
Bande de Plâtre Adhésive, Camouflé		8
Tampon Oculaire, Camouflé		2
Oxytetracycline-Polymixun - B Pommade Ophthalmique, tube de 1/8 once	6505-299-8608	2

A. Tirer en bas la paupiere inferieure et appliquer la pommade.

B. Apres avoir mis la pommade, couvrir avec le tampon oculaire et mettre en place à l'aide de bandes adhésives.

C. Blessures oculaires: En cas d'un corps étranger qui ne peut être enlevé facilement, exprimer un peu de pommade ophthalmique sur l'intérieur de la paupière inférieure et mettre le bandage ophthalmique. Aller voir un médecin immédiatement.

ATTENTION: Ne pas frotter l'oeil.

3.7 Workmanship. Workmanship shall be first class throughout. The individual components, as well as the completely assembled First Aid Kit, Eye Dressing, shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with Standard MIL-STD-105.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with Standard MIL-STD-105.

Component (unit of product)	Characteristic	Rqmt.	Test procedure	Sample unit	Insp. level	AQL (percent defective)
Eye pad Ophthalmic ointment	Sterility	3.3.1	See 3.3.1.1	Each	- - -	*
	Leakage of tube	3.3.3	4.4.1	Tube	L7	1.0

* Acceptance number shall be zero (that is, none of the "sterilization detectors" used in 3.3.1.1.1 shall show that sterilization conditions have not been attained. None of the samples in 3.3.1.1.2 for each sterilization lot shall fail).

4.2.3 Component materials. Sufficient samples of the components shall be selected from each lot of material used in the manufacture of the First Aid Kit, Eye Dressing, to perform tests as required by the applicable subsidiary specifications or by tests specified herein. When sample size for component materials covered by the subsidiary specifications is not specified, the sample size for each test shall be as indicated in 4.2.2. The unit of product for testing shall be the quantity of material necessary to perform all required tests one time each, taking into consideration the fact that, where possible, the same material may be used to perform two or more tests.

4.2.4 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

4.3 Examination. The individual components, as well as the completely assembled First Aid Kit, Eye Dressing, shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Leakage test. Each filled, unopened tube of each ophthalmic ointment shall be stored for 8 hours at a temperature of 140° F. No leakage of contents shall be evidenced during the test, nor at completion of test.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) first aid kit, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I. Packaging quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	576 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Unit package shall be specified in 3.2.4.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with Specification PPP-B-566 or PPP-B-676, except that commercial colors shall be acceptable. Each carton shall be adequately secured to prevent accidental opening.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose within the continental limits of the United States of America.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with Specification PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; or PPP-B-621, table III.

5.2.1.2 Case liner. Each level A box shall be lined with a waterproof case liner conforming to Specification MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

5.2.1.3 Strapping. Strapping, when required, shall be in accordance with applicable box specification.

Note. Strapping will not be required for shipment forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be a box of appropriate size constructed in accordance with Specification PPP-B-585, style 3, designed for class 2 use and type 2 load; PPP-B-621, table I; or PPP-B-636, type I, class 1.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Unit package. Each unit package shall be marked as specified in 3.4.4.1 through 3.4.4.4, inclusive.

5.3.2 Intermediate package. Each intermediate package shall be marked in accordance with Standard MIL-STD-129. Contract number and name of contractor shall be shown.

5.3.3 Exterior container. Exterior container shall be marked in accordance with Standard MIL-STD-129. In addition, the letter "L" shall precede the Federal stock number. One space shall be allowed between the letter "L" and the first digit of the stock number. The terms "Stock number" or "FSN" shall not be indicated. Service color and service symbol markings are required only on direct shipments from the contractor's plant to overseas activities.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Copies of Military Medical Purchase Description for stock number enumerated in 3.2.3 may be obtained from the procuring activity or as directed by the contracting officer.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 Availability of color chips. The individual 3-by 5-inch chips of the colors referenced in 3.2.5 are available from General Services Administration, Business Service Center, Region 3, Seventh and D Streets, S. W., Washington 25, D.C. Price 5 cents.

6.4 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

6.5 Transportation description. Transportation description applicable to this item is:

Kits, first aid

Carload minimum weight 24,000 pounds subject to Rule 34 of the Uniform Freight Classification.


Truckload minimum weight 24,000 pounds per National Motor Freight Classification.

Notice. When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

Custodians:
Army-Med
Navy-BuMed
Air Force-AFCSG

Preparing activity:
Defense Medical Supply Center

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.



Copy A. - Pull down lower eyelid and apply ointment.

Copy B. - After applying ointment, cover with eye pad and secure with adhesive strips.

Copy C. - Eye Injuries:

In the case of a foreign body which cannot be easily removed, squeeze a small amount of Ophthalmic Ointment on inner surface of lower lid and apply eye dressing. Seek medical aid at once.

CAUTION: Do not rub eye.

Figure 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

FOLD

DEPARTMENT OF THE NAVY

POSTAGE AND FEES PAID
NAVY DEPARTMENT

OFFICIAL BUSINESS

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity *(as indicated on reverse hereof)*.

SPECIFICATION

ORGANIZATION *(Of submitter)* | CITY AND STATE

CONTRACT NO. | QUANTITY OF ITEMS PROCURED | DOLLAR AMOUNT $

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES.

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES", IN WHAT WAY?

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity)* | DATE

DD FORM 1426, APR 63 REPLACES NAVSHIPS FORM 4863, WHICH IS OBSOLETE. S/N 0102-LF-001-4260 PLATE NO. 15413

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.