INCH-POUND

MIL-F-36043D
22 June 1992
SUPERSEDING
MIL-F-36043C
12 December 1978

MILITARY SPECIFICATION

FIRST AID KIT, EYE DRESSING

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a first aid kit suitable for treatment of eye injuries.

2 APPLICABLE DOCUMENTS

2.1 Government Documents.

2.1.1 Specifications, standards, and handbooks. The following specifications, standards, and handbooks form a part of this document to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DODISS) and supplement thereto, cited in the solicitation (see 6.2).

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-MST, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

AMSC N/A FSC 6545

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

FEDERAL

L-P-378	- Plastic Film, (Polyethylene Thin Gage)
CCC-G-101	- Gauze, Absorbent.
PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-585	- Boxes, Wood, Wirebound.
PPP-B-601	- Boxes, Wood, Cleated-Plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	- Boxes, Fiberboard.
PPP-B-676	- Boxes, Setup.

MILITARY

MIL-L-10547	- Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof Flexible.
MIL-S-36586	- Sterilization Test Strip Set, Bacterial Spore.

STANDARDS

FEDERAL

FED-STD-595	- Colors.

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.
MIL-STD-794	- Procedures for Packaging Parts and Equipment.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

2.1.2 Other Government documents, drawings, and publications. The following other Government documents, drawings, and publications form a part of this document to the extent specified herein. Unless otherwise specified, the issues are those cited in the solicitation.

DRAWING

Defense Personnel Support Center (DPSC)

24211 - Bag, First Aid Kit, Eye Dressing

(Copies of specifications, standards and drawing required by contractor in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Order of precedence. In the event of a conflict between the text of this document and the references cited herein (except for related associated detail specifications, specification sheets, or MS standards), the text of this document takes precedence. Nothing in this document, however, supersedes applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 End Item. The first aid kit shall be an assembly of components to facilitate four complete eye dressings with instructions contained in a sealed reclosable polyethylene bag.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

3.1.1 Component parts. Table I details the components required for each eye dressing first aid kit.

TABLE 1. First aid kit contents.

End Item	Component	Required Number
Eye Dressing, First Aid Kit	Package containing: 2 Individually wrapped eye pads (sterile) 4 Strips of adhesive tape 8 inches long by one inch wide.	2
	Ophthalmic ointment	2
	Reclosable Polyethylene bag	1
	Instruction Sheet	1

3.2 Materials.

3.2.1 Camouflaged Eye pad. Camouflaged eye pad shall be 2 ±1/16 inches in diameter and shall be composed of a filler of a suitable quantity (approximately 0.8 gram) of USP grade absorbent cotton, covered top and bottom with single thicknesses of type 1, 44-36 gauze, camouflaged, in accordance with the requirement of Specification CCC-G-101. The camouflaged eye pad shall be free from loose threads and ravelings. The finished color of the camouflaged eye pad shall be brown. Shall approximate Color 30219 or Color number 30324 of FED-STD-595.

3.2.2 Adhesive tape. The adhesive tape shall be clear, hyporeactive, porous, conforming, plastic surgical tape having nominal dimensions of 1 ±1/16 inch wide by 8 ±1/8 inches long. Eight (8) adhesive strips shall be placed on an easy-to-separate backing material. The backing material shall be split or tabbed to facilitate separation of the backing from the tape. Each piece of adhesive tape and backing shall be joined by perforation, cutting or rouletting so that individual adhesive tape strips may be easily pulled apart from the adjoining strip.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

3.2.3 Sulfacetamide sodium ophthalmic ointment, USP, 10%, 1/8 oz. (3.5 Gm). Shall be NSN 6505-00-982-0288. Not less than 24 months of expiration date shall remain at time of delivery to the government. Ophthalmic ointment shall be stored in a controlled environment of 59 to 86°F while not in an assembled kit.

3.2.4 Reclosable polyethylene bag. The plastic bag used to contain the components of the kit shall be a sealed, reclosable, polyethylene bag. Shall be fabricated from 0.004 inch clear polyethylene, conforming to L-P-378, type I, class 1, grade A, finish 2. Shall be fitted with a 3-1/8 inch reclosable, interlocking closure for easy opening and reclosing by finger pressure. Shall have a sealed overflap above the interlocking closure, which shall be readily removable by tearing. Shall be in accordance with DPSC Drawing 24211. The lot number printed on the kit container shall correlate the Sterilization Lot Number of the eye pads contained therein, see 3.4.1. Expiration date indicated on tubes of ophthalmic ointment shall be indicated on bag adjacent to ophthalmic ointment stock number as indicated in the drawing.

3.2.5 Component polyethylene bags. The polyethylene bags specified herein shall be made from extruded stainless "layflat tubing". Polyethylene shall have average wall thickness of 0.003 inch. Minimum thickness at any point shall be not less than 0.00225 inch. Both seam and closure shall be heat sealed to prevent moisture from entering or to prevent escape of fluid contents in the event of leakage; excess air shall have been excluded from the sealed bag so that bag conforms to shape of the component therein. To insure a minimum of excess material, the width of commercially available tubing shall be of a dimension which shall closely fit the individual component to be enclosed. The length of the tubing between seals shall be the minimum to contain the component. Excess tubing beyond the seals shall be cut off. Each sealed bag containing an individual item shall be capable of passing the leakage test specified (see 4.4.1).

3.2.6 Instruction sheet. A list of components, expiration date, and instructions for use shall be printed in black ink on a white background directly on the container with English on one side and French on the other. Alternately, the list of components and instructions may be printed in black ink on white paper and inserted in the container. The printing shall be in accordance with DPSC Drawing No. 24211. The contractor is responsible for adding the expiration dating of the ophthalmic ointment.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

3.3 Camouflaged Eye pad packaging. Each camouflaged eye pad (3.2.1) shall be individually packaged in an envelope fabricated of natural kraft-colored glassine or parchment paper and sealed. The envelope shall be so constructed and sealed that sterility of the content is maintained until the envelope is opened. Individually packaged eye pads shall be sterilized by steam sterilization. Eye pads shall be sterile when tested as specified in 4.4.2.

3.4 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item identification.
b. National Stock Number.
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, dry heat,, steam, etc).
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or a designated qualified alternate.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-MQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, PA 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

3.4.1 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and simultaneously sterilized in the same sterilizer chamber. Each sterility lot shall be traceable to the pertinent manufacturing lot(s) (see 3.2.4).

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:00Z
Check the source to verify that this is the current version before use.

3.5 Assembly of components.

3.5.1 Ophthalmic ointment. The contractor shall remove each tube of ophthalmic ointment (3.2.3) from its cardboard carton, and insert it in a polyethylene bag as specified in 3.2.5.

3.5.2 Adhesive tape strips and camouflaged eye pad. The contractor shall tear off a unit of 4 adhesive strips (3.2.2) and insert it together with two camouflaged eye pads (3.3) in a 3 mil polyethylene bag and heat seal as specified in 3.2.5.

3.5.3 Pre-assembly instructions. The specified components shall be placed within the bag in such a manner as to form a flat packet. The components shall be placed between the English and French instructions so that no component shall obscure any of the listing and instructions for use. The same guidance shall apply to using a white paper background. Care shall be taken to provide a clean environment for the purpose of staging and final assembly of the first aid kits.

3.5.4 Final assembly. The contractor shall place two (2) bagged tubes of ophthalmic ointment (3.5.1), two bagged units of adhesive tape strips with camouflaged eye pads (3.5.2), and instruction sheets (3.2.6) in one reclosable plastic bag specified in 3.2.4. The contractor shall heat seal the recloseable bag and trim excess material from the heat seal.

3.5.5 Storage of assembled kits. The assembled first aid kits shall be stored in a controlled environment and shall not be subjected to temperatures below 40°F or above 95°F for any length of time. This restriction does not apply to kits being shipped on common carriers to the Government.

3.6 Marking for identification.

3.6.1 Camouflaged eye pad. Each envelope containing the camouflaged eye pad (3.3) shall bear the sterility clause: "STERILITY IS NOT GUARANTEED IF INDIVIDUAL ENVELOPE HAS BEEN DAMAGED OR OPENED", or similar statement and the name or registered trademark of manufacturer. When the contractor is not the manufacturer, the contractor's name shall also appear.

3.7 Workmanship. The individual components, as well as the completely assembled First Aid Kit, Eye Dressing, shall be free from defects which detract from their appearance or impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS.

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Responsibility for compliance. All items shall meet all requirements of sections 3 and 5. The inspection set forth in this specification shall become a part of the contractor's overall inspection system or quality program. The absence of any inspection requirements in the specification shall not relieve the contractor of the responsibility of ensuring that all products or supplies submitted to the Government for acceptance comply with all requirements of the contract. Sampling inspection, as part of manufacturing operations, is an acceptable practice to ascertain conformance to requirements, however this does not authorize submission of known defective material, either indicated or actual, nor does it commit the Government to accept defective material.

4.1.2 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.3 Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:00Z
Check the source to verify that this is the current version before use.

4.1.4 Certificates of quality. Certificates of quality, supplied by the manufacturer of the component material, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government and provided also that the test methods and the results obtained comply with those described in their specification. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table 1. Unit of product for examination purposes shall be one kit.

Table I. For examination.

	Inspection level	AQL (percent defective)
For visual examination		
Major A	II	1.0
Major B	II	1.5
Minor	II	4.0
For dimensional examination	S-2	4.0

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II. The inspection level shall be S-2 and the acceptance number shall be zero. The unit of product for sampling is shown in table II.

Table II For tests.

Component	Characteristics	Requirements	Test procedure
Bagged Components	Leakage test	3.2.5	4.4.1
Eye pad	Sterilization Test	3.3	4.4.2*

* Five individually packaged envelopes from each lot shall be tested.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

4.3 Examination. The first aid kit shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table III. Examination shall not be restricted to the classified possible defects listed below.

TABLE III. Classification of defects.

Categories	Defects
Major A	
101	Plastic bag (kit container) not in conformance with drawing.
102	Plastic bag (kit container) heat seals not continuous.
103	Plastic bag (kit container) not tamper-proof type.
104	Plastic bag (components) cut or torn.
105	Any component missing.
106	Marking incorrect, incomplete, illegible, or non-permanent.
107	Ophthalmic ointment tube leak.
108	Eye pad paper envelope not intact.
Major B	
109	Polyethylene bag heat sealed in the closure seam of the plastic bag (kit container).
Minor	
201	Excess air not removed from polyethylene bag before sealing.
202	Excess material not removed from bottom of polyethylene bag after sealing.
203	Excess material not removed from bottom of plastic bag (kit container) after sealing.
204	Printing on plastic bag (kit container) off center.
205	Complete kit package not flat.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

4.3.2 Dimensional examination. The component parts shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defect.

4.3.3 Inspection of packaging. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Leakage tests. Sealed, filled, polyethylene bags (3.2.5) shall be submerged to a depth of one foot in tap water at room temperature for 15 minutes. Bags shall be removed, exterior surface dried, opened, and examined for leakage of water.

4.4.2 Sterility test. The eye pads shall be tested for sterility in accordance with the method specified in the USP for Purified Cotton, Gauze, Surgical Dressings, and Related Material using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days of inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in paragraph 4.4.2.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for type I sterilization.

4.4.2.1 Inoculated product or simulated product. For steam sterilization, the organism used shall be stearothermophilus at an incubation temperature of 55° - 65°C. The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

4.4.2.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.2. Sterility test and 4.4.2.1 Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.4.2.1. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:00Z
Check the source to verify that this is the current version before use.

4.4.2.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a Government contract.

4.4.2.4 Product testing. Testing for sterility of the product after receipt by the Government shall be without a biological indicator and in accordance with the USP.

5 PACKAGING

5.1 Preservation. Preservation shall be level A or C, as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Cleaning and drying. First Aid Kit shall be cleaned and dried in accordance with MIL-P-116.

5.1.1.2 Preservatives applications. Preservatives shall not be used.

5.1.1.3 Unit package. Each kit shall be packaged in a sealed, recloseable, polyethylene bag. Closure shall be as specified in the applicable box specification.

5.1.1.4 Intermediate package. Twelve unit packages shall be packaged in a box conforming to PPP-B-566, PPP-B-636, Type CF, class domestic or PPP-B-676. Closure shall be as specified in the applicable box specification.

5.1.2 Level C.

5.1.2.1 Unit package. Preservation for First Aid Kit shall conform to the requirements of MIL-STD 794 for level C.

5.1.2.2 Intermediate package. Twelve unit packages containing first aid kits shall be packaged in accordance with the requirements of MIL-STD-794 for level C.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

5.2.1 Level A. Forty-eight intermediate packages (576 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type, PPP-B-621, class 2 or PPP-B-636, class weather resistant. Closure and strapping shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Forty-eight intermediate packages (576 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type, PPP-B-621, class 1, or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. The level C packing for First Aid Kit shall be packed in accordance with the requirements of MIL-STD-794 for level C.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination exceed 250 pounds or 20 cubic feet. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 20 cubic feed need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in length, 52 inches in width, and 54 inches in height. Pallet shall conform to Type IV or Type V.

5.2.5.1 Levels A and B. First Aid Kit packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets as specified in 5.2.5.

5.2.5.2 Level C. First Aid Kit packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Levels A, B and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1 and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.3.3.

6. NOTES

6.1 Intended use. The First Aid Kit covered by this specification is intended for treatment of eye injuries.

6.2 Acquisition Requirements. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) National Stock Number (NSN).
(c) Selection of applicable level of packaging and packing (see 5.1 and 5.2).

6.3 Subject term (keyword) listing.

Eye Pad
Adhesive tape
Ophthalmic ointment
Plastic bag
Instruction sheet

6.4 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification or those which are commercially available, but is intended to cover the types which are normally procured to meet military requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

6.5 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6545-00-853-6309	FIRST AID KIT, Eye Dressing

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing activity:

DLA-DM

Review activity:

DoD-MB

Project No. 6545-0092

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	MIL-F-36043D	22 JUNE 1992

3. DOCUMENT TITLE
FIRST AID KIT, EYE DRESSING

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215) 737-2870	444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete. — 198/290

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:00Z
Check the source to verify that this is the current version before use.